



저작자표시-비영리-변경금지 2.0 대한민국

이용자는 아래의 조건을 따르는 경우에 한하여 자유롭게

- 이 저작물을 복제, 배포, 전송, 전시, 공연 및 방송할 수 있습니다.

다음과 같은 조건을 따라야 합니다:



저작자표시. 귀하는 원저작자를 표시하여야 합니다.



비영리. 귀하는 이 저작물을 영리 목적으로 이용할 수 없습니다.



변경금지. 귀하는 이 저작물을 개작, 변형 또는 가공할 수 없습니다.

- 귀하는, 이 저작물의 재이용이나 배포의 경우, 이 저작물에 적용된 이용허락조건을 명확하게 나타내어야 합니다.
- 저작권자로부터 별도의 허가를 받으면 이러한 조건들은 적용되지 않습니다.

저작권법에 따른 이용자의 권리는 위의 내용에 의하여 영향을 받지 않습니다.

이것은 [이용허락규약\(Legal Code\)](#)을 이해하기 쉽게 요약한 것입니다.

[Disclaimer](#)

약학박사학위논문

DNA의 특질 분석을 이용한 드롭렛 디지털 PCR 기반의 *EGFR*

돌연변이 검출 방법 연구

**Droplet digital PCR-based *EGFR* mutation detection with an  
internal quality control index to determine the quality of DNA**

서울대학교 대학원

약학과 병태생리학 전공

김 성 수

## Abstract

### **Droplet digital PCR-based *EGFR* mutation detection with an internal quality control index to determine the quality of DNA**

Sung-Su Kim

College of Pharmacy

The Graduate School

Seoul National University

Formalin-fixed paraffin-embedded tissue (FFPET) samples are invaluable sources for both translational clinical research and molecular *in vitro* diagnostics. However, accurate detection of genetic mutations in FFPET is a major challenge due to artifactual results, due to sample age and quality. In a pre-clinical study, we used 315 non-small cell lung cancer (NSCLC) FFPET-derived DNA (FFPET-DNA) samples to establish sample criteria reflecting the minimum DNA quality suitable for PCR by comparing the results of droplet digital PCR-based mutation test (ddEGFR test) and qPCR-based EGFR mutation test (cobas<sup>®</sup> EGFR test). Using this criteria, we conducted a retrospective comparative clinical study of the ddEGFR and cobas EGFR tests of 171 NSCLC FFPET-DNA samples. Based on the pre-clinical study, the FFPET-DNA sample

criterion was established as internal quality control (iQC) index  $\geq 0.5$  (iQC copies  $\geq 500$ , using 3.3 ng [1000 genome equivalents] of FFPET-DNA), indicating that more than half of the input DNA was amplifiable. Based on this iQC index, an independent clinical study revealed that both tests were significantly concordant (overall percent agreement (OPA) = 92.98%). Discordants not detected by the cobas EGFR test were detected using the ddEGFR test, indicating that the higher sensitivity of the ddEGFR test is due to its lower limit of detection.

iQC index is a reliable indicator of the quality of FFPET-DNA and could be used to prevent incorrect diagnoses arising from low-quality samples. Compared with the cobas EGFR test, the ddEGFR test exhibited superior analytical performance and at least equivalent clinical performance. ddEGFR test is expected to serve as an appropriate treatment guidance for NSCLC patients.

**Running title:** ddPCR-based *EGFR* mutation detection with DNA iQC index

**Key words:** *EGFR*, epidermal growth factor receptor; FFPE, Formalin-fixed paraffin embedded; DIN, DNA integrity number; ddPCR, Droplet digital PCR; Genetic mutation detection; iQC, internal quality control, Liquid biopsy; ctDNA, Circulating tumor DNA

**Student number:** 2011-30502

# Contents

|                                    |             |
|------------------------------------|-------------|
| <b>Abstract</b> .....              | <b>ii</b>   |
| <b>Contents</b> .....              | <b>iv</b>   |
| <b>List of Tables</b> .....        | <b>vi</b>   |
| <b>List of Figures</b> .....       | <b>viii</b> |
| <b>List of Abbreviations</b> ..... | <b>x</b>    |

## **I. Introduction**

|  |    |
|--|----|
| 1. <i>EGFR</i> mutations in NSCLC .....                                  | 2  |
| 2. Companion diagnostics (CDx) .....                                     | 9  |
| 3. DNA damage in FFPEs .....   | 13 |
| 3.1 DNA fragmentation .....  | 15 |
| 3.2 Hydrolytic deamination .....   | 15 |
| 3.3 Formaldehyde-induced abasic site .....                               | 17 |
| 3.4 Formaldehyde-induced cross-link .....                                | 17 |
| 3.5 DNA damage during preparation of FFPE blocks (physical stress) ..... | 17 |
| 3.6 Strategies for minimizing sequence artifact .....                    | 19 |
| 4. Liquid biopsy .....   | 22 |

|   |            |
|---|------------|
| 5. Third generation PCR: Droplet digital PCR (ddPCR) .....            | 24         |
| 6. Limit of blank, limit of detection and limit of quantitation ..... | 27         |
| 6.1 Limit of blank (LoB) .....  | 27         |
| 6.2 Limit of detection (LoD) .....                                    | 27         |
| 6.3 Limit of quantitation (LoQ) .....                                 | 28         |
| <br>  |            |
| <b>II. Purpose of the study .....</b>                                 | <b>30</b>  |
| <b>III. Materials and Methods .....</b>                               | <b>33</b>  |
| <b>IV. Results .....</b>  | <b>43</b>  |
| <b>V. Discussion .....</b>  | <b>81</b>  |
| <b>VI. Conclusion and perspectives .....</b>                          | <b>87</b>  |
| <b>VII. Reference .....</b>   | <b>94</b>  |
| <br>  |            |
| <b>국문 초록 .....</b>  | <b>114</b> |

## List of Tables

**Table 1.** Gefitinib, erlotinib, and afatinib versus platinum-doublet chemotherapy as first line therapy for advanced *EGFR*-mutated NSCLC

**Table 2.** Third-generation *EGFR*-TKIs developed for treatment of NSCLC

**Table 3.** Two commercial kits for *EGFR* mutation testing

**Table 4.** Strategies for minimizing sequence artifacts in FFPE-DNA

**Table 5.** qPCR versus ddPCR

**Table 6.** Forty-five *EGFR* mutations detected by the dd*EGFR* mutation test

**Table 7.** dd*EGFR* test coverage

**Table 8.** A total of 316 samples with valid dd*EGFR* and cobas *EGFR* test results were included in the agreement analysis

**Table 9.** Detail concordance table of 316 samples with valid dd*EGFR* and cobas *EGFR* test

**Table 10.** Concordance of *EGFR* mutation detection between the cobas *EGFR* test and Sanger sequencing, without application of exclusion criteria

**Table 11.** Detail concordance of *EGFR* mutation detection between the cobas *EGFR* test and Sanger sequencing, without application of exclusion criteria

**Table 12.** Comparison of dd*EGFR* and cobas *EGFR* results from FFPE samples (block storage duration > 6, DIN < 2.5)

**Table 13.** Detail concordance rate of group 1 samples (block storage duration > 6, DIN < 2.5)

**Table 14.** Correlation between ddEGFR iQC index and DIN value

**Table 15.** Comparison of ddEGFR and cobas EGFR results from FFPE samples eliminated by the sample criteria (DIN and iQC index)

**Table 16.** Comparison between the ddEGFR and cobas EGFR test results with application of the sample criteria (Group 2)

**Table 17.** Comparison of ddEGFR and cobas EGFR results from FFPE samples eliminated by the sample criteria (iQC index only)

**Table 18.** Comparison between the ddEGFR and cobas EGFR test results, with application of only the iQC criterion (Group 3)

**Table 19.** Re-analysis of concordance rate using excluded samples

**Table 20.** Detail concordance rate using excluded samples

**Table 21.** Re-analysis of eight of eleven discordant samples in Group 3

**Table 22.** Method correlation between ddEGFR and cobas EGFR test in Retrospective comparison study group

**Table 23.** Detailed concordance rate of the retrospective comparison study group

**Table 24.** Discordant analysis by Sanger sequencing in the retrospective comparison study group

## List of Figures

**Figure 1.** EGFR-TKI sensitizing mutations of *EGFR* in NSCLC

**Figure 2.** Hypothesis explaining T790M mutation-related resistance to EGFR-TKIs (gefitinib or erlotinib)

**Figure 3.** Various DNA damage in FFPE

**Figure 4.** Hydrolytic deamination of cytosine residues

**Figure 5.** Schematic showing the ddPCR workflow

**Figure 6.** Relationship between the LoB, LoD, and LoQ

**Figure 7.** Study design and specimen selection (work flow of pre-clinical study)

**Figure 8.** Study design and specimen selection (work flow of retrospective comparison study)

**Figure 9.** The concept of iQC index

**Figure 10.** Determination of a suitable cut-off based on false-positive analysis using normal FFPE blocks

**Figure 11.** Validation of internal quality control of ddEGFR test (MI validation)

**Figure 12.** Validation of internal quality control of ddEGFR test (iQC copy and iQC index)

**Figure 13.** Distributions of ddEGFR iQC index corresponding to sample storage time

**Figure 14.** Distributions of DIN value corresponding to sample storage time

**Figure 15.** Establishment of sample criteria using discordant samples (DIN value)

**Figure 16.** Establishment of sample criteria using discordant samples (iQC index)

**Figure 17.** Schematic of re-analysis of discordant samples

**Figure 18.** Distributions of iQC index and DIN in the retrospective comparative study group

**Figure 19.** Clinical meaning of iQC index (risk management)

**Figure 20.** Clinical meaning of iQC index (viewpoint of multi-center clinical and translational research)

**Figure 21.** Dynamic monitoring of ctDNA in NSCLC

## List of Abbreviations

| <b>Abbreviation</b> | <b>Word</b>                      |
|---------------------|----------------------------------|
| <b>EGFR</b>         | Epidermal growth factor receptor |
| <b>NSCLC</b>        | Non-small cell lung cancer       |
| <b>CDx</b>          | Companion diagnostics            |
| <b>TKI</b>          | Tyrosine kinase inhibitor        |
| <b>FFPE</b>         | Formalin-fixed paraffin embedded |
| <b>DIN</b>          | DNA integrity number             |
| <b>PCR</b>          | Polymerase chain reaction        |
| <b>ddPCR</b>        | Droplet digital PCR              |
| <b>iQC</b>          | Internal quality control         |
| <b>PFS</b>          | Progression free survival        |
| <b>SNV</b>          | Single nucleotide variation      |
| <b>UDG</b>          | Uracil-DNA glycosylase           |
| <b>MPS</b>          | Massive parallel sequencing      |
| <b>C/N ratio</b>    | Cancer/normal ratio              |
| <b>TPS</b>          | Tissue preparation system        |
| <b>gDNA</b>         | genomic DNA                      |
| <b>PPA</b>          | Positive percent agreement       |
| <b>NPA</b>          | Negative percent agreement       |
| <b>OPA</b>          | Overall percent agreement        |
| <b>PPV</b>          | Positive percent value           |
| <b>NPV</b>          | Negative percent value           |

|               |   |
|---------------|---|
| <b>MI</b>     | Mutation index                              |
| <b>LoB</b>    | Limit of blank                              |
| <b>LoD</b>    | Limit of detection                          |
| <b>LoQ</b>    | Limit of quantitation                       |
| <b>SD</b>     | Standard deviation                          |
| <b>G.E.</b>   | Genomic equivalence                         |
| <b>MD</b>     | Mutation detected                           |
| <b>MND</b>    | Mutation not detected                       |
| <b>COSMIC</b> | Catalogue of somatic mutations in cancer    |
| <b>CLSI</b>   | Clinical and laboratory standards institute |
| <b>MFDS</b>   | Ministry of food and drug safety            |
| <b>cfDNA</b>  | Cell free DNA                               |
| <b>ctDNA</b>  | Circulating tumor DNA                       |

# **I. Introduction**

## 1. *EGFR* mutations in NSCLC

In light of recent advancements in personalized medicine, nucleic acid-based diagnostics will play a pivotal role in implementation of targeted therapies. For example, the activation of epidermal growth factor receptor (*EGFR*) pathways results in the initiation of cancer proliferation, metastasis, and angiogenesis. *EGFR* and its relatives are known to be oncogenic drivers in cancers such as lung cancer (1), breast cancer (2), and glioblastoma (3-5).

Lung cancer, which accounts for one-third of cancer-related deaths worldwide, is divided into small cell lung cancer (SCLC; which accounts for 20% of cases) and non-small cell lung cancer (NSCLC; which accounts for the other 80%). SCLC is a neural crest tumor that responds well to chemotherapy, at least initially; however, it often recurs as treatment-resistant disease. NSCLC, thought to be derived from lung epithelial cells, includes various subtypes: adenocarcinoma, bronchioloalveolar carcinoma, squamous cell carcinoma, and anaplastic and large cell carcinoma. Most patients with advanced NSCLC have a very high probability of metastasis and, if not treated, the mean survival time and 1 year survival rate is at 4–5 months less than 10% respectively (6). Combination cytotoxic chemotherapy increases the survival rate only slightly and causes serious non-specific toxicity to the patient (7). However, the problem of chemotherapy has been raised in the appearance of molecular targeted therapeutic drugs at NSCLC, and small molecule inhibitor of *EGFR* kinase for treatment of NSCLC was approved in 2005 (8).


Inhibiting *EGFR* kinase activity using tyrosine kinase inhibitors (TKIs) such as erlotinib and gefitinib slows disease progression in patients with NSCLC (9-11). *EGFR* kinase domain mutations occur in four exons encoding part of the tyrosine kinase domain of the enzyme and are clustered around the ATP-binding pocket (12). Deletions in *EGFR* exon 19 and the point

mutation L858R constitute about 90% of all *EGFR*-activating mutations in NSCLC. These mutations correlate with a better TKI treatment response rate and with significant increases in progression free survival (PFS) (13-18) (Fig. 1 and Table. 1). Some reports suggest that *EGFR*-activating mutations may be a prognostic factor for EGFR-TKI efficacy.

Despite EGFR-TKIs such as erlotinib or gefitinib being used as a standard treatment for NSCLC patients harboring EGFR mutations (19), the median PFS still does not exceed 10 months (20). This is because of the secondary point mutation T790M, which is associated with acquired resistance to EGFR-TKIs and is found in up to 50% of post-TKI tumor samples (21-24). T790M increases receptor binding affinity for ATP, thereby reducing drug activity. Recent evidence suggests that the T790M mutation is present in tumor cells at low frequency during development of NSCLC, but there is evidence that tumor cells harboring this mutation might be enriched during treatment with EGFR-TKIs (25, 26) (Fig. 2). Also, pre-existing *EGFR* T790M mutations are associated with poor clinical outcome after EGFR-TKI therapy for NSCLC (27, 28).

Therefore, new EGFR-TKIs that bind specifically to T790M-mutated receptors have been developed and used to treat patients with acquired resistance (29-31). Osimertinib (AZD9291), rociletinib (CO-1686), olmutinib (HM61713), and others (Nazartinib, Naquotinib) are undergoing clinical trials, and osimertinib (TAGRISSO™) has been approved by the U.S. FDA as a treatment for T790M-positive NSCLC (32) (Table.2). Although exciting data and response rates have been registered in patients treated with third generation EGFR-TKIs, acquired resistance still occurs after about 10 months (29, 30). C797S is another acquired *EGFR* mutation that inhibits the activity of EGFR-TKIs. The first evidence of C797S isolated from NSCLC patients was reported by Thress *et al* (33). The authors analyzed plasma samples from 19 patients with acquired resistance to osimertinib and identified C797S in six. The C797S


mutation appears to be an ideal target for overcoming acquired resistance to third generation inhibitors. Thus, accurate detection of *EGFR* mutations is critical for individualized treatment strategies for NSCLC.



**Figure 1. EGFR-TKI sensitizing mutations of *EGFR* in NSCLC.** Exons 18–21 in the tyrosine kinase region, in which the relevant mutations are located, are expanded (cyan bars). A detailed list of *EGFR* mutations associated with sensitivity (magenta boxes) or resistance (yellow boxes) to first and third generation TKIs is show for each exon.

| Drug      | Clinical trial | RR (%)   |                  | PFS (months) |                  | OS (months)  |                  | Reference |
|-----------|----------------|----------|------------------|--------------|------------------|--------------|------------------|-----------|
|           |                | EGFR-TKI | Platinum-doublet | EGFR-TKI     | Platinum-doublet | EGFR-TKI     | Platinum-doublet |           |
| Gefitinib | IPASS          | 71.2     | 47.3             | 9.5          | 6.3              | 21.6         | 21.9             | (34, 35)  |
|           | NEJ002         | 73.7     | 30.7             | 10.8         | 5.4              | 30.5         | 23.6             | (17)      |
|           | WJTOG3405      | 62.1     | 32.2             | 9.2          | 6.3              | 30.9         | not reached      | (36)      |
| Erlotinib | OPTIMAL        | 83       | 36               | 13.1         | 4.6              | 22.6         | 28.8             | (37)      |
|           | EURTAC         | 58       | 15               | 9.7          | 5.2              | 19.3         | 19.5             | (20)      |
| Afatinib  | LUX-Lung 3     | 56       | 23               | 11.1         | 6.9              | not reported | not reported     | (38)      |
|           | LUX-Lung 6     | 66.9     | 23               | 11           | 5.6              | 22.1         | 22.2             | (39)      |

**Table 1. Gefitinib, erlotinib, and afatinib versus platinum-doublet chemotherapy as first line therapy for advanced *EGFR*-mutated NSCLC.**



**Figure 2. Hypothesis explaining T790M mutation-related resistance to EGFR-TKIs (gefitinib or erlotinib).** Treatment with EGFR-TKIs may result in selection of T790M mutant cells. Thus, even if a small number of T790M-positive tumor cells are present at the beginning of treatment, selective proliferation of these cells could lead to clinical resistance to erlotinib.

| Drug<br>[research<br>code]      | Trial or<br>ClinicalTrials.gov<br>no.* | ORR (%) |        | PFS (months) |        | OS (months)  |              | Status              | Ref. |
|---------------------------------|--|---------|--------|--------------|--------|--------------|--------------|---------------------|------|
|                                 |  | T790M+  | T790M- | T790M+       | T790M- | T790M+       | T790M-       |                     |      |
| <b>Osimertinib</b><br>[AZD9291] | AURA 3                                 | 71      | 31     | 10.1         | 4.4    | not reported | not reported | Approved            | (40) |
| <b>Rociletinib</b><br>[CO-1686] | NCT01526928                            | 59      | 29     | 13.1         | 5.6    | not reported | not reported | Stopped             | (30) |
| <b>Olmutinib</b> †<br>[HM61713] | ELUXA 1                                | 58.8    | -      | -            | -      | -            | -            | Phase II<br>ongoing | (41) |
| <b>Nazartinib</b><br>[EGF816]   | NCT02108964                            | 54.5    | -      | -            | -      | -            | -            | Phase II<br>ongoing | (42) |
| <b>Naquotinib</b><br>[ASP8273]  | NTC02192697                            | 64      | -      | -            | -      | -            | -            | Phase II<br>ongoing | (43) |

\* <https://clinicaltrials.gov/ct2/home>

† Olmutinib has been approved by the Korea Ministry of Food and Drug Safety (MFDS).

**Table 2. Third-generation EGFR-TKIs developed for treatment of NSCLC.**

## 2. Companion diagnostics (CDx)

It is increasingly common for clinicians to analyze molecular characteristics of tumors prior to choosing the most appropriate treatment. When used alongside clinically validated and accurate companion diagnostics, the therapeutic effect of EGFR-TKI can be maximized.

In July 2011, the U.S. FDA announced guidelines on the use of companion diagnostics as an important tool for guiding the selection and use of appropriate therapies (*In vitro* companion diagnostics devices – guidance for industry and FDA staff, US FDA, 2014). According to this definition, CDx analysis is an *in vitro* diagnostic device that provides essential information for the safe and effective use of therapeutic drugs. The FDA also specifies three areas in which a CDx assay is required: [i] to identify patients most likely to benefit from a specific therapeutic product; [b] to identify patients who may suffer serious side effects as a result of treatment with a specific therapeutic product; [c] to monitor treatment responses and adjust the regimen (e.g., drug dosage, treatment time) to increase safety and efficacy. The aim of companion diagnostics is to thoroughly understand the molecular pathology and mechanism of action of a drug so that the properties of a particular molecule are suited to the required treatment outcome.

A previous report highlights the importance of companion diagnostics comprising trastuzumab (Herceptin<sup>®</sup>, Roche/Genentech) and immunohistochemical (IHC) analysis of HER2-positive advanced breast cancer (44). In 1998, the FDA approved trastuzumab and IHC for analysis of HER2 overexpression (HercepTest<sup>™</sup>, Dako); since then, several new targeted anticancer agents have been approved and introduced into the clinic. If an appropriate drug and dose are to be tailored to suit a particular individual, robust and accurate analysis methods that have a short turnaround time are needed. Important factors include standardization of analytical methods, validation of reagents and methods, qualified laboratory experience, and involvement

of pathologists. However, recent quality assurance studies conducted to confirm the mutation status of a standard panel of tumors indicate that, despite using the same or similar methods, laboratories may obtain different results (45, 46). Not only differences in methods, but also differences in the characteristics of the sample tested and environmental factors at each test site, can affect the result; therefore, it is imperative that all procedures are standardized.

Many techniques were developed instead of direct sequencing for *EGFR* mutation analysis. This is because direct sequencing has low sensitivity, is complex and time consuming, and is very difficult to standardize (47). Therefore, new techniques have been developed to detect known and *de novo* mutations, albeit with varying sensitivity (48). The Roche Cobas® *EGFR* Mutation Test (Roche diagnostics, Basel, Switzerland) detects 42 mutations in *EGFR* exons 18, 19, 20, and 21. In parallel with approval of osimertinib as a second-line treatment for patients harboring the T790M mutation, the Cobas *EGFR* test has been approved by the U.S. FDA. In particular, the assay overcomes the limitation of requiring tumor tissues because *EGFR* mutation status is assessed in a non-invasive manner. The safety and effectiveness of the Cobas *EGFR* test was confirmed by retrospective clinical validation studies using specimens from NSCLC patients screened in the EURTAC trial (49) and establishment through AURA 3 clinical trial using matched plasma to the tissues (40). This test allows for *EGFR* mutation screening in a single reaction with a 1-day turnaround time, and has greater sensitivity and specificity than Sanger sequencing.

In July 2013, the Scorpion Amplified Refractory Mutation System (ARMS) theascreen® *EGFR* Rotor-Gene® Q PCR kit (therascreen *EGFR* test; QIAGEN, Hilden, Germany) was approved by the U.S. FDA for use as a companion diagnostic for afatinib. A recent study reported a high concordance rate between the theascreen *EGFR* test and the Cobas *EGFR* test (50). However, the Cobas *EGFR* test has some advantages. For example, the time from


obtaining a tumor sample to obtaining a result is shorter, meaning that treatment can begin more quickly. In addition, Cobas EGFR test requires a very small amount of DNA to detect the mutation status of a tumor sample. The comparison of the two kits are shown in Table 3.

|                            | Roche Cobas® EGFR mutation test   | QIAGEN theascreen EGFR RGQ mutation test   |
|----------------------------|---|--|
| <b>Platform</b>            | <b>Real time-PCR</b>  | <b>Real time-PCR</b>   |
| <b>Instrument</b>          | Cobas z 480 Analyzer  | Rotor-Gene® Q MDx  |
| <b>IVD, CDx Label</b>      | <i>EGFR</i> _CDx (US-FDA approved)  | <i>EGFR</i> _CDx (US-FDA approved)   |
| <b>Drug</b>                | TARCEVA® (Erlotinib), TAGRISSO™ (Osimertinib)   | GILOTRIF® (Afatinib)   |
| <b>Intended use</b>        | The test is intended to be used as an aid in selecting patients with NSCLC for whom TARCEVA® (erlotinib) and TAGRISSO™ (Osimertinib), an <i>EGFR</i> tyrosine kinase inhibitor (TKI), is indicated. | The test is intended to be used to select patients with NSCLC for whom GILOTRIF™ (afatinib), an <i>EGFR</i> tyrosine kinase inhibitor (TKI), is indicated. |
| <b>Mutation coverage</b>   | G719X, E19del, T790M, S768I, E20Ins, L858R, L861Q<br>Total: 42 mutation sites   | E19del, L858R(CDx) + L681Q, G719X, S768I, exon 20 insertions, and T790M<br>Total: 15 mutation sites  |
| <b>Detection method</b>    | Allele-specific primer  | ARMS and Scorpions   |
| <b>Input DNA</b>           | FFPE : 50 ng / well<br>cfDNA : 25 µl / well   | FFPE : 80 ng / well<br>-   |
| <b>Reaction (well) no.</b> | 3   | 8  |
| <b>Sensitivity</b>         | FFPE : 5%<br>cfDNA : <100 copies /ml  | FFPE : 1%<br>-   |
| <b>Application</b>         | FFPE, cfDNA   | FFPE   |
| <b>Running time</b>        | ~2 h  | ~3 h   |

**Table 3. Two commercial kits for *EGFR* mutation testing**

### **3. DNA damage in FFPETs**

Formalin-fixed, paraffin-embedded tissue (FFPET) is the most widely available material for molecular diagnostics and clinical research (51). Fixing cancer tissues in buffered formalin is a standard procedure because it preserves tissue and cellular morphology prior to examination by anatomical pathologists. The routine formalin-fixation process stabilizes proteins by cross-linking, thereby maintaining tissues in an excellent condition for histopathological analysis (52). FFPET samples can be stored at room temperature, which means that damage caused by freezing can be avoided. However, the FFPET fixation procedure and long-term storage at room temperature causes several kinds of damage to nucleic acids, creating challenges to molecular analyses using FFPET-derived DNA (FFPET-DNA) (Fig. 3) (53).




**Figure 3. Various DNA damage in FFPET.** The FFPETs have some advantages, such as easier storage, but it is well known that formalin fixation results in DNA damage.

### 3.1 DNA fragmentation

Increased storage time and low pH lead to fragmentation of DNA in FFPET. PCR performed by using fresh FFPET is usually much more successful than that of using FFPET that has been stored for some time, indicating that DNA fragmentation occurs continuously during storage (54-57). Fragmented DNAs reduce the success of PCR amplification since this depends on the size of the amplicon (58). Also, DNA fragmentation correlates inversely with sequence coverage and can lead to substantial changes in the amount of amplifiable template, which is in turn related to the presence of sequencing artifacts (57, 59).

### 3.2 Hydrolytic deamination

Hydrolytic deamination of cytosine residues yields uracil lesions, which are a major source of artifacts when sequencing FFPET-DNA (60-62). In the human genome, cytosine deamination occurs at an estimated rate of 60–500 events/day (63). Among the sequence artifacts detected in FFPET DNA, C:G > T:A transitions are the most common type of single nucleotide variant (Fig. 4) (64). Recent comparisons of sequencing data generated from fresh-frozen and FFPET demonstrated that the majority of false-positive and -negative mutation calls were made for FFPET samples, particularly in low quality samples (64, 65). Recently, Do, H *et al.* have examined sequence artifacts in FFPET DNA samples, highlighting the importance of DNA lesions (66). For example, the T790M mutation in *EGFR* is a predictor of TKI resistance in NSCLC patients (67); therefore, an FFPET artefact causing a false-positive result may lead to a clinician selecting the wrong treatment option (68, 69).



**Figure 4. Hydrolytic deamination of cytosine residues.** Deamination of cytosine residues results in non-reproducible artifacts. *EGFR* G719S and T790M are the most common false-positive results called after deamination.

### **3.3 Formaldehyde-induced abasic site**

Formaldehyde induced abasic sites are generated by depurination. When it reacts with oxygen in the atmosphere, formaldehyde is readily oxidized to formic acid, which reduces the pH of a formalin solution. Formalin solutions are therefore buffered to maintain a neutral pH. In DNA, the *N*-glycosidic bonds between a purine base and its deoxyribose moiety are most susceptible to hydrolysis at low pH (70); thus, fixation of tissues in unbuffered formalin will lead to a marked reduction in the amount of amplifiable DNA template (71). Abasic sites also cause problems with respect to sequence analysis. Chemical alterations in DNA block the passage of DNA polymerase, resulting in sequence artifacts (72, 73).

### **3.4 Formaldehyde-induced cross-link**

Cross-linking reduces the stability of double stranded DNA, which in turn affects the amount of FFPE DNA that can be amplified by PCR (74). Formaldehyde, the main component of formalin, induces crosslinks between proteins and between proteins and DNA; it also induces DNA-formaldehyde adducts and inter-strand DNA crosslinks, all of which have a primary genotoxic effect (75-77). Cross-linking not only causes problems in terms of DNA isolation, but also reduces PCR amplification efficiency (25). Therefore, formaldehyde-induced cross-linking reduces the overall stability of DNA (78), which makes downstream analysis difficult.

### **3.5 DNA damage during preparation of FFPE blocks (physical stress)**

As mentioned above, formalin fixation has a time-dependent adverse effect on the quality of FFPE DNA and subsequent PCR results (79). Because different laboratories use different

fixation protocols and store tissues for different times, the quality of FFPE DNA should be checked prior to its use in clinical studies. Other than storage duration and fixation procedures, fixation time is another critical factor for PCR analysis. Indeed, Inoue et al. revealed that the quality of DNA in FFPE tissues varied markedly depending on the fixation time (79). It is imperative that all nucleic acid assays used for clinical purposes use properly prepared and stored FFPE tissues. However, many hospitals lack appropriate FFPE storage facilities, resulting in numerous unqualifiable samples (80). Thus, for successful assay in the clinic, DNA quality is paramount. Furthermore, prior to the use of FFPE samples aged more than 10 years for retrospective studies, the quality of the DNA should be tested vigorously.

### 3.6 Strategies for minimizing sequence artifact

Minimizing sequence artifacts is critical for accurate detection of mutations in FFPE DNA. Accurate detection of EGFR-TKI actionable and/or resistance mutations enables identification of patients who will benefit from targeted therapy and avoid adverse effects associated with unnecessary treatment of non-responsive patients. Several strategies can be used to minimize DNA damage, which is the main limitation for any PCR-based assay and/or amplicon sequencing-based technique. Several papers describe strategies for minimizing sequence artifacts (Table. 4) (66, 81).

First, fragmentation of FFPE DNA reduces the amount of amplifiable template available. As mentioned earlier, the success rate of PCR is related to amplicon size. Therefore, it is important to consider short amplicons when designing primers and probes (82, 83). In addition, the design of shorter amplicons as well as more stricter conditions for PCR amplification could be a useful way to lower the amount of unspecific PCR products, which generate of high background (81). Moreover, careful design of primers and probes are necessary to prevent non-specific artefacts such as primer dimers and chimeric products.

Second, C:G>T:A artifacts are a common form of damage found in FFPE DNA (84). *In-vitro* removal of uracil bases from FFPE DNA using uracil-DNA glycosylase (UDG) before PCR amplification reduces this problem (84). Detection of true mutations is not affected by UDG treatment. UDG is an evolutionarily conserved DNA repair enzyme that imitates the base excision repair pathway and removes uracil from DNA (85). Thus, pre-treatment with UDG is a simple and effective strategy for reducing C:G>T:A artefacts.

Third, reversal of formaldehyde-induced DNA-DNA and DNA-protein crosslinks by heating samples to over 90°C increases the amount of amplifiable DNA template (82, 86-88).

Lastly, template quantity and quality are crucial for optimizing assay conditions and accurate detection of mutations. The same measured quantity of different FFPE DNAs can contain different amounts of amplifiable template, depending on the degree of DNA fragmentation (59). Spectrophotometric and fluorometric methods tend to seriously overestimate the actual amount of amplifiable FFPE-DNA (69). Thus, PCR-based methods such as qPCR and digital (d)PCR are more suitable for quantifying the amount of actual amplifiable FFPE DNA.

| <b>Damage</b>          | <b>Strategy</b>  |
|------------------------|--|
| <b>Fragmentation</b>   | 1. Design shorter amplicons  |
|                        | 2. Assessment of DNA integrity   |
|                        | 3. Quantification of amplifiable templates using PCR   |
| <b>Deamination</b>     | 1. Pre-treatment with UDG  |
| <b>Abasic sites</b>    | 1. Use of specific DNA polymerases (e.g. <i>pfu</i> polymerase) that have low bypass efficiency over DNA lesions such as uracil and abasic sites |
|                        | 1. Use of a high-fidelity DNA polymerase to reduce polymerase errors   |
| <b>Cross-linking</b>   | 1. Heat treatment of remove formaldehyde-induced crosslinks  |
|                        | 2. Extended proteinase K treatment to digest tissues and remove proteins cross-linked to DNA   |
| <b>Physical stress</b> | 1. Establishment of SOPs for sample storage conditions   |
|                        | 2. Establishment of SOPs for fixation procedures   |

**Table 4. Strategies for minimizing sequence artifacts in FFPET-DNA.**

## 4. Liquid biopsy

Direct testing of tumor tissue is used widely to detect gene mutations. However, methods used to determine mutation status in tumor tissue from patients with NSCLC have limitations (89). First, tumor samples are “contaminated” by normal tissue. Second, the biopsy may not represent the total burden of mutated cells, particularly in patients with metastatic disease. Third, the quality of extracted and stored DNA is variable; poor quality DNA may prevent accurate analyses. Also, patients with progressive disease cannot always be subjected to invasive procedures. In light of these limitations, new ways of monitoring tumor genetics and tumor dynamics have been developed. Recently, cell-free circulating tumor DNA (cfDNA and ctDNA) has attracted much attention as a potential tumor marker.

Analysis of “liquid biopsies” is a rapidly expanding field of cancer research because it can be useful for diagnosis and treatment of cancer patients, as well as its non-invasive and convenience. Analysis of liquid biopsies enables [a] early diagnosis, [b] estimation of the risk of cancer recurrence and metastasis (prognostic information), [c] observation and real-time monitoring of treatment responses, [d] better understanding of the therapeutic target and mechanism(s) of resistance, and [e] better understanding of metastasis (90).

However, it is unclear whether ctDNA in liquid biopsies can be used for non-invasive assessment of EGFR mutations in NSCLC patients. The first blood-based *EGFR* mutation analysis method to be developed was based on plasma obtained from pre-treatment NSCLC patients with gefitinib using the amplification refractory mutation system (ARMS)-based *EGFR* mutation detection kit. The *EGFR* mutation status derived from tumor tissue agreed with the results from matched plasma samples in 73% of cases (91). This result showed that it was possible to use ctDNA to detect *EGFR* mutations; other studies have also used various


serum- or plasma-based methods to assess *EGFR* mutation status (92-97). A previous paper reported the application of non-invasive procedures to monitor the *EGFR* T790M (98). Monitoring the mutation by using plasma in patients gives benefit to patients by suggesting proper treatment.

Because ctDNA analysis does not involve formaldehyde fixation, the number of false-positive results (caused by deamination) is low (99). Also, the amount of ctDNA in the plasma of healthy controls is much lower than that in NSCLC patients (100, 101). However, it is very difficult to analyze ctDNA from NSCLC patient due to the low amount of absolute ctDNA. The frequent fragmentation is another obstacle for analysis. It is known that double-stranded DNA fragmentations frequently occur in the serum or plasma of cancer patients (102). Also, high levels of DNA derived from tissues of non-tumor origin make detection of ctDNA technically difficult. Since ctDNA is derived directly from the tumor, it must be quantified accurately and with high sensitivity (103-108). Due to the varying quantity and quality of DNA derived from tumors, high sensitive and specialized equipment is required for ctDNA detection. Cutting-edge technologies can be used to detect somatic mutations, loss of heterozygotes, and mutation of tumor associated genes, at low frequencies as low as 0.01% (109-113). Although several methods can be used to analyze mutations in ctDNA [e.g., digital polymerase chain reaction (dPCR) (114), mutant-enriched PCR, and peptide nucleic acid-locked nucleic acid PCR], few have been approved for ctDNA analysis.


## **5. Third generation PCR: Droplet digital PCR (ddPCR)**

Real-time quantitative PCR (qPCR) is the most common method used to quantify expression of target genes (115, 116). However, either external calibrators or normalization to endogenous controls is required to estimate the concentration of an unknown (117). Imperfect amplification efficiency affects  $C_T$  values, which in turn limits the accuracy of this technique for absolute quantitation. ddPCR is an assay that combines state-of-the-art microfluidics technology with TaqMan-based PCR to achieve precise identification of target DNA with high sensitivity and specificity (Fig. 5, Table. 5) (118). This is because quantitation is achieved without the need for standard assays, allowing easy interpretation and an unambiguous digital readout. For ddPCR, a sample is fractionated into 20,000 droplets and PCR amplification of the template molecule(s) occurs in each individual droplet, effectively avoiding PCR bias, increasing the signal to noise ratio, removing the need for calibration or a reference, and providing absolute quantitation of a sample (117, 119). Hence, ddPCR is far simpler, faster, and less error prone than qPCR (120). Due to its technological advantages, which enable highly sensitive detection of mutations, the method has been adopted for clinical research (97, 119, 121-123). Notably, DNA integrity can also be assessed using ddPCR technology (119).

**STEP1: Prepare Sample and Partition into droplets**



**STEP2: Thermal cycle to Endpoint and Read droplets**



**STEP3: Apply Thresholds and Compute Concentrations**



Illustrated by Tae Won Yun

**Figure 5. Schematic showing the ddPCR workflow.** (A) Sample preparation and droplet generation. (B) Thermal cycling to endpoint and reading of each droplet. (C) Data analysis.

|                     | Real time-PCR   | ddPCR   |
|---------------------|---|---|
| <b>Advantage</b>    | <ol style="list-style-type: none"> <li>1. High throughput, routine experiments</li> <li>2. Relative quantification</li> <li>3. Short run time (~2 h / 96 samples)</li> <li>4. Low running cost</li> </ol> | <ol style="list-style-type: none"> <li>1. High Accuracy, Precision, Sensitivity<br/>(reduced error rate, 1.2-fold ↓)</li> <li>2. Relative and absolute quantification</li> <li>3. Simplified quantification<br/>: neither calibration standards nor a reference is required</li> <li>4. Removal of PCR bias</li> <li>5. Superior partitioning<br/>: 20,000 droplets per 20 µl → higher accuracy</li> <li>6. Sensitive detection of molecules with low-expression</li> </ol> |
| <b>Disadvantage</b> | <ol style="list-style-type: none"> <li>1. Dependent on C<sub>T</sub> value of calibration standards</li> <li>2. Inhibitor-dependent PCR efficiency</li> </ol>   | <ol style="list-style-type: none"> <li>1. Long run time (6 h / 96 samples)</li> <li>2. Higher running costs</li> <li>3. Requires attention compared to qPCR</li> </ol>  |

**Table 5. qPCR versus ddPCR.** Advantages and disadvantages of each assay.

## 6. Limit of blank, limit of detection and limit of quantitation

In the field of diagnostics, the ability to detect slight alterations in DNA is very important. This is known as sensitivity. Sensitivity is an important characteristic of any diagnostic tool because the ability to detect low level mutations increases the reliability of the assay. Sensitivity, limit of blank (LoB), limit of detection (LoD), and limit of quantitation (LoQ) are related by means of describing the lowest concentration of analytes (Fig. 6).

### 6.1 Limit of blank (LoB)

LoB is used when testing the samples that contain no analyte. It is a measure of a sample's expected maximum analyte concentration. Even though the samples lack analytes, the blank might contain low concentration of analytes. This is important when using very sensitive methods such as ddPCR. The LoB in ddPCR rules out false-positives. Therefore, it is very important to measure the LoB using standard guidelines.

LoB is estimated by blank sample's mean result and standard deviation (SD). The LoB is calculated according to:

$$\text{LoB} = \text{mean}_{\text{blank}} + 1.645 (\text{SD}_{\text{blank}})$$

### 6.2 Limit of detection (LoD)

LoD is the lowest analyte concentration and it is reliable. Opposite to LoB, which measures the highest concentration of analytes, LoD measures the lowest analyte concentration. LoD is an interval where analytes are detected while performing serial dilution. The LoD is


calculated by mean and SD of the low concentration of sample. The LoD is calculated according to:

$$\text{LoD} = \text{LoB} + 1.645 (\text{SD}_{\text{low concentration sample}})$$

### **6.3 Limit of quantitation (LoQ)**

LoQ is not only an analyte that can be reliably detected, but it is also the lowest concentration indicator for functional sensitivity. The LoQ value can be greater or equal to the LoD, even though the LoQ and LoD are calculated independently. In clinical trials, the LoQ indicates the minimal sensitivity to drug.

The Clinical and Laboratory Standards Institute (CLSI) have published standard guidelines for determining the LoB, LoD, and LoQ (CLSI document EP17. Wayne, PA USA: CLSI; 2004). It is important to characterize the analytical performance of diagnostics tests to understand their capabilities and limitations, and to ensure they are “fit for purpose”.



**Figure 6. Relationship between the LoB, LoD, and LoQ.** Distribution of replicates of a blank sample (solid black curve) and a specimen of low concentration (blue dashed curve). The LoB excludes a small proportion of blank results (“ $\alpha$ ”, Type I error). The LoD is then set such that only a small proportion (“ $\beta$ ”, Type II error) of these results will fall below the LoB. A sample measurement result exceeding the LoD threshold would represent quantitative detection of the measurand (analyte) (red dotted curve).

## **II. Purpose of the study**

Advances in science and technology have opened up new avenues in terms of molecular targeted therapy for cancer patients. For instance, several signal transduction pathways are activated in various cancers, which can be targeted by small molecules. Mutation analysis is a key factor for stratifying cancer patients according to the likely benefit of molecular targeted therapy. Therefore, accurate detection of mutations that may aid/retard cancer therapy is essential.

Imprecise mutation analysis caused by DNA damage is a very serious problem in many diagnostic fields. In particular, DNA fragmentation reduces the amount of amplifiable DNA in a tissue; this makes estimation of the mutation frequency value (based on the amount of input DNA) less accurate. FFPET is the most widely used resource for mutation analysis. But, artifactual sequence variants arising from DNA damage will be detected more frequently because stochastic enrichment (in the context of low copy targets) increases the risk of false-positives. DNA extracted from FFPET is subject to extensive fragmentation, which reduces data quality. None of the traditional tools used to check DNA integrity (e.g., UV absorbance or gel electrophoresis) actually measure the amount of amplifiable DNA or the amount of DNA fragmentation. Therefore, information on the PCR amplifiable templates will enable determining how reliable the accuracy of detected gene mutation.

Here, we report development of sample criteria to ensure minimum FFPET DNA quality for PCR. We then apply these criteria to a ddPCR-based EGFR mutation test. To establish the sample criteria, we first collected and analyzed 316 NSCLC FFPET DNA samples of various age and quality from three different sites (hospitals). We also conducted independent retrospective clinical studies using 228 NSCLC FFPET samples to verify the clinical implications of the established sample criteria. In addition, we compared the performance of the GenesWell™ ddEGFR mutation test (ddEGFR test) with that of the Cobas® EGFR mutation


test (Cobas EGFR test) in a pre-clinical and clinical studies based on 544 NSCLC FFPE samples.

The major study objectives were 1) to establish sample criteria to determine the minimum DNA quality suitable for PCR, 2) to compare the clinical performances of the ddEGFR and cobas EGFR tests.


### **III. Materials and Methods**

## 1. Study design

To establish sample criteria, a total of 316 samples obtained from NSCLC patients were tested for *EGFR* mutations. A post-hoc analysis of these pre-clinical data was conducted for both the ddEGFR and cobas EGFR test results (Fig. 7). Based on the established sample criteria, an independent retrospective comparison study was performed to estimate the concordance between the ddEGFR and cobas EGFR tests; for this purpose, 228 FFPE-DNA samples from NSCLC patients were analyzed by both tests (Fig. 8). Both EGFR mutation tests were performed in a double-blind fashion by an independent laboratory (Abion Inc., Seoul, Korea). The study design (workflow) is depicted in Figure 7, 8.



**Figure 7. Study design and specimen selection (work flow of pre-clinical study).** Patient sample dispositions for the applied sample criteria. For the pre-clinical study group, a total of 316 FFPE specimens were subjected to post-hoc analysis.



**Figure 8. Study design and specimen selection (work flow of retrospective comparison study).** Patient sample dispositions for the applied iQC index. For the retrospective comparison study group, a total of 228 FFPE specimens were analyzed (D/O, Drop Out).

## **2. FFPET collection**

FFPET blocks of resected or biopsy samples from NSCLC patients ( $n = 316$ ) collected from 2005 to 2014 were retrieved from the Department of Pathology, Samsung Medical Center ( $n = 200$ ) (SMC; Seoul, Korea), Asan Medical Center ( $n = 66$ ) (AMC; Seoul, Korea), and Severance Hospital ( $n = 50$ ) (Seoul, Korea). This study was approved by the Institutional Review Board (IRB) of SMC and Seoul National University (study ID: SMC-2014-05-084-002). For the retrospective comparison study, 228 archived FFPET blocks from NSCLC patients collected between 2010 and 2016 were obtained from Department of Pathology, SMC. This study was approved by the IRB of SMC and the Ministry of Food and Drug Safety (MFDS) of Korea (study ID: SMC-2016-07-104-002). From each FFPET, 10  $\mu\text{m}$  sections were cut and subjected to DNA extraction. H&E-stained sections containing tumor lesions marked by a pathologist (S.W.C.) were scanned, and the cancer/normal (C/N) ratio was calculated using the Panoramic Viewer Software v.1.15.4 (3DHISTECH, Budapest, Hungary). Patient information was anonymized and de-identified prior to analysis.

## **3. DNA extraction and determination of DNA quantity and quality**

DNA extraction from FFPEs was performed using an automated Tissue Preparation System (TPS; Siemens Healthcare, Erlangen, Germany) with the VERSANT<sup>®</sup> Tissue Preparation Reagents, as described previously (124). Total nucleic acids were eluted with 100  $\mu\text{L}$  elution buffer containing UDG provided by the manufacturer. For all samples, DNA concentration was assessed using the Qubit<sup>™</sup> 3.0 Fluorometer and Qubit dsDNA BR assay kit (ThermoFisher Scientific, MA, USA). The DNA integrity number (DIN), reflecting the DNA fragmentation level (125) of genomic DNA (gDNA), was analyzed on a 2200 TapeStation system with

Genomic DNA Screen Tape (Agilent Technologies, CA, USA). The samples were prepared by mixing 1  $\mu$ l of DNA sample with 10  $\mu$ l of genomic DNA sample buffer. A genomic DNA ladder was placed in the first lane of 8-strip tube followed by the samples. The prepared strip tube was mixed and placed in the 2200 TapeStation instrument. Samples were analyzed using TapeStation Analysis v.01.15 Software.

#### **4. Validation of internal quality control (iQC) of ddEGFR test**

The ddEGFR test (Gencurix Inc., Seoul, Korea) was designed as a highly sensitive ddPCR-based diagnostic test for detecting 45 mutation sites within the exon 18–21 region of the *EGFR* gene using four reactions. The amplified fragments, which contain the fluorophores FAM<sup>TM</sup> or HEX<sup>TM</sup>, are displayed as dots (droplets) and can be used to calculate concentrations (copies/20  $\mu$ L) based on the Poisson distribution (118). The details of specific mutations detected by the assay are provided in Table 6 and Table 7.

| No. | Target exon | COSMIC No. * | Target mutant A.a   | Target mutant site          | Result |                          |        |
|-----|-------------|--------------|---------------------|-----------------------------|--------|--------------------------|--------|
| 1   | Exon18      | 6239         | p.G719A             | c.2156G>C                   | G719X  |                          |        |
| 2   |             | 6252         | p.G719S             | c.2155G>A                   |        |                          |        |
| 3   |             | 6253         | p.G719C             | c.2155G>T                   |        |                          |        |
| 4   | Exon19      | 6223         | p.E746_A750del      | c.2235_2249 del 15          | 19del  |                          |        |
| 5   |             | 13551        | p.E746_T751>I       | c.2235_2252 > AAT           |        |                          |        |
| 6   |             | 12728        | p.E746_T751del      | c.2236_2253 del 18          |        |                          |        |
| 7   |             | 12678        | p.E746_T751>A       | c.2237_2251 del 15          |        |                          |        |
| 8   |             | 12367        | p.E746_S752>A       | c.2237_2254 del 18          |        |                          |        |
| 9   |             | 12384        | p.E746_S752>V       | c.2237_2255>T               |        |                          |        |
| 10  |             | 6225         | p.E746_A750del      | c.2236_2250 del 15          |        |                          |        |
| 11  |             | 6220         | p.E746_S752>D       | c.2238_2255 del 18          |        |                          |        |
| 12  |             | 13550        | p.E746_A750>IP      | c.2235_2248>AATTC           |        |                          |        |
| 13  |             | 12403        | p.L747_S752>Q       | c.2239_2256>CAA             |        |                          |        |
| 14  |             | 12422        | p.L747_A750>P       | c.2238_2248 >GC             |        |                          |        |
| 15  |             | 12419        | p.L747_T751>Q       | c.2238_2252 >GCA            |        |                          |        |
| 16  |             | 6218         | p.L747_E749del      | c.2239_2247 del 9           |        |                          |        |
| 17  |             | 12382        | p.L747_ A750>P      | c.2239_2248 TTAAGAGAAG>C    |        |                          |        |
| 18  |             | 6210         | p.L747_T751>S       | c.2240_2251 del 12          |        |                          |        |
| 19  |             | 12383        | p.L747_T751>P       | c.2239_2251>C               |        |                          |        |
| 20  |             | 13552        | p.E746_T751>IP      | c.2235_2251>AATTC           |        |                          |        |
| 21  |             | 6254         | p.L747_T751del      | c.2239_2253 del 15          |        |                          |        |
| 22  |             | 6255         | p.L747_S752del      | c.2239_2256 del 18          |        |                          |        |
| 23  |             | 12387        | p.L747_P753>Q       | c.2239_2258 >CA             |        |                          |        |
| 24  |             | 12370        | p.L747_P753>S       | c.2240_2257 del 18          |        |                          |        |
| 25  |             | 12416        | p.E746_T751>VA      | c.2237_2253>TTGCT           |        |                          |        |
| 26  |             | -            | -                   | c.2239_2257>GT              |        |                          |        |
| 27  |             | 26038        | p.K745_E749del      | c.2233_2247del15            |        |                          |        |
| 28  |             | 13556        | p.S752_I759del      | c.2253_2276del124           |        |                          |        |
| 29  |             | 12386        | p.E746_T751>V       | c.2237_2252>T               |        |                          |        |
| 30  |             | 12385        | p.E746_S752>I       | c.2235_2255>AAT             |        |                          |        |
| 31  |             | 18427        | p.E746_P753>VS      | c.2237_2257>TCT             |        |                          |        |
| 32  |             | 12369        | p.L747_T751del      | c.2240_2254 del 15          |        |                          |        |
| 33  |             | 23571        | p.L747_T751delLREAT | c.2238_2252 del 15          |        |                          |        |
| 34  |             | Exon20       | 12376               | p.V769_D770insASV           |        | c.2307_2308 insGCCAGCGTG | E20Ins |
| 35  |             |              | 12377               | p.H773_V774insH             |        | c.2319_2320 insCAC       |        |
| 36  |             |              | 12378               | p.D770_N771insG             |        | c.2310_2311 insGGT       |        |
| 37  | 13428       |              | p.D770_N771insSVD   | c.2311_2312insGCGTGGA CA    |        |                          |        |
| 38  | 13558       |              | p.V769_D770insASV   | c.2309_2310A C>CCA GCGTGGAT |        |                          |        |
| 39  | Exon20      | 6241         | p.S768I             | c.2303G>T                   | S768I  |                          |        |
| 40  |             | 6240         | p.T790M             | c.2369C>T                   | T790M  |                          |        |
| 41  |             | -            | p.C797S             | c.2389T>A                   | C797S  |                          |        |

|    |               |       |         |                  |              |
|----|---------------|-------|---------|------------------|--------------|
| 42 |               | -     | p.C797S | c.2390G>C        |              |
| 43 | <b>Exon21</b> | 6224  | p.L858R | c.2573T>G        | <b>L858R</b> |
| 44 |               | 12429 | p.L858R | c.2573_2574TG>GT |              |
| 45 |               | 6213  | p.L861Q | c.2582T>A        | <b>L861Q</b> |

\* Catalogue of Somatic Mutations in Cancer (COSMIC). 2016, v.78.

<http://cancer.sanger.ac.uk/cosmic>

**Table 6. Forty-five *EGFR* mutations detected by the ddEGFR mutation test.**

The non-clinical performance studies followed the guidelines approved by the Clinical and Laboratory Standards Institute (CLSI) and the Korea-MFDS. For validation of internal quality control of ddEGFR, FFPE reference standard DNA extracts (HDx™ Reference Standard, Horizon Discovery, Cambridge, UK) for EGFR mutations were blended with fixed amounts of wild-type gDNA (3.3 ng, 1000 GE; Promega) and each sample, with a target MI of 1.5%. In addition, four serial dilutions of each sample (9.9 ng, 6.6 ng, 3.3 ng, and 1.65 ng) were prepared and analyzed using the ddEGFR test. The iQC copies and target MI of each sample was confirmed based on the input DNA concentration and target MI (1.5%).

## **5. Biomarker analysis**

The ddEGFR test was performed in a 20 µL volume containing 3.3 ng (1000 GE)/reaction of template DNA on a Droplet Digital™ PCR (ddPCR) system (Bio-Rad, Hercules, CA, USA). The ddPCR assay was conducted as described previously (124). Amplification conditions consisted of a 10 min activation period at 95 °C, followed by 40 cycles of a 30 sec at 94 °C for denaturation, 1 min at 60 °C for annealing/extension, and a final 10 min inactivation step at 98 °C. After thermal cycling, plates were transferred to a QX200 droplet reader to read the droplets using the QuantaSoft v1.7.4 software (Bio-Rad) to assess the number of droplet positive for *EGFR*. Thresholds for detection were set manually based on results from non-template control wells and negative control wells containing wild-type gDNA (Promega).

PCR amplification for the cobas EGFR test (Roche Molecular Systems Inc., Branchburg, NJ, USA) was performed on a cobas® z 480 Analyzer. The cobas EGFR test requires 150 ng total input DNA. The mutation result reporting is fully automated, mutation was reported as either G719X in exon 18, 19del in exon 19, S768I in exon 20, T790M in exon 20, exon 20

insertion, or L858R in exon 21 (126). Both mutation tests were analyzed in a double-blind fashion, and the results were matched after analysis.

For mutation screening of *EGFR* exons 18, 19, 20, and 21 by 2× bidirectional Sanger sequencing, regions of interest were amplified by PCR, and the amplified samples were processed at an independent laboratory (Macrogen, Seoul, Korea) using a validated protocol. Sanger sequencing results were cross-checked and interpreted by a pathologist (Y.L.C).

## **6. Methods correlation and statistical analysis**

Positive percent agreement (PPA), negative percent agreement (NPA), and overall percent agreement (OPA) of the ddEGFR and cobas EGFR tests were compared in pre-clinical and retrospective comparison studies. Agreement analysis for all methods was based on mutation report calls (Table. 7). Statistical analysis was performed using GraphPad Prism™ (GraphPad Software Inc., San Diego, USA) and the R 1.6.12 package ‘psych’ (<http://CRAN.R-project.org/package=psych>). For the agreement analyses, PPA, NPA, and OPA were calculated with their corresponding 95% confidence intervals (CIs).

## **IV. Results**

## 1. Validation of internal quality control (iQC) of ddEGFR test

Because the ddPCR-based test can lead to the false positive results by both its intrinsic high sensitivity and FFPE characteristics, the cut-offs of ddEGFR test were determined based on false-positive analyses using normal FFPE. Mutation calls were identified based on true-positive mutation values higher than limit of blank (LoB) and limit of detection (LoD) mutation index (MI) [1], which were established by analytical performance studies (Table. 7). MI is a numerical value representing the ratio of mutant to internal quality control (iQC) copies, calculated as follows:

$$\text{MI (\%)} = [\text{Mutant copies} / \text{iQC copies}] \times 100 \quad [1]$$

In the ddEGFR test, iQC copies can be converted to concentration of input DNA using an FFPE reference standard, suggesting that iQC index [2] is a representative index of amplifiable DNA. Because iQC copies were analyzed using 3.3 ng (1,000 genomic equivalents [GE]) of input DNA per reaction well, iQC index was calculated as follows:

$$\text{iQC index} = [\text{iQC copies} / \text{input DNA copies}] \quad [2]$$

An iQC index value close to “1” means that the quality of DNA is better. Figure 9 shows the concept and example of iQC index.

Using 40 wild-type FFPE samples, false-positive rates were determined for mutant calling of eight targets. The maximum number of copies was 5.4 per reaction, and the false-positive rates were below 0.5% of MI (Fig. 10).

| Exon | Mutations detected  | Mutation report call | LoB copies / MI (%) * | LoD MI (%) † |
|------|---------------------|----------------------|-----------------------|--------------|
| 18   | G719A, G719C, G719S | G719X                | 5.6 / 0.22            | 0.77         |
| 19   | 30 deletions        | 19del                | 3.0 / 0.09            | 0.83         |
| 20   | S768I               | S768I                | 1.5 / 0.05            | 0.83         |
|      | T790M               | T790M                | 6.8 / 0.34            | 0.78         |
|      | C797S §             | C797S                | 1.6 / 0.03            | 0.75         |
|      | 5 Insertions        | E20Ins               | 1.6 / 0.06            | 0.62         |
| 21   | L858R               | L858R                | 1.6 / 0.03            | 0.71         |
|      | L861Q               | L861Q                | 1.4 / 0.05            | 0.74         |


\* LoB =  $\text{mean}_{\text{blank}} + 1.645 (\text{SD}_{\text{blank}})$  (127)

† LoD MI (%) =  $\text{mean}_{\text{low concentration sample}} + 1.645 (\text{SD}_{\text{low concentration sample}})$

§ using the plasmids containing the non-predominant mutation for analytical performance

**Table 7. ddEGFR test coverage.** The test is designed to detect G719X (G719A, G719C, and G719S) mutations in exon 18, deletions, complex mutations (S768I, T790M, and C797S) in exon 19, insertions in exon 20, and the L858R and L861Q mutations in exon 21. The LoB were calculated using 9.9 ng of normal FFPET-DNA. The LoD was measured 6 points (3-0.1%) serially diluted using the NSCLC FFPET-DNA (19del, L858R) and reference standard FFPET-DNA (the other mutations) with wild-type FFPET-DNA. The LoD was determined by the lowest amount of DNA that gave an *EGFR* “Mutation Detected” rate of at least 95% for the target mutation.







\* Mutation Index (MI %) = [Mutation copies / iQC copies] x 100

**Figure 10. Determination of a suitable cut-off based on false-positive analysis using normal FFPE blocks.** Forty normal FFPE specimens were tested, and eight targets were evaluated for each specimen. The maximum number of copies was 5.4 per reaction (red dotted line) and the false-positive rate was below 0.5%.

We evaluated iQC using reference standard. Reference standard FFPET-DNA extract for *EGFR* mutations was blended with fixed amount of wild-type gDNA (3.3 ng, 1,000GE) targeting a 1.5% mutation level. The expected iQC index and MI were given by the calculated from the quantity of input DNA. As a result, the measured MI (%) and iQC index were observed close to match with expected value (iQC index = 1, MI = 1.5%) (Fig. 11). Moreover, iQC was also validated by serial dilution with four concentrations of reference standard FFPET-DNA, and observed that the measured values close match with the each expected values (Fig. 12) suggesting that iQC copies can represent the amount of input DNA.



**Figure 11. Validation of internal quality control of ddEGFR test (MI validation).** Internal quality control (iQC) validation using reference standard. Each FFPE reference standard DNA extract for the *EGFR* mutations was blended with a fixed amount of wild-type gDNA (3.3 ng, 1000GE) targeting a 1.5% mutation level, which was validated for use in the ddEGFR test. Values are expressed as the mean  $\pm$  SD of nine experiments.



**Figure 12. Validation of internal quality control of ddEGFR test (iQC copy and iQC index).** Internal quality control (iQC) validation using the reference standard. Four serial dilutions of each FFPET reference standard DNA extract were prepared and subjected to the ddEGFR test. Values are expressed as the mean  $\pm$  SD of three experiments.

## **2. Comparison of ddEGFR and cobas EGFR tests without sample criteria**

The *EGFR* mutations in 316 NSCLC FFPET samples were analyzed using the ddEGFR and cobas tests. Both methods yielded valid results for all but one of the samples. Surprisingly, the ddEGFR and cobas EGFR tests were low concordant (PPA = 94.04%, NPA = 63.41%, OPA = 78.10%, kappa coefficient value ( $\kappa$ ) = 0.6650) (Table. 8, Table. 9). The cobas EGFR test also exhibited very low concordance with Sanger sequencing of 299 samples (PPA = 59.30%, NPA = 75.00%, OPA = 65.63%,  $\kappa$  = 0.4526) (Table. 10, Table. 11).

| All Samples<br>(n = 316) |       | cobas EGFR Test       |      |       |
|--------------------------|-------|-----------------------|------|-------|
|                          |       | MD                    | MND  | Total |
| ddEGFR<br>Test           | MD    | 139                   | 62 * | 201   |
|                          | MND   | 9                     | 104  | 113   |
|                          | Total | 148                   | 166  | 314   |
| PPA (95% C.I.)           |       | 94.04% (88.99–97.24%) |      |       |
| NPA (95% C.I.)           |       | 63.41% (55.55–70.79%) |      |       |
| OPA (95% C.I.)           |       | 78.10% (73.12–82.54%) |      |       |
| PPV (95% C.I.)           |       | 70.30% (63.48–76.51%) |      |       |
| NPV (95% C.I.)           |       | 92.04% (85.42–96.29%) |      |       |

\* 3 samples: cobas, 19del; ddEGFR, 19del/T790M

1 sample: cobas, 19del; ddEGFR, 19del/L858R

1 sample: cobas, E20Ins; ddEGFR, E20Ins/T790M

1 sample: cobas, S768I; ddEGFR, S768I/L858R

**Table 8. A total of 316 samples with valid ddEGFR and cobas EGFR test results were included in the agreement analysis. PPA, NPA, and OPA between the ddEGFR and cobas EGFR tests for the detection of *EGFR* mutations were 94.04%, 63.41%, and 78.10%, respectively.**

| All Samples<br>(n = 316) |              | cobas EGFR Test |       |       |        |       |       |                 |                 |                 |                 |                 |                  |                 |     |       |
|--------------------------|--------------|-----------------|-------|-------|--------|-------|-------|-----------------|-----------------|-----------------|-----------------|-----------------|------------------|-----------------|-----|-------|
|                          |              | G719X           | 19del | T790M | E20Ins | S768I | L858R | G719X,<br>T790M | G719X,<br>S768I | 19del,<br>T790M | 19del,<br>L858R | T790M,<br>L858R | T790M,<br>E20Ins | S768I,<br>L858R | MND | Total |
| ddEGFR<br>Test           | G719X        | 1               |       |       |        |       |       |                 |                 |                 |                 |                 |                  |                 | 1   | 2     |
|                          | 19del        |                 | 85    |       |        |       |       |                 |                 |                 |                 |                 |                  |                 | 4   | 89    |
|                          | T790M        |                 |       | 0     |        |       |       |                 |                 |                 |                 |                 |                  |                 |     | 0     |
|                          | E20Ins       |                 |       |       | 0      |       |       |                 |                 |                 |                 |                 |                  |                 | 2   | 2     |
|                          | S768I        |                 |       |       |        | 0     |       |                 |                 |                 |                 |                 |                  |                 |     | 0     |
|                          | L858R        |                 | 1 †   |       |        |       | 48    |                 |                 |                 |                 |                 |                  |                 | 47  | 96    |
|                          | G719X,T790M  |                 |       |       |        |       |       | 0               |                 |                 |                 |                 |                  |                 | 1   | 1     |
|                          | G719X,S768I  |                 |       |       |        |       |       |                 | 3               |                 |                 |                 |                  |                 |     | 3     |
|                          | 19del,T790M  |                 | 3     |       |        |       |       |                 |                 | 0               |                 |                 |                  |                 |     | 3     |
|                          | 19del,L858R  |                 | 1     |       |        |       |       |                 |                 |                 | 0               |                 |                  |                 |     | 1     |
|                          | T790M,L858R  |                 |       |       |        |       |       |                 |                 |                 |                 | 0               |                  |                 | 1   | 1     |
|                          | T790M,E20Ins |                 |       |       | 1      |       |       |                 |                 |                 |                 |                 | 0                |                 |     | 1     |
|                          | S768I, L858R |                 |       |       |        | 1     |       |                 |                 |                 |                 |                 |                  | 2               |     | 3     |
|                          | MND          |                 | 9     |       |        |       |       |                 |                 |                 |                 |                 |                  |                 | 104 | 113   |
| Total                    | 1            | 99              | 0     | 1     | 1      | 48    | 0     | 3               | 0               | 0               | 0               | 0               | 2                | 160             | 315 |       |

† 1 sample was excluded from the analysis, 1 sample: Invalid

$\kappa$  coefficient = 0.6650 (95% C.I. 59.98–73.03%)

**Table 9.** Detail concordance table of 316 samples with valid ddEGFR and cobas EGFR test.

| All samples<br>(n = 299) |       | Sanger sequencing     |      |       |
|--------------------------|-------|-----------------------|------|-------|
|                          |       | MD                    | MND  | Total |
| cobas<br>EGFR test       | MD    | 102                   | 29 * | 131   |
|                          | MND   | 70 †                  | 87   | 157   |
|                          | Total | 172                   | 116  | 288   |
| PPA (95% C.I.)           |       | 59.30% (61.56–66.72%) |      |       |
| NPA (95% C.I.)           |       | 75.00% (66.11–82.57%) |      |       |
| OPA (95% C.I.)           |       | 65.63% (59.83–71.10%) |      |       |
| PPV (95% C.I.)           |       | 77.86% (69.78–84.65%) |      |       |
| NPV (95% C.I.)           |       | 55.41% (47.28–63.34%) |      |       |

\* 1 sample: Sanger, L858R; cobas, L858R, S768I

† 4 samples: Sanger, 19del, L858R; cobas, 19del

1 sample: Sanger, 19del, L858R; cobas, L858R

**Table 10. Concordance of *EGFR* mutation detection between the cobas EGFR test and Sanger sequencing, without application of exclusion criteria.** A total of 299 samples with valid Sanger sequencing and cobas EGFR test results were included in the analysis. Seventeen samples could not be evaluated for concordance because Sanger sequencing did not yield results.

| All samples<br>(n = 299) |              | Sanger Sequencing |       |       |        |       |       |             |                 |                 |                 |                 |                 |                 |     |       |
|--------------------------|--------------|-------------------|-------|-------|--------|-------|-------|-------------|-----------------|-----------------|-----------------|-----------------|-----------------|-----------------|-----|-------|
|                          |              | G719X             | 19del | T790M | E20Ins | S768I | L858R | 19del,G719X | 19del,<br>T790M | 19del,<br>L858R | G719X,<br>T790M | G719X,<br>S768I | T790M,<br>L858R | S768I,<br>L858R | MND | Total |
| cobas<br>EGFR<br>test    | G719X        | 0                 |       |       |        |       |       |             |                 |                 |                 |                 |                 |                 | 1   | 1     |
|                          | 19del        |                   | 62    |       |        |       | 8 §   |             |                 | 4               |                 |                 |                 |                 | 18  | 92    |
|                          | T790M        |                   |       | 0     |        |       |       |             |                 |                 |                 |                 |                 |                 |     | 0     |
|                          | E20Ins       |                   |       |       | 0      |       |       |             |                 |                 |                 |                 |                 |                 | 1   | 1     |
|                          | S768I        |                   |       |       |        | 0     |       |             |                 |                 |                 |                 |                 |                 |     | 0     |
|                          | L858R        |                   | 2 §   |       |        |       | 40    |             |                 | 1               |                 |                 |                 |                 | 5   | 48    |
|                          | 19del,G719X  |                   |       |       |        |       |       | 0           |                 |                 |                 |                 |                 |                 |     | 0     |
|                          | 19del, T790M |                   |       |       |        |       |       |             | 0               |                 |                 |                 |                 |                 |     | 0     |
|                          | 19del, L858R |                   |       |       |        |       |       |             |                 | 0               |                 |                 |                 |                 |     | 0     |
|                          | G719X,T790M  |                   |       |       |        |       |       |             |                 |                 | 0               |                 |                 |                 |     | 0     |
|                          | G719X,S768I  |                   | 1 §   |       |        |       |       |             |                 |                 |                 | 0               |                 |                 | 2   | 3     |
|                          | T790M,L858R  |                   |       |       |        |       |       |             |                 |                 |                 |                 | 0               |                 |     | 0     |
|                          | S768I, L858R |                   |       |       |        |       | 1     |             |                 |                 |                 |                 |                 | 0               | 1   | 2     |
|                          | MND          |                   | 24    |       |        |       | 27    | 1           |                 | 13              |                 |                 |                 |                 | 87  | 152   |
| Total                    | 0            | 89                | 0     | 0     | 0      | 76    | 1     | 0           | 18              | 0               | 0               | 0               | 0               | 115             | 299 |       |


§ 11 samples were excluded from the analysis

$\kappa$  coefficient = 0.4526 (95% C.I. 37.45–53.07%)

**Table 11. Detail concordance of *EGFR* mutation detection between the cobas EGFR test and Sanger sequencing, without application of exclusion criteria.**

### **3. Proof-of-concept for determination of minimum DNA quality suitable for PCR using the ddPCR method**


To increase the concordance between the ddEGFR and cobas EGFR tests, we investigated the minimum DNA quality suitable for PCR analysis by re-analyzing the ddEGFR data, which provided iQC copies and index (Fig. 11, 12). FFPET block storage duration was reflected in the amount of amplifiable DNA, decreasing below 50% (mean of iQC index = 0.31, Standard deviation, SD = 0.57) of 1000 GE in 7–11-year-old samples. By contrast, in 2–6-year-old samples, the amount of amplifiable DNA was around 100% (mean of iQC index = 1.07, SD = 0.69) (Fig. 13). Therefore, iQC index decreased with storage duration when FFPET was stored at room temperature. The pattern of DIN values from 315 FFPET-DNA samples measured in parallel was similar to those of iQC copies and index (Fig. 14).



**Figure 13. Distributions of ddEGFR iQC index corresponding to sample storage time.**

The black line indicates the median value (\*Genomic Equivalence; \*\*\* $p < 0.0001$ ).

$n = 315$




**Figure 14. Distributions of DIN value corresponding to sample storage time.** The black line indicates the median value (\*Genomic Equivalence; \*\*\* $p. < 0.0001$ ).

#### **4. Establishment of iQC index for ddEGFR using FFPET samples**

We classified all 316 samples into four groups according to their storage durations, DIN values, and iQC copies, as illustrated in Figure 4A. First, we analyzed DIN values for 57 discordant samples from among 169 samples derived from blocks stored for more than 6 years. The DIN value was less than 2.5 for most of the discordant samples (Fig. 15). Among 147 samples with a block storage duration  $\leq 6$  years, 26 samples did not satisfy our DIN criteria, and were included in Group 1 (Table 12, Table 13). To establish iQC index criteria, the 60 discordant samples in Group 1 were re-analyzed, and almost all (58/60) had an iQC index  $< 0.5$  (Fig. 16).

Based on this result, the sample criteria were established as follows: block storage duration  $\leq 6$  years,  $DIN > 2.5$ , and  $iQC\ index \geq 0.5$ . In addition, we observed a strong correlation between ddEGFR iQC index and DIN value, further supporting the idea that iQC index represents the quality of FFPET-DNA ( $p. < 0.0001$ , Table. 14).



**Figure 15. Establishment of sample criteria using discordant samples (DIN value).** Discordant samples were analyzed by comparing the results of the Cobas EGFR and ddEGFR tests. Plots show distributions of the correlation between DIN value and sample storage duration of discordant samples. The black line represents the median value. The DIN values of discordant samples are distributed under the black dotted line.



| Group 1 ( <i>n</i> = 194) |       | cobas EGFR Test           |      |       |
|---------------------------|-------|---------------------------|------|-------|
|                           |       | MD                        | MND  | Total |
| ddEGFR<br>Test            | MD    | 49                        | 56 * | 105   |
|                           | MND   | 9                         | 79   | 88    |
|                           | Total | 58                        | 135  | 193   |
| PPA (95% C.I.)            |       | 84.48% (72.58% to 92.65%) |      |       |
| NPA (95% C.I.)            |       | 58.52% (49.73% to 66.93%) |      |       |
| OPA (95% C.I.)            |       | 66.32% (59.18% to 72.95%) |      |       |
| PPV (95% C.I.)            |       | 46.67% (36.87% to 56.66%) |      |       |
| NPV (95% C.I.)            |       | 89.77% (81.47% to 95.22%) |      |       |

**Table 12. Comparison of ddEGFR and cobas EGFR results from FFPE samples (block storage duration > 6, DIN < 2.5).** PPA, NPA, and OPA between the ddEGFR and cobas EGFR tests for the detection of *EGFR* mutations were 84.48%, 58.52%, and 66.32%, respectively. MD, mutation detected; MND, mutation not detected.

| Group 1<br>(n = 194)       |              | Cobas EGFR Test |       |       |        |       |       |                 |                 |                 |                 |                 |                  |                 |     |       |
|----------------------------|--------------|-----------------|-------|-------|--------|-------|-------|-----------------|-----------------|-----------------|-----------------|-----------------|------------------|-----------------|-----|-------|
|                            |              | G719X           | 19del | T790M | E20Ins | S768I | L858R | G719X,<br>T790M | G719X,<br>S768I | 19del,<br>T790M | 19del,<br>L858R | T790M,<br>L858R | T790M,<br>E20Ins | S768I,<br>L858R | N/A | Total |
| ddEGFR<br>Mutation<br>Test | G719X        | 0               |       |       |        |       |       |                 |                 |                 |                 |                 |                  |                 |     | 0     |
|                            | 19del        |                 | 33    |       |        |       |       |                 |                 |                 |                 |                 |                  |                 | 3   | 36    |
|                            | T790M        |                 |       | 0     |        |       |       |                 |                 |                 |                 |                 |                  |                 |     | 0     |
|                            | E20Ins       |                 |       |       | 0      |       |       |                 |                 |                 |                 |                 |                  |                 | 1   | 1     |
|                            | S768I        |                 |       |       |        | 0     |       |                 |                 |                 |                 |                 |                  |                 |     | 0     |
|                            | L858R        |                 | 1 †   |       |        |       | 14    |                 |                 |                 |                 |                 |                  |                 | 44  | 59    |
|                            | G719X,T790M  |                 |       |       |        |       |       | 0               |                 |                 |                 |                 |                  |                 | 1   | 1     |
|                            | G719X,S768I  |                 |       |       |        |       |       |                 | 1               |                 |                 |                 |                  |                 |     | 1     |
|                            | 19del,T790M  |                 | 3     |       |        |       |       |                 |                 | 0               |                 |                 |                  |                 |     | 3     |
|                            | 19del,L858R  |                 | 1     |       |        |       |       |                 |                 |                 | 0               |                 |                  |                 |     | 1     |
|                            | T790M,L858R  |                 |       |       |        |       |       |                 |                 |                 |                 | 0               |                  |                 | 1   | 1     |
|                            | T790M,E20Ins |                 |       |       | 1      |       |       |                 |                 |                 |                 |                 | 0                |                 |     | 1     |
|                            | S768I, L858R |                 |       |       |        | 1     |       |                 |                 |                 |                 |                 |                  | 1               |     | 2     |
|                            | N/A          |                 | 9     |       |        |       |       |                 |                 |                 |                 |                 |                  |                 | 79  | 88    |
| <b>Total</b>               | 0            | 47              | 0     | 1     | 1      | 14    | 0     | 1               | 0               | 0               | 0               | 0               | 1                | 129             | 194 |       |

$\kappa$  coefficient = **0.4611** (95% C.I. 36.56–55.66%)

**Table 13. Detail concordance rate of group 1 samples (block storage duration > 6, DIN < 2.5).**

|   | iQC index \ DIN       | Bad (< 2.5) | Good (> 2.5) |
|---|-----------------------|-------------|--------------|
| <b>Group 1 (n = 194)</b><br><i>(p. &lt; 0.0001)</i> | <b>Low (&lt; 0.5)</b> | 155         | 2            |
|   | <b>High (≥ 0.5)</b>   | 22          | 15           |
| <b>Group 2 (n = 113)</b>                            | <b>Low (&lt; 0.5)</b> | 0           | 8            |
|   | <b>High (≥ 0.5)</b>   | 0           | 113          |
| <b>Group 3 (n = 150)</b>                            | <b>Low (&lt; 0.5)</b> | 0           | 0            |
|   | <b>High (≥ 0.5)</b>   | 22          | 128          |
| <b>Group 4 (n = 166)</b>                            | <b>Low (&lt; 0.5)</b> | 156         | 10           |
|   | <b>High (≥ 0.5)</b>   | 0           | 0            |

**Table 14. Correlation between ddEGFR iQC index and DIN value.** The ddEGFR iQC index and DIN value of excluded samples were lower than the values of the sample criteria. Results from included samples were opposite to those of the excluded samples. In addition, the ddEGFR iQC index and DIN values were strongly correlated.

## 5. Comparison of ddEGFR and cobas EGFR tests with iQC index criteria

When the iQC index cut-off was applied to 121 samples (block storage duration  $\leq 6$  years, DIN  $> 2.5$ ), 113 samples remained (Fig. 7, Group 2). Group 2 samples had a very high concordance rate between the ddEGFR and cobas EGFR tests (PPA = 100.00%, NPA = 76.00%, OPA = 94.69%,  $\kappa = 0.9197$ ) (Table. 15, Table. 16). To determine the clinical implications of the iQC index, we applied this criterion to 316 FFPETs, resulting in the classification of 150 samples into Group 3 (Fig. 7), and re-analyzed the concordance rate between the ddEGFR and cobas EGFR results. Group 3 samples had a very high concordance rate (PPA = 100.00%, NPA = 75.00%, OPA = 92.67%,  $\kappa = 0.8923$ ) (Table. 17, Table. 18), similar to the results from Group 2 (applying all criteria). By contrast, Group 4 samples, in which the iQC index criterion was not satisfied, exhibited a very low concordance rate (PPA = 78.57%, NPA = 58.20%, OPA = 63.41%,  $\kappa = 0.3862$ ) (Table. 19, Table. 20). Therefore, we suggested that iQC index is a key factor in determining whether DNA is of sufficient quality for the ddEGFR test.

| <b>Group 2</b><br><i>(n = 113)</i> |              | <b>cobas EGFR Test</b> |            |              |
|------------------------------------|--------------|------------------------|------------|--------------|
|                                    |              | <b>MD</b>              | <b>MND</b> | <b>Total</b> |
| <b>ddEGFR Test</b>                 | <b>MD</b>    | 88                     | 6          | 94           |
|                                    | <b>MND</b>   | 0                      | 19         | 19           |
|                                    | <b>Total</b> | 88                     | 25         | <b>113</b>   |
| <b>PPA (95% C.I.)</b>              |              | 100.0% (95.89–100.0%)  |            |              |
| <b>NPA (95% C.I.)</b>              |              | 76.00% (54.87–90.64%)  |            |              |
| <b>OPA (95% C.I.)</b>              |              | 94.69% (88.80–98.03%)  |            |              |
| <b>PPV (95% C.I.)</b>              |              | 93.62% (86.62–97.62%)  |            |              |
| <b>NPV (95% C.I.)</b>              |              | 100.0% (82.35–100.0%)  |            |              |

MD, mutation detected; MND, mutation not detected

**Table 15. Comparison of ddEGFR and cobas EGFR results from FFPE samples eliminated by the sample criteria (DIN and iQC index).** Analysis of concordance rates between ddEGFR and cobas EGFR tests when the sample criteria were applied (Group 2). PPA, NPA, and OPA between the ddEGFR and cobas EGFR tests for the detection of *EGFR* mutations were 100%, 76.00%, and 94.69%, respectively. MD, mutation detected; MND, mutation not detected.

| Group 2<br>(n = 113) |                 | cobas EGFR Test |       |       |        |       |       |                 |                 |                 |     |       |
|----------------------|-----------------|-----------------|-------|-------|--------|-------|-------|-----------------|-----------------|-----------------|-----|-------|
|                      |                 | G719X           | 19del | T790M | E20Ins | S768I | L858R | G719X,<br>T790M | G719X,<br>S768I | S768I,<br>L858R | MND | Total |
| ddEGFR<br>Test       | G719X           | 1               |       |       |        |       |       |                 |                 |                 | 1   | 2     |
|                      | 19del           |                 | 51    |       |        |       |       |                 |                 |                 | 1   | 52    |
|                      | T790M           |                 |       | 0     |        |       |       |                 |                 |                 |     | 0     |
|                      | E20Ins          |                 |       |       | 0      |       |       |                 |                 |                 | 1   | 1     |
|                      | S768I           |                 |       |       |        | 0     |       |                 |                 |                 |     | 0     |
|                      | L858R           |                 |       |       |        |       | 33    |                 |                 |                 | 3   | 36    |
|                      | G719X,<br>T790M |                 |       |       |        |       |       | 0               |                 |                 |     | 0     |
|                      | G719X,<br>S768I |                 |       |       |        |       |       |                 | 2               |                 |     | 2     |
|                      | S768I,<br>L858R |                 |       |       |        |       |       |                 |                 | 1               |     | 1     |
|                      | MND             |                 |       |       |        |       |       |                 |                 |                 | 19  | 19    |
| <b>Total</b>         | 1               | 51              | 0     | 0     | 0      | 33    | 0     | 2               | 1               | 25              | 113 |       |

$\kappa$  coefficient = **0.9197** (95% C.I. 85.85–98.09%)

**Table 16. Comparison between the ddEGFR and cobas EGFR test results with application of the sample criteria (Group 2).**

| <b>Group 3</b><br><i>(n = 150)</i> |              | <b>cobas EGFR Test</b> |            |              |
|------------------------------------|--------------|------------------------|------------|--------------|
|                                    |              | <b>MD</b>              | <b>MND</b> | <b>Total</b> |
| <b>ddEGFR Test</b>                 | <b>MD</b>    | 106                    | 11 *       | 117          |
|                                    | <b>MND</b>   | 0                      | 33         | 33           |
|                                    | <b>Total</b> | 106                    | 44         | 150          |
| <b>PPA (95% C.I.)</b>              |              | 100.0% (96.58–100.0%)  |            |              |
| <b>NPA (95% C.I.)</b>              |              | 75.00% (59.66–86.81%)  |            |              |
| <b>OPA (95% C.I.)</b>              |              | 92.67% (87.26–96.28%)  |            |              |
| <b>PPV (95% C.I.)</b>              |              | 90.60% (83.80–95.21%)  |            |              |
| <b>NPV (95% C.I.)</b>              |              | 100.0% (87.26–100.0%)  |            |              |

\* 3 samples: cobas, 19del; ddEGFR, 19del/T790M

**Table 17. Comparison of ddEGFR and cobas EGFR results from FFPE samples eliminated by the sample criteria (iQC index only).** Analysis of concordance rates between ddEGFR and cobas EGFR tests when the sample criteria were applied (Group 3). PPA, NPA, and OPA between the ddEGFR and cobas EGFR tests for the detection of *EGFR* mutations were 100%, 75%, and 92.67%, respectively. MD, mutation detected; MND, mutation not detected.

| Group 3<br>(n = 150) |              | cobas EGFR Test |       |       |        |       |       |                 |                 |             |                 |     |
|----------------------|--------------|-----------------|-------|-------|--------|-------|-------|-----------------|-----------------|-------------|-----------------|-----|
|                      |              | G719X           | 19del | T790M | E20Ins | S768I | L858R | G719X,<br>T790M | G719X,<br>S768I | 19del,T790M | S768I,<br>L858R | MND |
| ddEGFR<br>Test       | G719X        | 1               |       |       |        |       |       |                 |                 |             | 1               | 2   |
|                      | 19del        |                 | 61    |       |        |       |       |                 |                 |             | 1               | 62  |
|                      | T790M        |                 |       | 0     |        |       |       |                 |                 |             |                 | 0   |
|                      | E20Ins       |                 |       |       | 0      |       |       |                 |                 |             | 1               | 1   |
|                      | S768I        |                 |       |       |        | 0     |       |                 |                 |             |                 | 0   |
|                      | L858R        |                 |       |       |        |       | 41    |                 |                 |             | 4               | 45  |
|                      | G719X,T790M  |                 |       |       |        |       |       | 0               |                 |             | 1               | 1   |
|                      | G719X,S768I  |                 |       |       |        |       |       |                 | 2               |             |                 | 2   |
|                      | 19del,T790M  |                 | 3     |       |        |       |       |                 |                 | 0           |                 | 3   |
|                      | S768I, L858R |                 |       |       |        |       |       |                 |                 |             | 1               | 1   |
|                      | MND          |                 |       |       |        |       |       |                 |                 |             | 33              | 33  |
| Total                | 1            | 64              | 0     | 0     | 0      | 41    | 0     | 2               | 0               | 1           | 41              | 150 |

K coefficient = **0.8923** (95% C.I. 83.25–95.22%)

**Table 18. Comparison between the ddEGFR and cobas EGFR test results, with application of only the iQC criterion (Group 3).**

| Group 4<br>(n = 166)  |       | cobas EGFR Test       |      |       |
|-----------------------|-------|-----------------------|------|-------|
|                       |       | MD                    | MND  | Total |
| ddEGFR<br>Test        | MD    | 33                    | 51 * | 84    |
|                       | MND   | 9                     | 71   | 80    |
|                       | Total | 42                    | 122  | 164   |
| <b>PPA (95% C.I.)</b> |       | 78.57% (63.19–89.70%) |      |       |
| <b>NPA (95% C.I.)</b> |       | 58.20% (48.93–67.06%) |      |       |
| <b>OPA (95% C.I.)</b> |       | 63.41% (55.55–70.79%) |      |       |
| <b>PPV (95% C.I.)</b> |       | 39.29% (28.80–50.55%) |      |       |
| <b>NPV (95% C.I.)</b> |       | 88.75% (79.72–94.72%) |      |       |

\* 1 sample: cobas, 19del; ddEGFR, 19del/L858R

1 sample: cobas, E20Ins; ddEGFR, E20Ins/T790M

1 sample: cobas, S768I; ddEGFR, S768I/L858R

**Table 19. Re-analysis of concordance rate using excluded samples.** Samples were selected by iQC index (< 0.5). The 166 excluded samples with valid ddEGFR test and cobas EGFR test results were included in the agreement analysis. PPA, NPA, and OPA between the ddEGFR and cobas EGFR tests for the detection of *EGFR* mutations were 78.57%, 58.20%, and 63.41%, respectively.

| Group 4<br>(n = 166) |                 | cobas EGFR Test |       |       |        |       |       |                 |                 |                 |                 |                  |                 |                 |     |       |
|----------------------|-----------------|-----------------|-------|-------|--------|-------|-------|-----------------|-----------------|-----------------|-----------------|------------------|-----------------|-----------------|-----|-------|
|                      |                 | G719X           | 19del | T790M | E20Ins | S768I | L858R | G719X,<br>T790M | G719X,<br>S768I | 19del,<br>T790M | 19del,<br>L858R | T790M,<br>E20Ins | T790M,<br>L858R | S768I,<br>L858R | N/A | Total |
| ddEGFR<br>Test       | G719X           | 0               |       |       |        |       |       |                 |                 |                 |                 |                  |                 |                 |     | 0     |
|                      | 19del           |                 | 24    |       |        |       |       |                 |                 |                 |                 |                  |                 |                 | 3   | 27    |
|                      | T790M           |                 |       | 0     |        |       |       |                 |                 |                 |                 |                  |                 |                 |     | 0     |
|                      | E20Ins          |                 |       |       | 0      |       |       |                 |                 |                 |                 |                  |                 |                 | 1   | 1     |
|                      | S768I           |                 |       |       |        | 0     |       |                 |                 |                 |                 |                  |                 |                 |     | 0     |
|                      | L858R           |                 | 1 †   |       |        |       | 7     |                 |                 |                 |                 |                  |                 |                 | 43  | 51    |
|                      | G719X,<br>T790M |                 |       |       |        |       |       | 0               |                 |                 |                 |                  |                 |                 |     | 0     |
|                      | G719X, S768I    |                 |       |       |        |       |       |                 | 1               |                 |                 |                  |                 |                 |     | 1     |
|                      | 19del, T790M    |                 |       |       |        |       |       |                 |                 | 0               |                 |                  |                 |                 |     | 0     |
|                      | 19del, L858R    |                 | 1     |       |        |       |       |                 |                 |                 | 0               |                  |                 |                 |     | 1     |
|                      | T790M, E20Ins   |                 |       |       | 1      |       |       |                 |                 |                 |                 | 0                |                 |                 |     | 1     |
|                      | T790M, L858R    |                 |       |       |        |       |       |                 |                 |                 |                 |                  | 0               |                 | 1   | 1     |
|                      | S768I, L858R    |                 |       |       |        | 1     |       |                 |                 |                 |                 |                  |                 | 1               |     | 2     |
|                      | N/A             |                 | 9     |       |        |       |       |                 |                 |                 |                 |                  |                 |                 | 71  | 80    |
| <b>Total</b>         | 0               | 35              | 0     | 1     | 1      | 7     | 0     | 1               | 0               | 0               | 0               | 0                | 1               | 119             | 165 |       |


† 1 sample was excluded from the analysis, 1 sample: invalid

$\kappa$  coefficient = 0.3862 (95% C.I. 27.85–49.40%)

**Table 20. Detail concordance rate using excluded samples.** Samples were selected by iQC index (< 0.5). The 166 excluded samples with valid ddEGFR test and cobas EGFR test results were included in the agreement analysis.

## 6. Analysis of discordant samples in pre-clinical study

Applying the iQC index criterion, we re-analyzed the remaining 11 discordant samples in Group 3. A schematic representation of the re-analysis workflow for discordant samples is depicted in Figure 17. In three of the samples, the ddEGFR test reported a double mutation (19del/T790M), whereas the cobas EGFR test and Sanger sequencing reported only a single mutation (19del) (Table 17, Group 3). This may be a result of the low detection sensitivity of Sanger sequencing (around 15%) (128) and cobas EGFR tests (LoD of T790M = ~ 3%; cobas EGFR v2). Based on the ddEGFR results of these three samples, the MI of T790M was ~1% (1.11%, 1.16%, and 1.03%). Additionally, eight discordant samples were verified by Sanger sequencing, and no mutations were found (Table. 21). Moreover, to observe the effect of tumor ratio, we performed macrodissection to enrich for tumor tissue, and then re-analyzed *EGFR* mutations in eight samples for which the cobas EGFR test had yielded negative results but the ddEGFR test yielded positive results. After macrodissection, the cobas EGFR test gave the same results as the ddEGFR test for four of the eight samples (Table. 21). Thus, our results indicate that the ddEGFR test is more sensitive for EGFR mutation detection, independent of tumor ratio. Strangely, one discordant case was a mutation detected (T790M/G719X) of preliminary result, whereas invalid by the ddEGFR test after macrodissection. Because iQC index is very low (0.37, data not shown), it is estimated that DNA degradation progressed during the macrodissection process.



**Figure 17. Schematic of re-analysis of discordant samples.** Eight of eleven discordant samples were verified by Sanger sequencing. In addition, after increasing the tumor-to-normal tissue ratio, *EGFR* mutations were re-analyzed using the cobas EGFR and ddEGFR tests.

| Sample No | Block Storage age | C/N Ratio (%) | DIN | Preliminary result |   |        |             |              |         | After Macrodissection |       |         |        |               |
|-----------|-------------------|---------------|-----|--------------------|---|--------|-------------|--------------|---------|-----------------------|-------|---------|--------|---------------|
|           |                   |               |     | cobas              |   | ddEGFR |             | ddEGFR MI(%) | Sanger  | cobas                 |       | ddEGFR  |        | ddEGFR MI (%) |
| 1         | 11                | 23            | 3.2 | MND                | - | MD     | L858R       | 0.8          | WT      | N/A                   | N/A   | N/A     | N/A    | -             |
| 2         | 11                | 51            | 2.2 | MND                | - | MD     | T790M/G719X | 1.08 / 1.02  | Invalid | MND                   | -     | Invalid | -      | -             |
| 3         | 6                 | 41            | 4.6 | MND                | - | MD     | L858R       | 1.57         | WT      | MD                    | L858R | MD      | L858R  | 2.57          |
| 4         | 6                 | 28            | 3.7 | MND                | - | MD     | G719X       | 10.17        | WT      | MD                    | G719X | MD      | G719X  | 20.17         |
| 5         | 6                 | 12            | 3.9 | MND                | - | MD     | L858R       | 5.91         | WT      | MD                    | L858R | MD      | L858R  | 4.55          |
| 6         | 6                 | 15            | 3.1 | MND                | - | MD     | L858R       | 7.4          | WT      | N/A                   | N/A   | N/A     | N/A    | -             |
| 7         | 3                 | 43            | 4.4 | MND                | - | MD     | 20Ins       | 12.88        | WT      | MND                   | -     | MD      | E20Ins | 7.4           |
| 8         | 3                 | 34            | 4.1 | MND                | - | MD     | 19del       | 15.72        | WT      | MD                    | 19del | MD      | 19del  | 3.15          |

**MD, mutation detected; MND, mutation not detected.**

**N/A, FFPE blocks not available**

**Table 21. Re-analysis of eight of eleven discordant samples in Group 3.** Eight of eleven discordant samples were verified by Sanger sequencing.

In addition, after increasing the tumor-to-normal tissue ratio, detection of EGFR mutations was re-analyzed using the cobas EGFR and ddEGFR tests. After enrichment of tumor tissue, the cobas EGFR test could identify previously undetected *EGFR* mutations in four out of the eight samples.

However, the ddEGFR analysis yielded data identical to the preliminary results.

## 7. Retrospective comparative clinical study for clinical utility of iQC index

Next, we analyzed the *EGFR* mutation status of 228 samples using the ddEGFR and cobas EGFR tests; 57 samples were excluded based on the iQC index (Fig. 8). The remaining 171 samples with iQC index  $\geq 0.5$  gave PPA of 98.23%, NPA of 82.76%, and OPA of 92.98% between the ddEGFR and cobas EGFR tests ( $\kappa = 0.9029$ , Table. 22, Table. 23). Among 12 discordant samples, six were reported to have a double mutation according to the ddEGFR test but only a single mutation according to the cobas EGFR test. As expected, the MI of the additional detected mutation was very low. One discordant case was a mutation not detected by the cobas EGFR test, but detected (L861Q) by both ddEGFR test and Sanger. Conversely, another discordant case was a mutation not detected by the ddEGFR test but detected (19del) by the cobas EGFR test and Sanger (Table. 24). This was a rare mutation of the 19del subtype (c.2239\_2264del\_insGCGAA) caused by a non-specific reaction that is not designed in the cobas EGFR test, and thus it cannot be employed to discriminate the possibility potential erroneous detection (129) and beneficial cross-reaction of commercial diagnostic kits.

| Retrospective comparison study,<br>Applied iQC index ( <i>n</i> = 171) |       | cobas EGFR Test       |      |       |
|--|-------|-----------------------|------|-------|
|  |       | MD                    | MND  | Total |
| ddEGFR<br>Test   | MD    | 111                   | 10 * | 121   |
|  | MND   | 2                     | 48   | 50    |
|  | Total | 113                   | 58   | 171   |
| PPA (95% C.I.)   |       | 98.23% (93.75–99.78%) |      |       |
| NPA (95% C.I.)   |       | 82.76% (70.57–91.41%) |      |       |
| OPA (95% C.I.)   |       | 92.98% (88.06–96.32%) |      |       |
| PPV (95% C.I.)   |       | 91.74% (85.33–95.97%) |      |       |
| NPV (95% C.I.)   |       | 96.00% (86.29–99.51%) |      |       |

\* 3 samples: Cobas, 19del; ddEGFR, 19del/T790M

1 sample: Cobas, G719X; ddEGFR, G719X/L861Q

1 sample: Cobas, L858R; ddEGFR, G719X/L858R

1 sample: Cobas, L858R; ddEGFR, T790M/L858R

**Table 22. Method correlation between ddEGFR and cobas EGFR test in Retrospective comparison study group.** Samples with valid ddEGFR and cobas EGFR test results were included in the agreement analysis. PPA, NPA, and OPA between the ddEGFR and cobas EGFR tests for the detection of *EGFR* mutations were 98.23%, 82.76%, and 92.98%, respectively. MD, mutation detected; MND, mutation not detected.

| Retrospective comparison study, Applied iQC index (n = 171) |              | cobas EGFR Test |          |          |          |           |          |          |              |              |              |              |              |            |       |
|---|--------------|-----------------|----------|----------|----------|-----------|----------|----------|--------------|--------------|--------------|--------------|--------------|------------|-------|
|   |              | G719X           | 19del    | T790M    | E20Ins   | S768I     | L858R    | L861Q    | 19del, T790M | G719X, S768I | G719X, L858R | G719X, L861Q | T790M, L858R | N/A        | Total |
| ddEGFR Test   | G719X        | 2               |          |          |          |           |          |          |              |              |              |              |              |            | 2     |
|   | 19del        |                 | 46       |          |          |           |          |          |              |              |              |              |              |            | 46    |
|   | T790M        |                 |          | 0        |          |           |          |          |              |              |              |              |              |            | 0     |
|   | E20Ins       |                 |          |          | 2        |           |          |          |              |              |              |              |              | 1          | 3     |
|   | S768I        |                 |          |          |          | 0         |          |          |              |              |              |              |              |            | 0     |
|   | L858R        |                 |          |          |          |           | 54       |          |              |              |              |              |              | 2          | 56    |
|   | L861Q        |                 |          |          |          |           |          | 0        |              |              |              |              |              | 1          | 1     |
|   | 19del, T790M |                 | 3        |          |          |           |          |          | 1            |              |              |              |              |            | 4     |
|   | G719X, S768I |                 |          |          |          |           |          |          |              | 3            |              |              |              |            | 3     |
|   | G719X, L858R |                 |          |          |          |           | 1        |          |              |              | 0            |              |              |            | 1     |
|   | G719X, L861Q | 1               |          |          |          |           |          |          |              |              |              | 0            |              |            | 1     |
|   | T790M, L858R |                 |          |          |          |           | 1        |          |              |              |              |              | 3            |            | 4     |
|   | N/A          |                 | 2        |          |          |           |          |          |              |              |              |              |              | 48         | 50    |
| <b>Total</b>  | <b>3</b>     | <b>51</b>       | <b>0</b> | <b>2</b> | <b>0</b> | <b>56</b> | <b>0</b> | <b>1</b> | <b>3</b>     | <b>0</b>     | <b>0</b>     | <b>3</b>     | <b>52</b>    | <b>171</b> |       |

$\kappa$  coefficient = **0.9029** (95% C.I. 85.08–95.49%)


**Table 23. Detailed concordance rate of the retrospective comparison study group.**

| Sample No. | cobas EGFR test |       | ddEGFR test |             | ddEGFR MI      |               | Sanger Seq |
|------------|-----------------|-------|-------------|-------------|----------------|---------------|------------|
|            |                 |       |             |             |                |               |            |
| 1          | MND             | -     | MD          | L858R       | L858R, 25.47%  |               | Invalid    |
| 2          | MD              | L858R | MD          | L858R/T790M | L858R, 42.9%   | T790M, 1.01%  | L858R      |
| 3          | MD              | 19del | MD          | 19del/T790M | 19del, 12.79%  | T790M, 0.96%  | 19del      |
| 4          | MD              | L858R | MD          | L858R/G719X | L858R, 43.22%  | G719X, 1.98%  | Invalid    |
| 5          | MD              | 19del | MD          | 19del/T790M | 19del, 22.45%  | T790M, 1.22%  | 19del      |
| 6          | MD              | 19del | MD          | 19del/T790M | 19del, 47.94%  | T790M, 0.87%  | 19del      |
| 7          | MD              | 19del | MND         | -           | -              |               | WT         |
| 8          | MND             | -     | MD          | E20Ins      | E20Ins, 14.55% |               | Invalid    |
| 9          | MD              | 19del | MND         | -           | -              |               | 19del *    |
| 10         | MND             | -     | MD          | L861Q       | L861Q, 64.93%  |               | L861Q      |
| 11         | MD              | G719X | MD          | G719X/L861Q | G719X, 11.36%  | L861Q, 13.64% | Invalid    |
| 12         | MND             | -     | MD          | L858R       | L858R, 6.46%   |               | Invalid    |

\* This was a rare mutation of the 19del subtype (c.2239\_2264del\_insGCGAA) caused by a non-specific reaction that is not designed in the cobas EGFR test, and thus it cannot be employed to discriminate the possibility potential erroneous detection (129).

**Table 24. Discordant analysis by Sanger sequencing in the retrospective comparison study group.** Based on the results of the retrospective comparison with cobas EGFR test, 12 samples were valid and discordant with the ddEGFR test. Samples that gave discordant results between the cobas EGFR and ddEGFR tests were analyzed by Sanger sequencing. MD, mutation detected; MND, mutation not detected.

In addition, we measured DIN values from 228 FFPET-DNA samples and observed a pattern similar to that of iQC index. Furthermore, the majority of the most recent samples (within 1 year) had  $DIN > 2.5$  and  $iQC \text{ index} \geq 0.5$  (Fig. 18). These data revealed that iQC index is a very powerful indicator of the quality of FFPET-DNA. In addition, these observations demonstrate that the ddEGFR test is a robust diagnostic tool for the accurate detection of *EGFR* mutations in clinical practice.



**Figure 18. Distributions of iQC index and DIN in the retrospective comparative study group.** Most of the more recent samples (taken within 1 year) satisfied the sample criteria defined above (red dotted line) (\*\*\*) ( $p. < 0.0001$ ).

## **V. Discussion**

FFPETs undergo effective preservation of the cellular, architectural, and morphological details and allow easy storage at room temperature for extensive periods. For these reasons, FFPET is useful source for molecular diagnostics. But, the quality of FFPET-DNA has been largely ignored in the clinical research and diagnosis field, and internal control of most commercial diagnostic kits were used only to validate assays. Changes in mutation status due to the low quality of FFPET-DNA may result in incorrect diagnoses. Therefore, considerable effort is required to optimize the sample criteria for determining the quality of FFPET-DNA that is suitable for PCR. In this study, we established iQC index criteria to determine the minimum quality of FFPET-DNA and demonstrated the benefits of implementing these criteria benefits in a real-world clinical application.

In our experiments, both automated tissue preparation system (TPS; Siemens Healthcare, Erlangen, Germany), which can minimize handling errors and decrease the effect of formaldehyde-induced DNA–DNA and DNA–protein crosslinks (51, 58), and UDG treatment are powerful strategy for reducing false positives caused by sequence artifacts (61, 130-132). However, sequence artifacts due to DNA fragmentation remained a problem. Limitations in the availability of many clinical specimens drive the need for low DNA inputs into molecular assays. Cutting edge technologies such as NGS required good quality of DNA material for in-depth molecular profiling in various cancer (133). None of traditional tools such as UV absorbance, gel electrophoresis, assays with intercalating dyes measure the amount of amplifiable DNA, and each of these tools has an additional limitations (134). Also, depending on the degree of fragmentation, the same quantities of DNA from different FFPET samples can contain widely different amounts of amplifiable DNA templates (59). For this reason, PCR-based methods such as qPCR, ddPCR, and next-generation sequencing (NGS) are preferable for quantifying the amount of amplifiable template in FFPET-DNA (119).

iQC copies can represent the concentration of input DNA (Fig. 11, 12) and a novel concept for mutation calling, MI, was introduced as an indicator of the mutation level, which in turn reflects DNA quality. MI provides a much more accurate indication of the mutation level than mutation frequency, which is simply calculated based on input DNA concentration. Accordingly, we designed an iQC that measures *EGFR* exon 20 to assess the amount of amplifiable FFPET-DNA. Because elevated *EGFR* expression have been observed in NSCLC (128), the internal control should be designed to represent the levels of the *EGFR* gene copy number.

Interestingly, some discordant samples reported as single mutations by the cobas EGFR test were shown to have double mutation by the ddEGFR test in pre-clinical (Table. 17) and clinical studies (Table. 22). The additional detected MI was around 1%, as expected. According to a clonal selection model, EGFR-TKI treatment may lead to the selection of T790M mutant cells, and thus even a small fraction of T790M positive tumor cells at the beginning of treatment could lead to clinical EGFR-TKI resistance (25). Therefore, is very important to detect even a small fraction of mutations.

The discordant results were largely due to the difference in LoD between both tests. Specifically, some mutations detected by the ddEGFR test cannot be detected by the cobas EGFR test because the sensitivity of the ddEGFR test is higher. In the pre-clinical study Group 3, L858R MIs were relatively low in cobas EGFR-negative samples (around 5%, Table. 21). Considering that the LoD of the cobas EGFR test was 5%, these MI values, which reflect sample quality and mutation level, are reasonable. When the iQC index was applied, the proportion of such discordant samples was small.

The cobas EGFR test recommends macrodissection of low-percent tumor tissues (below

10%) to improve detectability (126). When the cobas EGFR test yielded a result of ‘mutation not detected’ (MND), even though the C/N ratio was over 10%, the *EGFR* mutation was re-analyzed after increasing the tumor ratio by macrodissection. The cobas EGFR test yielded the same results as the ddEGFR test for four of the eight samples (Table. 21), showing that ddEGFR exhibits superior analytical performance. Thus, the detectability of the cobas EGFR test may be increased by enriching for tumor tissue, e.g., by macrodissection, but this manipulation requires extra time and effort. By contrast, the ddEGFR test does not require macrodissection and improves reliability.

For 1-year-old samples ( $n = 46$ ) in the clinical study group, all iQC index and DIN values satisfied the sample criteria (iQC index  $\geq 0.5$ , DIN  $> 2.5$ ) (Fig. 18), suggesting that the low age samples used in clinical practice will have minimal problems with DNA quality. Previous reported that, cobas EGFR test and Sanger are highly concordant because used to low age samples (50, 135, 136). In contrast, concordance rate between cobas EGFR test and Sanger of the pre-clinical study samples were very low (Table. 10, Table. 11), but when iQC index was applied, the concordance rate was increased (Data not shown). Given the additional cost and effort required for DIN measurement, the iQC index is a robust indicator of the minimum DNA quality required for PCR amplification, as it takes into account the various types of DNA damage caused by FFPE processing and long-term storage.

In addition, iQC index  $\geq 0.5$  is an indicator that guarantees clinical equivalence obtained through comparative clinical studies with existing approved diagnostics test (cobas EGFR test). Clinical equivalence cannot be guaranteed for samples with an iQC index  $< 0.5$ . However, discordance of samples with iQC index  $< 0.5$  can be explained by two causes. The first cause may be due to the possibility of false negative results of cobas EGFR test. Second, there is a possibility of false positive result due to ultra-sensitivity of ddEGFR test. The analytical

performance of the ddEGFR test indicates a high possibility of false negative results on the cobas EGFR test. However, it is questionable whether mutations detected in samples with an iQC index  $< 0.5$  (especially mutations detected at low MI %) are related to drug treatment response. This must be verified through prospective clinical trials. Small-scale independent retrospective clinical studies for the ddEGFR test have shown responses of patients treated with an EGFR-TKI, even of patients with an iQC index  $< 0.5$  (unpublished). Therefore, an iQC index  $\geq 0.5$  is a numerical value used in comparative clinical studies to guarantee clinical equivalence between the results of the ddEGFR and cobas EGFR tests. This means that iQC index may be changed through an independent prospective clinical trial for ddEGFR test. Also, using another approved kit for comparative clinical study, the iQC index may be changed for clinical equivalence. This is the result of considering clinical cut-off not only to compare the mutation results but also to ensure EGFR-TKI response.

We also observed that the iQC index could be used to obtain a diagnosis using other cancer type samples (colorectal cancer) and other diagnostic kits, including the ddPCR-based KRAS mutation test and the cobas<sup>®</sup> KRAS mutation test. Similar to the results of this study, we observed that when the iQC index was applied, the concordance rate between the results of the two kits increased significantly (data not shown). Overall, the iQC index could be a useful criterion for judging the quality of FFPE-DNA.

The ddEGFR test is designed to detect the C797S mutation in *EGFR* exon 20. NSCLC patients with advanced disease are usually treated with first or second generation EGFR-TKIs (erlotinib, gefitinib, or afatinib) (137, 138). However, patients acquire resistance after 9 to 13 months (34, 139, 140), the most common being T790M within the *EGFR* kinase domain (22, 24). Third generation EGFR-TKI have been designed to mutant-selectively target highly active against T790M mutation (141, 142). Recently, ctDNA evaluation was used to identify a novel

mechanism underlying acquired resistance to the third generation EGFR-TKI osimertinib; specifically, emergence of an acquired *EGFR* C797S mutation (33). The pre-clinical and retrospective clinical studies based on the ddEGFR test did not detect the C797S mutation. However, we subsequently found out that none of the NSCLC patients in those studies actually received a third generation EGFR-TKI.

Diagnostic kits must be able to detect mutations in very small amounts of DNA with high sensitivity. One of the problems associated with detecting *EGFR* mutations in NSCLC patients is not obtaining a sufficient quantity of specimen to confirm the status of several biomarkers. The ddEGFR test can detect *EGFR* mutations using only 3.3 ng (total 13.2 ng/4 wells) of FFPE-DNA per well, with a sensitivity of 0.78%. With cobas EGFR test, 3 reactions were performed using 50 ng DNA per well (total 150 ng / 3 well) to detect 42 *EGFR* mutations.

The presence of EGFR mutations in NSCLC is a major prognostic biomarker in terms of treatment selection. Thus, sensitive and accurate screening for EGFR mutations facilitates a personalized and specific approach to patient treatment / management.

In conclusion, detecting genetic mutations in FFPE samples is difficult because the DNA often becomes fragmented during storage, resulting in sequence artifacts and low-quality data. The iQC index allows selection of appropriate FFPE-DNA samples for companion diagnostics using a ddPCR-based mutation test. Furthermore, we suggest that clinical trials using FFPE should involve criteria that reflect the quality of the DNA in samples. The performance of the ddPCR-based EGFR mutation test developed herein was superior to that of the commercially available Cobas EGFR test.

## **VI. Conclusion and perspectives**

Personalized treatment for cancer patients should include a genetic assessment of mutation status. FFPET is the most widely available material for molecular diagnostics and research. FFPE samples are routinely used for histopathological analysis because they are convenient, cost-effective, and easy to store. However, detecting genetic mutations in FFPE samples is difficult because DNA fragments during storage, resulting in sequence artifacts. DNA fragmentation is an irreversible reaction that not only reduces the amount of amplifiable template but also leads to stochastic enrichment of artifactual changes. However, since mutational analyses of FFPEs is a standard procedure used to inform therapeutic decisions, appropriate measures must be incorporated to allow assessment of DNA damage and prevent the calling of false-positive mutations.

For reliable analysis of mutations, it is necessary to pre-analyze and evaluate the amount/quality of amplifiable template FFPE DNA. The more sensitive test method, the higher purity DNA is required. Traditional quantitation methods such as UV absorbance and spectrophotometry lack accuracy when it comes to measuring the amount of amplifiable template DNA. Evaluating the amount of amplifiable DNA is very important in diagnosis using PCR-based methods. Up to this time, the PCR-based diagnostic test reflected only the validity and invalidity of the assay, not considering the quality of the DNA. The iQC index established herein enables accurate detection of *EGFR* mutations. If using PCR-based assays to detect *EGFR* mutations, we believe that the results can be trusted if at least half of the input DNA is amplifiable. The iQC index allows appropriate selection of FFPE DNA samples for companion diagnosis using a ddPCR-based mutation test. In addition, DNA quality and EGFR mutations are analyzed simultaneously in same NSCLC FFPE sample. The results are reliable, even though only a small amount of FFPE DNA is required. Indeed, we found that the analytical performance of the ddPCR-based EGFR mutation test was superior to that of the

Cobas EGFR test. Measuring the quality of DNA extracted from patient samples makes it possible to prevent diagnostic errors, allow re-testing where appropriate, and make decisions regarding appropriate treatment (Fig. 19). The results of this study will also be of benefit to clinical researchers. As mentioned above, handling and storage procedures vary from site to site, and multi-center clinical trials that do not take sample quality into account can yield unreliable results; both of these problems can be overcome by application of the iQC index and standardization of sample quality (Fig. 20).

The iQC index seems to be most useful for diagnoses based on liquid biopsy samples (e.g., ctDNA). The iQC index may be necessary for liquid biopsy specimens since the quality of ctDNA can vary diversely by frequent fragmentation, which is a biological property of ctDNA, variations in sample collection and storage by each hospital, and by logistic issues.

We are currently conducting prospective clinical studies to confirm the clinical utility of the ddPCR-based EGFR mutation test for liquid biopsy samples. A ddEGFR test based on plasma samples may be both feasible and effective for identifying EGFR mutations, predicting treatment responses, and monitoring acquired resistance during EGFR-TKI treatment. T790M is the most common mutation associated with acquired resistance to first-generation EGFR-TKIs (23). T790M is present in a small population of tumor cells during development of NSCLC; therefore, it may be detected at very low frequency (25). A very low level T790M not detected by Cobas EGFR can be sensitively detected with ddEGFR test. However, it is very difficult to verify using the gold standard method such as Sanger sequencing due to technical limitations (low sensitivity). Therefore, we intend to undertake a prospective clinical trial to examine the usefulness of the ddEGFR test for monitoring the activity of EGFR-TKIs in NSCLC patients harboring T790M at low frequency; this will inform us as to whether the test can predict patients that will benefit from treatment with third generation EGFR-TKIs.

In the near future, mutation detection tests will play a very important role in the clinical setting as they will identify biomarkers for molecular targeted therapy. The benefits of treatment will be maximized only if accurate companion diagnostics are used; these diagnostics must be validated according to strict standards. The ddEGFR test not only detects *EGFR* mutations but also provides information about the quality of DNA, thereby preventing diagnostics errors. It also offers the opportunity to decide when to appropriate treatment through serial monitoring using liquid biopsy (Fig. 21).



ex> Input DNA 1000 copies

| Ex> Input DNA 1000 copies |                        |                | ddEGFR test  |  |         | Cobas EGFR test                                       |        |
|---------------------------|------------------------|----------------|--|--|---------|---|--------|
| Patient No.               | Amplifiable DNA copies | Mut DNA copies | iQC index<br>(amplifiable DNA copy/<br>Input DNA copies) | MI %<br>(Mut copy/Amplifiable<br>DNA copy) x 100 | Result  | Mutation freq %<br>(Mut copy/Input DNA copy)<br>x 100 | Result |
| Patient 1                 | 1000                   | 100            | 1000/1000 = 1  | 100/1000 x 100 = 10                              | MD      | 100/1000 x 100 = 10                                   | MD     |
| Patient 2                 | 500                    | 50             | 500/1000 = 0.5   | 50/500 x 100 = 10                                | MD      | 50/1000 x 100 = 5.0                                   | MD     |
| Patient 3                 | 250                    | 25             | 250/1000 = 0.25  | 25/250 x 100 = 10                                | Invalid | 25/1000 x 100 = 2.5                                   | MND    |


MD: Mutation detected  
MND: Mutation not detected




**Re-test ability**


  
**Lose opportunity of re-test and treatment**

Figure 19. Clinical meaning of iQC index (risk management). Prevent diagnostic errors due to DNA quality and risk management is possible.



**Figure 20. Clinical meaning of iQC index (viewpoint of multi-center clinical and translational research).** Since conditions for processing and storage of samples from each clinical institution are different, it is necessary to provide an indicator for the quality of DNA.



Illustrated by Tae Won Yun

**Figure 21. Dynamic monitoring of ctDNA in NSCLC.** ddPCR-based mutation test can be used for the quantification of dynamic changes in the *EGFR*, *c-Met* mutations in ctDNA, and can be used to assess the clinical outcomes of therapy.

## VII. References

1. Mok TS. Personalized medicine in lung cancer: What we need to know. *Nature reviews Clinical oncology* 2011;8:661-8.
2. Arteaga CL, Sliwkowski MX, Osborne CK, Perez EA, Puglisi F, Gianni L. Treatment of her2-positive breast cancer: Current status and future perspectives. *Nature reviews Clinical oncology* 2012;9:16-32.
3. Libermann TA, Nusbaum HR, Razon N, Kris R, Lax I, Soreq H, et al. Amplification, enhanced expression and possible rearrangement of egf receptor gene in primary human brain tumours of glial origin. *Nature* 1985;313:144-7.
4. Lee JC, Vivanco I, Beroukhim R, Huang JH, Feng WL, DeBiasi RM, et al. Epidermal growth factor receptor activation in glioblastoma through novel missense mutations in the extracellular domain. *PLoS medicine* 2006;3:e485.
5. Vivanco I, Robins HI, Rohle D, Campos C, Grommes C, Nghiemphu PL, et al. Differential sensitivity of glioma- versus lung cancer-specific egfr mutations to egfr kinase inhibitors. *Cancer discovery* 2012;2:458-71.
6. Brambilla E, Travis WD, Colby TV, Corrin B, Shimosato Y. The new world health organization classification of lung tumours. *The European respiratory journal* 2001;18:1059-68.
7. Breathnach OS, Freidlin B, Conley B, Green MR, Johnson DH, Gandara DR, et al. Twenty-two years of phase iii trials for patients with advanced non-small-cell lung cancer: Sobering results. *Journal of clinical oncology : official journal of the American Society of Clinical Oncology* 2001;19:1734-42.
8. Blackhall F, Ranson M, Thatcher N. Where next for gefitinib in patients with lung cancer? *The Lancet Oncology* 2006;7:499-507.

9. Mendelsohn J, Baselga J. Status of epidermal growth factor receptor antagonists in the biology and treatment of cancer. *Journal of clinical oncology : official journal of the American Society of Clinical Oncology* 2003;21:2787-99.
10. Herbst RS, Maddox AM, Rothenberg ML, Small EJ, Rubin EH, Baselga J, et al. Selective oral epidermal growth factor receptor tyrosine kinase inhibitor zolmitriptan is generally well-tolerated and has activity in non-small-cell lung cancer and other solid tumors: Results of a phase I trial. *Journal of clinical oncology : official journal of the American Society of Clinical Oncology* 2002;20:3815-25.
11. Thatcher N, Chang A, Parikh P, Rodrigues Pereira J, Ciuleanu T, von Pawel J, et al. Gefitinib plus best supportive care in previously treated patients with refractory advanced non-small-cell lung cancer: Results from a randomised, placebo-controlled, multicentre study (Iressa survival evaluation in lung cancer). *Lancet* 2005;366:1527-37.
12. Kosaka T, Yatabe Y, Endoh H, Kuwano H, Takahashi T, Mitsudomi T. Mutations of the epidermal growth factor receptor gene in lung cancer: Biological and clinical implications. *Cancer research* 2004;64:8919-23.
13. Han SW, Kim TY, Hwang PG, Jeong S, Kim J, Choi IS, et al. Predictive and prognostic impact of epidermal growth factor receptor mutation in non-small-cell lung cancer patients treated with gefitinib. *Journal of clinical oncology : official journal of the American Society of Clinical Oncology* 2005;23:2493-501.
14. Lynch TJ, Bell DW, Sordella R, Gurubhagavatula S, Okimoto RA, Brannigan BW, et al. Activating mutations in the epidermal growth factor receptor underlying responsiveness of non-small-cell lung cancer to gefitinib. *The New England journal of medicine* 2004;350:2129-39.
15. Paez JG, Janne PA, Lee JC, Tracy S, Greulich H, Gabriel S, et al. Egfr mutations in

- lung cancer: Correlation with clinical response to gefitinib therapy. *Science* 2004;304:1497-500.
16. Janne PA, Engelman JA, Johnson BE. Epidermal growth factor receptor mutations in non-small-cell lung cancer: Implications for treatment and tumor biology. *Journal of clinical oncology : official journal of the American Society of Clinical Oncology* 2005;23:3227-34.
  17. Maemondo M, Inoue A, Kobayashi K, Sugawara S, Oizumi S, Isobe H, et al. Gefitinib or chemotherapy for non-small-cell lung cancer with mutated egfr. *The New England journal of medicine* 2010;362:2380-8.
  18. Fang S, Wang Z. Egfr mutations as a prognostic and predictive marker in non-small-cell lung cancer. *Drug design, development and therapy* 2014;8:1595-611.
  19. Rosell R, Bivona TG, Karachaliou N. Genetics and biomarkers in personalisation of lung cancer treatment. *Lancet* 2013;382:720-31.
  20. Rosell R, Carcereny E, Gervais R, Vergnenegre A, Massuti B, Felip E, et al. Erlotinib versus standard chemotherapy as first-line treatment for european patients with advanced egfr mutation-positive non-small-cell lung cancer (eurtac): A multicentre, open-label, randomised phase 3 trial. *The Lancet Oncology* 2012;13:239-46.
  21. Kosaka T, Yatabe Y, Endoh H, Yoshida K, Hida T, Tsuboi M, et al. Analysis of epidermal growth factor receptor gene mutation in patients with non-small cell lung cancer and acquired resistance to gefitinib. *Clinical cancer research : an official journal of the American Association for Cancer Research* 2006;12:5764-9.
  22. Sequist LV, Waltman BA, Dias-Santagata D, Digumarthy S, Turke AB, Fidias P, et al. Genotypic and histological evolution of lung cancers acquiring resistance to egfr inhibitors. *Science translational medicine* 2011;3:75ra26.

23. Yu HA, Arcila ME, Rekhtman N, Sima CS, Zakowski MF, Pao W, et al. Analysis of tumor specimens at the time of acquired resistance to egfr-tyki therapy in 155 patients with egfr-mutant lung cancers. *Clinical cancer research : an official journal of the American Association for Cancer Research* 2013;19:2240-7.
24. Arcila ME, Oxnard GR, Nafa K, Riely GJ, Solomon SB, Zakowski MF, et al. Rebiopsy of lung cancer patients with acquired resistance to egfr inhibitors and enhanced detection of the t790m mutation using a locked nucleic acid-based assay. *Clinical cancer research : an official journal of the American Association for Cancer Research* 2011;17:1169-80.
25. Inukai M, Toyooka S, Ito S, Asano H, Ichihara S, Soh J, et al. Presence of epidermal growth factor receptor gene t790m mutation as a minor clone in non-small cell lung cancer. *Cancer research* 2006;66:7854-8.
26. Pao W, Miller VA, Politi KA, Riely GJ, Somwar R, Zakowski MF, et al. Acquired resistance of lung adenocarcinomas to gefitinib or erlotinib is associated with a second mutation in the egfr kinase domain. *PLoS medicine* 2005;2:e73.
27. Lee Y, Lee GK, Lee YS, Zhang W, Hwang JA, Nam BH, et al. Clinical outcome according to the level of preexisting epidermal growth factor receptor t790m mutation in patients with lung cancer harboring sensitive epidermal growth factor receptor mutations. *Cancer* 2014;120:2090-8.
28. Su KY, Chen HY, Li KC, Kuo ML, Yang JC, Chan WK, et al. Pretreatment epidermal growth factor receptor (egfr) t790m mutation predicts shorter egfr tyrosine kinase inhibitor response duration in patients with non-small-cell lung cancer. *Journal of clinical oncology : official journal of the American Society of Clinical Oncology* 2012;30:433-40.

29. Janne PA, Yang JC, Kim DW, Planchard D, Ohe Y, Ramalingam SS, et al. Azd9291 in egfr inhibitor-resistant non-small-cell lung cancer. *The New England journal of medicine* 2015;372:1689-99.
30. Sequist LV, Soria JC, Goldman JW, Wakelee HA, Gadgeel SM, Varga A, et al. Rociletinib in egfr-mutated non-small-cell lung cancer. *The New England journal of medicine* 2015;372:1700-9.
31. Kim ES. Olmutinib: First global approval. *Drugs* 2016;76:1153-7.
32. Minari R, Bordi P, Tiseo M. Third-generation epidermal growth factor receptor-tyrosine kinase inhibitors in t790m-positive non-small cell lung cancer: Review on emerged mechanisms of resistance. *Translational lung cancer research* 2016;5:695-708.
33. Thress KS, Paweletz CP, Felip E, Cho BC, Stetson D, Dougherty B, et al. Acquired egfr c797s mutation mediates resistance to azd9291 in non-small cell lung cancer harboring egfr t790m. *Nature medicine* 2015;21:560-2.
34. Mok TS, Wu YL, Thongprasert S, Yang CH, Chu DT, Saijo N, et al. Gefitinib or carboplatin-paclitaxel in pulmonary adenocarcinoma. *The New England journal of medicine* 2009;361:947-57.
35. Fukuoka M, Wu YL, Thongprasert S, Sunpaweravong P, Leong SS, Sriuranpong V, et al. Biomarker analyses and final overall survival results from a phase iii, randomized, open-label, first-line study of gefitinib versus carboplatin/paclitaxel in clinically selected patients with advanced non-small-cell lung cancer in asia (ipass). *Journal of clinical oncology : official journal of the American Society of Clinical Oncology* 2011;29:2866-74.
36. Mitsudomi T, Morita S, Yatabe Y, Negoro S, Okamoto I, Tsurutani J, et al. Gefitinib versus cisplatin plus docetaxel in patients with non-small-cell lung cancer harbouring

- mutations of the epidermal growth factor receptor (wjtog3405): An open label, randomised phase 3 trial. *The Lancet Oncology* 2010;11:121-8.
37. Zhou C, Wu YL, Chen G, Feng J, Liu XQ, Wang C, et al. Erlotinib versus chemotherapy as first-line treatment for patients with advanced egfr mutation-positive non-small-cell lung cancer (optimal, ctong-0802): A multicentre, open-label, randomised, phase 3 study. *The Lancet Oncology* 2011;12:735-42.
  38. Sequist LV, Yang JC, Yamamoto N, O'Byrne K, Hirsh V, Mok T, et al. Phase iii study of afatinib or cisplatin plus pemetrexed in patients with metastatic lung adenocarcinoma with egfr mutations. *Journal of clinical oncology : official journal of the American Society of Clinical Oncology* 2013;31:3327-34.
  39. Wu YL, Zhou C, Hu CP, Feng J, Lu S, Huang Y, et al. Afatinib versus cisplatin plus gemcitabine for first-line treatment of asian patients with advanced non-small-cell lung cancer harbouring egfr mutations (lux-lung 6): An open-label, randomised phase 3 trial. *The Lancet Oncology* 2014;15:213-22.
  40. Mok TS, Wu YL, Ahn MJ, Garassino MC, Kim HR, Ramalingam SS, et al. Osimertinib or platinum-pemetrexed in egfr t790m-positive lung cancer. *The New England journal of medicine* 2017;376:629-40.
  41. Park K, Han JY, Kim DW, Bazhenova LA, Ou SH, Pang YK, et al. 190tip: Eluxa 1: Phase ii study of bi 1482694 (hm61713) in patients (pts) with t790m-positive non-small cell lung cancer (nscl) after treatment with an epidermal growth factor receptor tyrosine kinase inhibitor (egfr tki). *Journal of thoracic oncology : official publication of the International Association for the Study of Lung Cancer* 2016;11:S139.
  42. Jia Y, Juarez J, Li J, Manuia M, Niederst MJ, Tompkins C, et al. Egf816 exerts anticancer effects in non-small cell lung cancer by irreversibly and selectively targeting

- primary and acquired activating mutations in the egf receptor. *Cancer research* 2016;76:1591-602.
43. Liao BC, Lin CC, Lee JH, Yang JC. Update on recent preclinical and clinical studies of t790m mutant-specific irreversible epidermal growth factor receptor tyrosine kinase inhibitors. *Journal of biomedical science* 2016;23:86.
  44. Slamon DJ, Leyland-Jones B, Shak S, Fuchs H, Paton V, Bajamonde A, et al. Use of chemotherapy plus a monoclonal antibody against her2 for metastatic breast cancer that overexpresses her2. *The New England journal of medicine* 2001;344:783-92.
  45. Beau-Faller M, Degeorges A, Rolland E, Mounawar M, Antoine M, Poulot V, et al. Cross-validation study for epidermal growth factor receptor and kras mutation detection in 74 blinded non-small cell lung carcinoma samples: A total of 5550 exons sequenced by 15 molecular french laboratories (evaluation of the egfr mutation status for the administration of egfr-tkis in non-small cell lung carcinoma [ermetic] project--part 1). *Journal of thoracic oncology : official publication of the International Association for the Study of Lung Cancer* 2011;6:1006-15.
  46. Bellon E, Ligtenberg MJ, Tejpar S, Cox K, de Hertogh G, de Stricker K, et al. External quality assessment for kras testing is needed: Setup of a european program and report of the first joined regional quality assessment rounds. *The oncologist* 2011;16:467-78.
  47. Penzel R, Sers C, Chen Y, Lehmann-Muhlenhoff U, Merkelbach-Bruse S, Jung A, et al. Egfr mutation detection in nslc--assessment of diagnostic application and recommendations of the german panel for mutation testing in nslc. *Virchows Arch* 2011;458:95-8.
  48. Nagai Y, Miyazawa H, Huqun, Tanaka T, Udagawa K, Kato M, et al. Genetic heterogeneity of the epidermal growth factor receptor in non-small cell lung cancer cell

- lines revealed by a rapid and sensitive detection system, the peptide nucleic acid-locked nucleic acid pcr clamp. *Cancer research* 2005;65:7276-82.
49. Benloch S, Botero ML, Beltran-Alamillo J, Mayo C, Gimenez-Capitan A, de Aguirre I, et al. Clinical validation of a pcr assay for the detection of egfr mutations in non-small-cell lung cancer: Retrospective testing of specimens from the eurtac trial. *PloS one* 2014;9:e89518.
  50. Kimura H, Ohira T, Uchida O, Matsubayashi J, Shimizu S, Nagao T, et al. Analytical performance of the cobas egfr mutation assay for japanese non-small-cell lung cancer. *Lung cancer* 2014;83:329-33.
  51. Lehmann A, Schewe C, Hennig G, Denkert C, Weichert W, Budczies J, Dietel M. Applicability of a system for fully automated nucleic acid extraction from formalin-fixed paraffin-embedded sections for routine kras mutation testing. *Diagnostic molecular pathology : the American journal of surgical pathology, part B* 2012;21:114-9.
  52. Becker KF, Schott C, Hipp S, Metzger V, Porschewski P, Beck R, et al. Quantitative protein analysis from formalin-fixed tissues: Implications for translational clinical research and nanoscale molecular diagnosis. *The Journal of pathology* 2007;211:370-8.
  53. Specht K, Richter T, Muller U, Walch A, Werner M, Hofler H. Quantitative gene expression analysis in microdissected archival formalin-fixed and paraffin-embedded tumor tissue. *The American journal of pathology* 2001;158:419-29.
  54. Ludyga N, Grunwald B, Azimzadeh O, Englert S, Hofler H, Tapio S, Aubele M. Nucleic acids from long-term preserved ffpe tissues are suitable for downstream analyses. *Virchows Arch* 2012;460:131-40.
  55. Gallegos Ruiz MI, Floor K, Rijmen F, Grunberg K, Rodriguez JA, Giaccone G. Egfr

- and k-ras mutation analysis in non-small cell lung cancer: Comparison of paraffin embedded versus frozen specimens. *Cellular oncology : the official journal of the International Society for Cellular Oncology* 2007;29:257-64.
56. Lin J, Kennedy SH, Svarovsky T, Rogers J, Kemnitz JW, Xu A, Zondervan KT. High-quality genomic DNA extraction from formalin-fixed and paraffin-embedded samples deparaffinized using mineral oil. *Analytical biochemistry* 2009;395:265-7.
  57. Carrick DM, Mehaffey MG, Sachs MC, Altekruze S, Camalier C, Chuaqui R, et al. Robustness of next generation sequencing on older formalin-fixed paraffin-embedded tissue. *PLoS one* 2015;10:e0127353.
  58. Hennig G, Gehrman M, Stropp U, Brauch H, Fritz P, Eichelbaum M, et al. Automated extraction of DNA and rna from a single formalin-fixed paraffin-embedded tissue section for analysis of both single-nucleotide polymorphisms and mrna expression. *Clinical chemistry* 2010;56:1845-53.
  59. Wong SQ, Li J, Tan AY, Vedururu R, Pang JM, Do H, et al. Sequence artefacts in a prospective series of formalin-fixed tumours tested for mutations in hotspot regions by massively parallel sequencing. *BMC medical genomics* 2014;7:23.
  60. Sakurada A, Lara-Guerra H, Liu N, Shepherd FA, Tsao MS. Tissue heterogeneity of egfr mutation in lung adenocarcinoma. *Journal of thoracic oncology : official publication of the International Association for the Study of Lung Cancer* 2008;3:527-9.
  61. Do H, Wong SQ, Li J, Dobrovic A. Reducing sequence artifacts in amplicon-based massively parallel sequencing of formalin-fixed paraffin-embedded DNA by enzymatic depletion of uracil-containing templates. *Clinical chemistry* 2013;59:1376-83.
  62. Chen L, Liu P, Evans TC, Jr., Ettwiller LM. DNA damage is a pervasive cause of

- sequencing errors, directly confounding variant identification. *Science* 2017;355:752-6.
63. Krokan HE, Drablos F, Slupphaug G. Uracil in DNA--occurrence, consequences and repair. *Oncogene* 2002;21:8935-48.
64. Oh E, Choi YL, Kwon MJ, Kim RN, Kim YJ, Song JY, et al. Comparison of accuracy of whole-exome sequencing with formalin-fixed paraffin-embedded and fresh frozen tissue samples. *PLoS one* 2015;10:e0144162.
65. Astolfi A, Urbini M, Indio V, Nannini M, Genovese CG, Santini D, et al. Whole exome sequencing (wes) on formalin-fixed, paraffin-embedded (ffpe) tumor tissue in gastrointestinal stromal tumors (gist). *BMC genomics* 2015;16:892.
66. Do H, Dobrovic A. Sequence artifacts in DNA from formalin-fixed tissues: Causes and strategies for minimization. *Clinical chemistry* 2015;61:64-71.
67. Ye X, Zhu ZZ, Zhong L, Lu Y, Sun Y, Yin X, et al. High t790m detection rate in tki-naive nslc with egfr sensitive mutation: Truth or artifact? *Journal of thoracic oncology : official publication of the International Association for the Study of Lung Cancer* 2013;8:1118-20.
68. Do H, Dobrovic A. Limited copy number-high resolution melting (lcn-hrm) enables the detection and identification by sequencing of low level mutations in cancer biopsies. *Molecular cancer* 2009;8:82.
69. Sah S, Chen L, Houghton J, Kemppainen J, Marko AC, Zeigler R, Latham GJ. Functional DNA quantification guides accurate next-generation sequencing mutation detection in formalin-fixed, paraffin-embedded tumor biopsies. *Genome medicine* 2013;5:77.
70. Suzuki T, Ohsumi S, Makino K. Mechanistic studies on depurination and apurinic site

- chain breakage in oligodeoxyribonucleotides. *Nucleic acids research* 1994;22:4997-5003.
71. Zsikla V, Baumann M, Cathomas G. Effect of buffered formalin on amplification of DNA from paraffin wax embedded small biopsies using real-time pcr. *Journal of clinical pathology* 2004;57:654-6.
  72. Heyn P, Stenzel U, Briggs AW, Kircher M, Hofreiter M, Meyer M. Road blocks on paleogenomes--polymerase extension profiling reveals the frequency of blocking lesions in ancient DNA. *Nucleic acids research* 2010;38:e161.
  73. Sikorsky JA, Primerano DA, Fenger TW, Denvir J. DNA damage reduces taq DNA polymerase fidelity and pcr amplification efficiency. *Biochemical and biophysical research communications* 2007;355:431-7.
  74. Anderson S, Bloom KJ, Vallera DU, Rueschoff J, Meldrum C, Schilling R, et al. Multisite analytic performance studies of a real-time polymerase chain reaction assay for the detection of braf v600e mutations in formalin-fixed, paraffin-embedded tissue specimens of malignant melanoma. *Archives of pathology & laboratory medicine* 2012;136:1385-91.
  75. Fraenkel-Conrat H, Olcott HS. The reaction of formaldehyde with proteins; cross-linking between amino and primary amide or guanidyl groups. *Journal of the American Chemical Society* 1948;70:2673-84.
  76. Speit G, Schutz P, Merk O. Induction and repair of formaldehyde-induced DNA-protein crosslinks in repair-deficient human cell lines. *Mutagenesis* 2000;15:85-90.
  77. Merk O, Speit G. Significance of formaldehyde-induced DNA-protein crosslinks for mutagenesis. *Environmental and molecular mutagenesis* 1998;32:260-8.
  78. McGhee JD, von Hippel PH. Formaldehyde as a probe of DNA structure. Ii. Reaction

- with endocyclic imino groups of DNA bases. *Biochemistry* 1975;14:1297-303.
79. Inoue T, Nabeshima K, Kataoka H, Koono M. Feasibility of archival non-buffered formalin-fixed and paraffin-embedded tissues for pcr amplification: An analysis of resected gastric carcinoma. *Pathology international* 1996;46:997-1004.
80. Hewitt SM, Lewis FA, Cao Y, Conrad RC, Cronin M, Danenberg KD, et al. Tissue handling and specimen preparation in surgical pathology: Issues concerning the recovery of nucleic acids from formalin-fixed, paraffin-embedded tissue. *Archives of pathology & laboratory medicine* 2008;132:1929-35.
81. Ofner R, Ritter C, Ugurel S, Cerroni L, Stiller M, Bogenrieder T, et al. Non-reproducible sequence artifacts in ffpe tissue: An experience report. *Journal of cancer research and clinical oncology* 2017.
82. Shi SR, Cote RJ, Wu L, Liu C, Datar R, Shi Y, et al. DNA extraction from archival formalin-fixed, paraffin-embedded tissue sections based on the antigen retrieval principle: Heating under the influence of ph. *The journal of histochemistry and cytochemistry : official journal of the Histochemistry Society* 2002;50:1005-11.
83. Gillio-Tos A, De Marco L, Fiano V, Garcia-Bragado F, Dikshit R, Boffetta P, Merletti F. Efficient DNA extraction from 25-year-old paraffin-embedded tissues: Study of 365 samples. *Pathology* 2007;39:345-8.
84. Do H, Dobrovic A. Dramatic reduction of sequence artefacts from DNA isolated from formalin-fixed cancer biopsies by treatment with uracil- DNA glycosylase. *Oncotarget* 2012;3:546-58.
85. Schormann N, Ricciardi R, Chattopadhyay D. Uracil-DNA glycosylases-structural and functional perspectives on an essential family of DNA repair enzymes. *Protein science : a publication of the Protein Society* 2014;23:1667-85.

86. Jackson V. Studies on histone organization in the nucleosome using formaldehyde as a reversible cross-linking agent. *Cell* 1978;15:945-54.
87. Sepp R, Szabo I, Uda H, Sakamoto H. Rapid techniques for DNA extraction from routinely processed archival tissue for use in pcr. *Journal of clinical pathology* 1994;47:318-23.
88. Wu L, Patten N, Yamashiro CT, Chui B. Extraction and amplification of DNA from formalin-fixed, paraffin-embedded tissues. *Applied immunohistochemistry & molecular morphology : AIMM* 2002;10:269-74.
89. Aung KL, Board RE, Ellison G, Donald E, Ward T, Clack G, et al. Current status and future potential of somatic mutation testing from circulating free DNA in patients with solid tumours. *The HUGO journal* 2010;4:11-21.
90. Alix-Panabieres C, Pantel K. Circulating tumor cells: Liquid biopsy of cancer. *Clinical chemistry* 2013;59:110-8.
91. Kimura H, Kasahara K, Kawaishi M, Kunitoh H, Tamura T, Holloway B, Nishio K. Detection of epidermal growth factor receptor mutations in serum as a predictor of the response to gefitinib in patients with non-small-cell lung cancer. *Clinical cancer research : an official journal of the American Association for Cancer Research* 2006;12:3915-21.
92. Horiike A, Kimura H, Nishio K, Ohyanagi F, Satoh Y, Okumura S, et al. Detection of epidermal growth factor receptor mutation in transbronchial needle aspirates of non-small cell lung cancer. *Chest* 2007;131:1628-34.
93. He C, Liu M, Zhou C, Zhang J, Ouyang M, Zhong N, Xu J. Detection of epidermal growth factor receptor mutations in plasma by mutant-enriched pcr assay for prediction of the response to gefitinib in patients with non-small-cell lung cancer. *International*

- journal of cancer 2009;125:2393-9.
94. Taniguchi K, Uchida J, Nishino K, Kumagai T, Okuyama T, Okami J, et al. Quantitative detection of egfr mutations in circulating tumor DNA derived from lung adenocarcinomas. *Clinical cancer research : an official journal of the American Association for Cancer Research* 2011;17:7808-15.
  95. Maheswaran S, Sequist LV, Nagrath S, Ulkus L, Brannigan B, Collura CV, et al. Detection of mutations in egfr in circulating lung-cancer cells. *The New England journal of medicine* 2008;359:366-77.
  96. Yung TK, Chan KC, Mok TS, Tong J, To KF, Lo YM. Single-molecule detection of epidermal growth factor receptor mutations in plasma by microfluidics digital pcr in non-small cell lung cancer patients. *Clinical cancer research : an official journal of the American Association for Cancer Research* 2009;15:2076-84.
  97. Oxnard GR, Paweletz CP, Kuang Y, Mach SL, O'Connell A, Messineo MM, et al. Noninvasive detection of response and resistance in egfr-mutant lung cancer using quantitative next-generation genotyping of cell-free plasma DNA. *Clinical cancer research : an official journal of the American Association for Cancer Research* 2014;20:1698-705.
  98. Nakahara Y, Takagi Y, Hosomi Y, Kagei A, Yamamoto T, Sawada T, et al. Noninvasive monitoring of the genetic evolution of egfr-mutant non-small-cell lung cancer by analyzing circulating tumor DNA during combination chemotherapy with gefitinib and pemetrexed or s-1. *OncoTargets and therapy* 2016;9:5287-95.
  99. Denis MG, Vallee A, Theoleyre S. Egfr t790m resistance mutation in non small-cell lung carcinoma. *Clinica chimica acta; international journal of clinical chemistry* 2015;444:81-5.

100. Ulivi P, Mercatali L, Casoni GL, Scarpi E, Bucchi L, Silvestrini R, et al. Multiple marker detection in peripheral blood for nsclc diagnosis. *PloS one* 2013;8:e57401.
101. Catarino R, Coelho A, Araujo A, Gomes M, Nogueira A, Lopes C, Medeiros R. Circulating DNA: Diagnostic tool and predictive marker for overall survival of nsclc patients. *PloS one* 2012;7:e38559.
102. Jahr S, Hentze H, Englisch S, Hardt D, Fackelmayer FO, Hesch RD, Knippers R. DNA fragments in the blood plasma of cancer patients: Quantitations and evidence for their origin from apoptotic and necrotic cells. *Cancer research* 2001;61:1659-65.
103. Sorenson GD. Detection of mutated kras2 sequences as tumor markers in plasma/serum of patients with gastrointestinal cancer. *Clinical cancer research : an official journal of the American Association for Cancer Research* 2000;6:2129-37.
104. Sorenson GD. A review of studies on the detection of mutated kras2 sequences as tumor markers in plasma/serum of patients with gastrointestinal cancer. *Annals of the New York Academy of Sciences* 2000;906:13-6.
105. Mouliere F, Robert B, Arnau Peyrotte E, Del Rio M, Ychou M, Molina F, et al. High fragmentation characterizes tumour-derived circulating DNA. *PloS one* 2011;6:e23418.
106. Thierry AR, Mouliere F, Gongora C, Ollier J, Robert B, Ychou M, et al. Origin and quantification of circulating DNA in mice with human colorectal cancer xenografts. *Nucleic acids research* 2010;38:6159-75.
107. Frattini M, Gallino G, Signoroni S, Balestra D, Lusa L, Battaglia L, et al. Quantitative and qualitative characterization of plasma DNA identifies primary and recurrent colorectal cancer. *Cancer letters* 2008;263:170-81.
108. Chen XQ, Stroun M, Magnenat JL, Nicod LP, Kurt AM, Lyautey J, et al. Microsatellite alterations in plasma DNA of small cell lung cancer patients. *Nature medicine*

- 1996;2:1033-5.
109. Misale S, Yaeger R, Hobor S, Scala E, Janakiraman M, Liska D, et al. Emergence of kras mutations and acquired resistance to anti-egfr therapy in colorectal cancer. *Nature* 2012;486:532-6.
  110. Chan KC, Jiang P, Zheng YW, Liao GJ, Sun H, Wong J, et al. Cancer genome scanning in plasma: Detection of tumor-associated copy number aberrations, single-nucleotide variants, and tumoral heterogeneity by massively parallel sequencing. *Clinical chemistry* 2013;59:211-24.
  111. Diehl F, Li M, Dressman D, He Y, Shen D, Szabo S, et al. Detection and quantification of mutations in the plasma of patients with colorectal tumors. *Proceedings of the National Academy of Sciences of the United States of America* 2005;102:16368-73.
  112. Leary RJ, Sausen M, Kinde I, Papadopoulos N, Carpten JD, Craig D, et al. Detection of chromosomal alterations in the circulation of cancer patients with whole-genome sequencing. *Science translational medicine* 2012;4:162ra54.
  113. Forshew T, Murtaza M, Parkinson C, Gale D, Tsui DW, Kaper F, et al. Noninvasive identification and monitoring of cancer mutations by targeted deep sequencing of plasma DNA. *Science translational medicine* 2012;4:136ra68.
  114. Ishii H, Azuma K, Sakai K, Kawahara A, Yamada K, Tokito T, et al. Digital pcr analysis of plasma cell-free DNA for non-invasive detection of drug resistance mechanisms in egfr mutant nsccl: Correlation with paired tumor samples. *Oncotarget* 2015;6:30850-8.
  115. Gilbert Y, Le Bihan Y, Aubry G, Veillette M, Duchaine C, Lessard P. Microbiological and molecular characterization of denitrification in biofilters treating pig manure. *Bioresource technology* 2008;99:4495-502.
  116. Dong L, Cordova-Kreylos AL, Yang J, Yuan H, Scow KM. Humic acids buffer the

- effects of urea on soil ammonia oxidizers and potential nitrification. *Soil biology & biochemistry* 2009;41:1612-21.
117. Hindson BJ, Ness KD, Masquelier DA, Belgrader P, Heredia NJ, Makarewicz AJ, et al. High-throughput droplet digital pcr system for absolute quantitation of DNA copy number. *Analytical chemistry* 2011;83:8604-10.
  118. Pinheiro LB, Coleman VA, Hindson CM, Herrmann J, Hindson BJ, Bhat S, Emslie KR. Evaluation of a droplet digital polymerase chain reaction format for DNA copy number quantification. *Analytical chemistry* 2012;84:1003-11.
  119. Didelot A, Kotsopoulos SK, Lupo A, Pekin D, Li X, Atochin I, et al. Multiplex picoliter-droplet digital pcr for quantitative assessment of DNA integrity in clinical samples. *Clinical chemistry* 2013;59:815-23.
  120. Hindson CM, Chevillet JR, Briggs HA, Gallichotte EN, Ruf IK, Hindson BJ, et al. Absolute quantification by droplet digital pcr versus analog real-time pcr. *Nature methods* 2013;10:1003-5.
  121. Taly V, Pekin D, Benhaim L, Kotsopoulos SK, Le Corre D, Li X, et al. Multiplex picodroplet digital pcr to detect kras mutations in circulating DNA from the plasma of colorectal cancer patients. *Clinical chemistry* 2013;59:1722-31.
  122. Beaver JA, Jelovac D, Balukrishna S, Cochran RL, Croessmann S, Zabransky DJ, et al. Detection of cancer DNA in plasma of patients with early-stage breast cancer. *Clinical cancer research : an official journal of the American Association for Cancer Research* 2014;20:2643-50.
  123. Ono A, Kenmotsu H, Watanabe M, Serizawa M, Mori K, Imai H, et al. Mutant allele frequency predicts the efficacy of egfr-tkis in lung adenocarcinoma harboring the l858r mutation. *Annals of oncology : official journal of the European Society for Medical*

- Oncology / ESMO 2014;25:1948-53.
124. Rajasekaran N, Oh MR, Kim SS, Kim SE, Kim YD, Choi HJ, et al. Employing digital droplet pcr to detect braf v600e mutations in formalin-fixed paraffin-embedded reference standard cell lines. *Journal of visualized experiments : JoVE* 2015.
  125. Permenter J, Ishwar A, Rounsavall A, Smith M, Faske J, Sailey CJ, Alfaro MP. Quantitative analysis of genomic DNA degradation in whole blood under various storage conditions for molecular diagnostic testing. *Molecular and cellular probes* 2015;29:449-53.
  126. O'Donnell P, Ferguson J, Shyu J, Current R, Rehage T, Tsai J, et al. Analytic performance studies and clinical reproducibility of a real-time pcr assay for the detection of epidermal growth factor receptor gene mutations in formalin-fixed paraffin-embedded tissue specimens of non-small cell lung cancer. *BMC cancer* 2013;13:210.
  127. Armbruster DA, Pry T. Limit of blank, limit of detection and limit of quantitation. *The Clinical biochemist Reviews* 2008;29 Suppl 1:S49-52.
  128. Tsiatis AC, Norris-Kirby A, Rich RG, Hafez MJ, Gocke CD, Eshleman JR, Murphy KM. Comparison of sanger sequencing, pyrosequencing, and melting curve analysis for the detection of kras mutations: Diagnostic and clinical implications. *The Journal of molecular diagnostics : JMD* 2010;12:425-32.
  129. Walsh K, Wallace WA, Butler R, Mackean MJ, Harrison DJ, Stirling D, Oniscu A. A cautionary lesson on the use of targeted methods for egfr mutation analysis: A case report. *Journal of clinical pathology* 2014;67:734-5.
  130. Serizawa M, Yokota T, Hosokawa A, Kusafuka K, Sugiyama T, Tsubosa Y, et al. The efficacy of uracil DNA glycosylase pretreatment in amplicon-based massively parallel

- sequencing with DNA extracted from archived formalin-fixed paraffin-embedded esophageal cancer tissues. *Cancer Genet* 2015;208:415-27.
131. Bourgon R, Lu S, Yan Y, Lackner MR, Wang W, Weigman V, et al. High-throughput detection of clinically relevant mutations in archived tumor samples by multiplexed PCR and next-generation sequencing. *Clinical cancer research : an official journal of the American Association for Cancer Research* 2014;20:2080-91.
  132. Kim S, Park C, Ji Y, Kim DG, Bae H, van Vrancken M, et al. Deamination effects in formalin-fixed, paraffin-embedded tissue samples in the era of precision medicine. *The Journal of molecular diagnostics : JMD* 2017;19:137-46.
  133. Hadd AG, Houghton J, Choudhary A, Sah S, Chen L, Marko AC, et al. Targeted, high-depth, next-generation sequencing of cancer genes in formalin-fixed, paraffin-embedded and fine-needle aspiration tumor specimens. *The Journal of molecular diagnostics : JMD* 2013;15:234-47.
  134. White RA, 3rd, Blainey PC, Fan HC, Quake SR. Digital PCR provides sensitive and absolute calibration for high throughput sequencing. *BMC genomics* 2009;10:116.
  135. Lopez-Rios F, Angulo B, Gomez B, Mair D, Martinez R, Conde E, et al. Comparison of molecular testing methods for the detection of EGFR mutations in formalin-fixed paraffin-embedded tissue specimens of non-small cell lung cancer. *Journal of clinical pathology* 2013;66:381-5.
  136. Ardakani NM, Giardina T, Grieco-Iacopetta F, Tesfai Y, Carrello A, Taylor J, et al. Detection of epidermal growth factor receptor mutations in lung adenocarcinoma: Comparing cobas 4800 EGFR assay with Sanger bidirectional sequencing. *Clinical lung cancer* 2016;17:e113-e9.
  137. Cataldo VD, Gibbons DL, Perez-Soler R, Quintas-Cardama A. Treatment of non-small-

- cell lung cancer with erlotinib or gefitinib. *The New England journal of medicine* 2011;364:947-55.
138. Giordano P, Manzo A, Montanino A, Costanzo R, Sandomenico C, Piccirillo MC, et al. Afatinib: An overview of its clinical development in non-small-cell lung cancer and other tumors. *Critical reviews in oncology/hematology* 2016;97:143-51.
  139. Rosell R, Moran T, Queralt C, Porta R, Cardenal F, Camps C, et al. Screening for epidermal growth factor receptor mutations in lung cancer. *The New England journal of medicine* 2009;361:958-67.
  140. Chi A, Remick S, Tse W. Egfr inhibition in non-small cell lung cancer: Current evidence and future directions. *Biomarker research* 2013;1:2.
  141. Cross DA, Ashton SE, Ghiorghiu S, Eberlein C, Nebhan CA, Spitzler PJ, et al. Azd9291, an irreversible egfr tki, overcomes t790m-mediated resistance to egfr inhibitors in lung cancer. *Cancer discovery* 2014;4:1046-61.
  142. Walter AO, Sjin RT, Haringsma HJ, Ohashi K, Sun J, Lee K, et al. Discovery of a mutant-selective covalent inhibitor of egfr that overcomes t790m-mediated resistance in nscl. *Cancer discovery* 2013;3:1404-15.

## DNA의 특질 분석을 이용한 드롭렛 디지털 PCR 기반의

### *EGFR* 돌연변이 검출 방법 연구

서울대학교 대학원

약학과 병태생리학 전공

김 성수

포르말린 고정 파라핀 내장 조직 (formalin-fixed paraffin-embedded tissue, FFPET) 샘플은 임상 연구와 체외 진단분야에 있어서 매우 중요한 자원이다. 하지만, 포르말린 고정 파라핀 내장 조직 샘플을 이용한 유전자 돌연변이 검출은 포르말린 고정 파라핀 내장 조직의 처리 과정 및 다양한 보존기간과 품질로부터 기인하는 문제점으로 인하여 부정확한 결과를 초래한다.

본 연구는 315 개의 비소세포 폐암 환자의 FFPET-DNA 샘플을 사용한 전임상 연구를 통하여 ddPCR 기반 (GenesWell™ ddEGFR Mutation test, “ddEGFR test”)과 Real-time PCR 기반의 EGFR 돌연변이 테스트 (cobas® EGFR Mutation test, “cobas EGFR test”) 를 비교함으로써 PCR에 적합한 최소 DNA 품질을 반영하는 샘플 기준을 확립했다. 확립된 샘플 기준을 사용하여 171개의 비소세포 폐암 환자의

FFPET-DNA 샘플을 사용한 독립된 후향적 임상 연구에서 ddEGFR 및 cobas EGFR test에 대한 비교 임상 및 샘플 기준의 임상적 유용성을 평가 하였다.

전임상 연구를 통하여 FFPET-DNA의 샘플 기준 (iQC index  $\geq 0.5$ , iQC copies 500 이상, FFPET-DNA가 3.3 ng [1000 GE]을 사용)을 확립 하였고, 이는 Input DNA 농도를 기준으로 절반 이상이 증폭 가능함을 의미한다. iQC index를 적용한 전임상 샘플 150개를 재 분석 하였고, ddEGFR test와 Cobas EGFR test의 일치율이 92.67%로 증가 하였다 (전임상 샘플 315개 전체 일치율 = 78.10%). iQC index를 기반으로, 독립적인 후향적 임상 시험을 실시 하였고, ddEGFR test와 Cobas EGFR test의 일치율이 매우 높게 관찰 되었다 (OPA = 92.98%). 또한, cobas EGFR test에서 검출하지 못하는 샘플들을 ddEGFR test를 사용하여 민감하게 검출 하였다.

iQC index는 FFPET-DNA 의 품질을 반영하는 신뢰할 수 있는 지표이며 매우 낮은 품질의 샘플로부터 기인하는 부정확한 진단을 방지할 수 있다. 또한, 기허가 받은 cobas EGFR test와 비교하여, ddEGFR test는 우수한 분석적 성능 및 적어도 동등한 임상적 성능을 나타냈다.

**주요어:** 상피세포성장수용체; 포르말린 고정 파라핀 내장 조직; 핵산 무결성 지수; 드롭렛 디지털 중합효소 연쇄반응; 유전자 돌연변이 검출; 핵산 내부 품질 지수

**학번:** 2011-30502