



저작자표시-비영리-변경금지 2.0 대한민국

이용자는 아래의 조건을 따르는 경우에 한하여 자유롭게

- 이 저작물을 복제, 배포, 전송, 전시, 공연 및 방송할 수 있습니다.

다음과 같은 조건을 따라야 합니다:



저작자표시. 귀하는 원저작자를 표시하여야 합니다.



비영리. 귀하는 이 저작물을 영리 목적으로 이용할 수 없습니다.



변경금지. 귀하는 이 저작물을 개작, 변형 또는 가공할 수 없습니다.

- 귀하는, 이 저작물의 재이용이나 배포의 경우, 이 저작물에 적용된 이용허락조건을 명확하게 나타내어야 합니다.
- 저작권자로부터 별도의 허가를 받으면 이러한 조건들은 적용되지 않습니다.

저작권법에 따른 이용자의 권리는 위의 내용에 의하여 영향을 받지 않습니다.

이것은 [이용허락규약\(Legal Code\)](#)을 이해하기 쉽게 요약한 것입니다.

[Disclaimer](#)

치의과학박사 학위논문

Sandblasting 후 산부식 처리한
티타늄 디스크 상의
Porphyromonas gingivalis
바이오필름에 대한
Atmospheric-Pressure Plasma
Jet의 살균 효과

The bactericidal effect of an
atmospheric-pressure plasma jet on
Porphyromonas gingivalis biofilms on
sandblasted and acid-etched titanium discs

2020년 2월

서울대학교 대학원
치의과학과 치의과학 전공
이 지 윤

Abstract

The bactericidal effect of an
atmospheric-pressure plasma jet
on *Porphyromonas gingivalis*
biofilms on sandblasted and
acid-etched titanium discs

JIYOON LEE D.D.S., M.S.D.

Program in Periodontology

Department of Dental Science

Graduate school, Seoul National University

(Directed by Professor Seol, Yang-Jo D.D.S., M.S.D.,
Ph.D.)

Direct application of atmospheric-pressure plasma jets (APPJs) has
been established as an effective method of microbial decontamination.

This study aimed to investigate the bactericidal effect of direct application of an APPJ using helium gas (He-APPJ) on *Porphyromonas gingivalis* biofilms on sandblasted and acid-etched (SLA) titanium discs.

On the SLA discs covered by *P. gingivalis* biofilms, an He-APPJ was applied at 3 or 5 minutes. To evaluate the effect of the plasma itself, the He gas without plasma generation was also applied at 3 or 5 minutes and untreated discs were used as the control group. The bactericidal effect of the He-APPJ was determined by the number of colony-forming units. Bacterial viability was observed by confocal laser scanning microscopy (CLSM), and bacterial morphology was examined by scanning electron microscopy (SEM).

As the plasma treatment time increased, the amount of *P. gingivalis* decreased, and the difference was statistically significant. In the SEM images, compared to the control group, the bacterial biofilm structure on SLA discs treated by the He-APPJ for more than 3 minutes was destroyed. The CLSM images also showed consistent results with the results of cell assay and SEM. Even in sites distant from the area of direct He-APPJ exposure, decontamination effects were observed in both SEM and CLSM images.

He-APPJ application was effective in removing *P. gingivalis* biofilm on SLA titanium discs in an *in vitro* experiment.

Keywords: Bacterial Load; Confocal Microscopy; Plasma Gases;
Porphyromonas gingivalis; Scanning Electron Microscopy; Titanium
Student Number: 2010-30638

Contents

Abstract (In English)	
I . INTRODUCTION	1
II. MATERIALS AND METHODS.....	4
III. RESULTS	8
IV. DISCUSSION	10
V. CONCLUSIONS	16
VI. REFERENCES	17
Table and Figures	23
Abstract (In Korean)	31

I . INTRODUCTION

Since their introduction in 1965, dental implants have been widely used to replace missing teeth. However, the accumulation of bacteria around dental implants results in inflammatory disease, such as peri-implant mucositis and peri-implantitis [1]. Moreover, peri-implantitis shows refractory features, and one of the reasons is the difficulties in decontaminating the implant surface after it is covered by bacterial deposits [2].

No gold-standard decontamination procedure has yet emerged for implant surfaces. In several *in vitro* and animal experiments, mechanical debridement using ultrasonic scalers and curettes, laser treatment, air-powder abrasive treatment have been applied with or without antiseptics or antibiotics [3-5]. However, mechanical debridement of implants with microthreads using conventional curettes is difficult because of the narrow interthread space. Air-powder abrasive treatments are effective in microthreaded implants, although alterations of the implant surface have been observed in some cases [6,7].

Atmospheric-pressure plasma jets (APPJs) are a candidate for decontaminating implant surfaces. Plasma is one of the 4 fundamental states of matter and can be artificially generated by adding energy to a gas in an electromagnetic field. If the gas is electrically charged, it radiates radicals, charged species and metastable species, which induce bactericidal effects *in vitro* [8,9]. Generally, because the energy

heating and energy transition processes increase the temperature of the target, the applicability of plasma under *in vivo* conditions is limited. However, since APPJs generate ions and neutral species at cold temperatures while electrons are hot along only a short mean free path, APPJs do not cause any thermal damage to their surroundings [10]. Due to this feature, APPJs have been applied to sterilize materials in biological research and medical fields [11]. In dentistry, APPJs are used to control endodontic infections, remove dental decay, increase the bonding efficiency of composite resin, and even to bleach teeth [12–15]. In the periodontal field, argon (Ar) plasma treatment on implant abutments was found to improve tissue adaptation in patients with periodontal disease and a thin biotype [16]. For implant treatment, Koban et al. [17] applied an APPJ to machined titanium discs contaminated by *Streptococcus mutans* or multiple species from human saliva and observed a superior antiseptic effect of the APPJ compared to that of chlorhexidine rinse. Recently, Carreiro et al. [18] evaluated an Ar gas - based APPJ treatment on sandblasted and acid-etched (SLA) surfaces, which are among the most widely used surfaces of currently available commercial dental implants. They showed the possibility of using APPJs on SLA surfaces as a decontamination method for *P. gingivalis* biofilms.

The preliminary study to evaluate the surface change of titanium discs after He-APPJ treatment, differences were not found in SEM images of the He-APPJ-treated discs for 3 or 5 minutes with the untreated control in machined titanium discs (Figure 1-A). The evaluation of the roughness of the SLA titanium discs after He-APPJ

for 0, 3 or 5 minutes also showed no marked changes regardless of the treatment time (Figure 1-B). The effect of pre-treatment on growth of *P. gingivalis* and osteoblasts were assessed by the treatment of sterilized discs with He-APPJ for 0, 3 or 5 minutes followed by cultivating *P. gingivalis* or osteoblast. The result showed that plasma pre-treatment did not affect the growth of *P. gingivalis* or osteoblasts (Figure 2).

In other study, we revealed that in APPJs, helium (He) gas is superior to Ar gas because the energy of metastable species of He is higher than that of Ar, so He gas-based APPJ (He-APPJ) should be investigated for clinical applications [19]. This study aimed to evaluate a He-APPJ as a potential decontamination method for peri-implantitis treatment. For this purpose, the bactericidal effect of the He-APPJ was tested using *P. gingivalis* biofilms on SLA titanium discs.

II. MATERIALS AND METHODS

1. Disc preparation

SLA titanium discs of 10 mm in diameter and 1 mm in thickness (Osstem Implant Co., Ltd., Busan, Korea) were used. Each SLA disc was cut evenly into 4 quarter circles for the experiments. Before cell culture or bacterial biofilm formation, 1-quarter of a SLA disc was placed into each well. The discs were rinsed with ethanol and sterilized with ethylene oxide gas.

2. Bacterial strain and biofilm formation

P. gingivalis (American Type Culture Collection 33277) was used for biofilm formation. Bacteria were grown in brain heart infusion growth medium (BHI medium) supplemented with hemin and menadione and incubated in an anaerobic environment (GasPak-EZ anaerobic container system; Becton Dickinson Microbiology Systems, Cockeysville, MD, USA) for 3 days. The bacterial concentration was adjusted by spectrophotometry in BHI medium containing 1×10^6 colony-forming units (CFU)/mL. For the development of biofilms, the sterilized discs were placed in 4-well polystyrene plates. The bacteria were inoculated and incubated in anaerobic environment for 6 days until confluency was attained, with a change to fresh medium at

48-hour intervals after 72 hours of incubation.

3. He-APPJ application

The design of the apparatus for generating an APPJ was described by Yoon et al. [19]. The APPJ source consists of a very-low-frequency (VLF)-driven 0.2 mm tungsten electrode surrounded by an alumina tube with an outer diameter of 0.6 mm and a coaxial quartz gas guiding tube with an inner diameter of 1 mm and a thickness of 1 mm. He was used as the discharge gas, which was injected into the gap between the alumina and quartz tubes and then released into the open air. The gas flow rate was maintained at 5 standard liters per minute and controlled by a mass flow controller (TN 2900, Brooks Instrument, Hatfield, PA, USA). The VLF frequency was generated using a function generator connected to a voltage amplifier and set to 7 kV under a pulse of 10 kHz (MF plasma power supply, Dawonsys, Ansan, Korea). The plasma device was fixed on a stand to prevent the device from moving and to maintain the distance between the plasma tip and specimens (Figure 3). The distance from the device tip to the specimens was 30 mm and adjusted to keep the target located at the end of the plasma plume.

The experimental groups were divided as follows: 1) the untreated control group, 2) the He without plasma generation group, and 3) the He-APPJ group. Each group was treated with APPJ or He for 3 or 5

minutes, except for control. In one of our preliminary study conducted by this He-APPJ to determine the time required for proper sterilization, 1-minute treatment was assessed not to be sufficient to have adequate sterilizing effect. Therefore, the treatment time was determined to 3 and 5 minutes.

4. Bacterial count assay

After the plasma treatment, the bacteria on each disc (n=5 per group) were transferred to 0.5 mL of a dispersion solution and then dispersed by vortexing for 1 minute to remove the remaining bacteria. The solutions were serially diluted and plated on Brucella agar plates (Hanil Komed Co., Seongnam, Korea). The bacteria were incubated anaerobically at 37°C for 7 days, and the number of CFUs was counted. All experiments were performed in triplicate.

5. Confocal laser scanning microscopy (CLSM) examinations

To evaluate cell viability, the bacteria on each disc after the abovementioned plasma treatment (n=2 per group) were stained with LIVE/DEAD® BacLight™ bacterial viability kits (Thermo Fisher Scientific, Waltham, MA, USA) and examined using CLSM (Carl Zeiss MicroImaging GmbH, Jena, Germany). Viable bacteria with

intact membranes showed green fluorescence, and dead bacteria with damaged membranes showed red fluorescence.

6. Scanning electron microscopy (SEM) examinations

The structural and morphological changes of the bacteria after the abovementioned plasma treatment (n=2 per group) were evaluated using SEM (HITACHI S-4700, Tokyo, Japan). The plasma-treated discs were fixed with 2.5% glutaraldehyde for 2 hours and rinsed in a secondary fixative of osmium tetroxide and cacodylate solution, followed by rinsing and dehydration in ethanol. The discs were then dried and coated with a thin layer of platinum. Each specimen was examined with SEM.

7. Statistical analysis

The data are presented as the mean \pm standard deviation. Statistical analysis was performed using 1-way analysis of variance for comparison of the control, 3-minute, and 5-minute treatment groups and the Student *t*-test for comparisons between treatment groups. *P* values <0.05 were considered to indicate statistical significance.

III. RESULTS

1. Changes in the number of *P. gingivalis* CFUs after He-APPJ treatment

As the SLA surface treatment time increased, the number of *P. gingivalis* CFUs prominently decreased in both the He gas and He-APPJ groups, and the changes were statistically significant (Table 1). After 5 minutes of treatment, the number of *P. gingivalis* CFUs in the He-APPJ group was below the threshold level (1×10^3) and was significantly different from that in the He gas group ($P=0.033$).

2. *P. gingivalis* biofilm and bacterial viability after He-APPJ treatment

In the SEM images, no treatment group showed multiple, thick layers of clustered bacterial biofilm on the titanium disc surface (Figure 4). In the He gas - only treatment group, remaining bacterial biofilm was observed in both the central and marginal areas in the 3-minute treatment group, while bacterial rods had been eliminated from the central area, but not the marginal area, of the 5-minute treatment group. However, the He-APPJ group showed a clear SLA

surface and destroyed bacterial biofilm structure, and no intact bacteria were detected in the central or marginal areas of both the 3-minute and 5-minute treatment groups.

In the CLSM examinations, heavy live bacterial accumulation on the discs was observed in the control group (Figure 5). In the He gas - only group, the amount of bacteria dramatically decreased in the central area. However, in the marginal area, 3 minutes of treatment did not induce a reduction in the amount of bacteria. The amount of bacteria decreased in the 5-minute treatment group, although a moderate amount of live bacteria were still observed. Among the He gas - only treatment groups, the amount of bacteria also remarkably decreased in the central area compared to the marginal area in both the 3-minute and 5-minute treatment groups. In the He-APPJ treatment group, only a few viable bacteria were observed in the central area of the 3-minute treatment group. However, in the marginal area, a moderate amount of *P. gingivalis* remained; a mixture of live and dead bacteria was observed in the marginal area of the 3-minute treatment group, whereas most of the remaining bacteria were dead in the 5-minute treatment group.

IV. DISCUSSION

The treatment of peri-implantitis is mainly focused on the removal of bacterial biofilms. However, peri-implant biofilms are more difficult to remove than periodontal biofilms, because both the implant design and rough surface texture impede the removal of the biofilm. The decontamination of rough surfaces, such as SLA surfaces, is more difficult than that of machined surfaces, which is one of the main reasons why special care is needed to treat the disease, as shown in the spontaneous peri-implantitis model proposed by Berglundh et al. [20]. In this study, a He-APPJ was used to decontaminate the SLA surface of titanium discs because blowing He gas, with or without plasma, can be done without altering the surface texture. The bactericidal effect of the He-APPJ on *P. gingivalis* biofilms on titanium discs was confirmed using several methods.

Microbial decontamination by APPJs is mediated by radicals that destroy the bacterial cell wall [21]. The mechanism of plasma-induced cell membrane rupture is mediated by the action of charged particles and reactive oxygen and nitrogen species (RONS). Charged particles and RONS produced by plasma can damage some small surface irregularities of the bacterial membrane and cause the membrane to rupture [8,22,23]. Because RONS are radicals produced by APPJs [24,25], and this investigation was performed in ambient air which is a good source of oxygen and nitrogen, the bactericidal effect of RONS cannot be excluded. These RONS attack membrane lipids,

which are vulnerable macromolecules of the cell, resulting in the destruction of the membrane [26]. This APPJ system has additional benefits for the treatment of peri-implantitis, as APPJ has shown the capacity for high levels of radical generation in humid environments and with fixed targets [27]. However, the generation of RONS is sensitive to the discharge conditions [21]. Unlike the temperature of Ar gas, which reaches 45°C during the plasma treatment, that of He gas does not exceed 25°C [17]. Previously, we conducted experiments to compare Ar-APPJ to He-APPJ, and observed that He-APPJ yielded more metastable species, such as NO, O, and O₃, than Ar did and achieved bacterial inactivation in significantly less time than Ar-APPJ did (1.63 minutes for He vs. 12.1 minutes for Ar) [19]. Therefore, Ar-APPJ and He-APPJ show significantly different bactericidal effects, and He-APPJ is considered to be an appropriate application for the decontamination of titanium surfaces.

To determine the proper treatment time of He-APPJ, we designed an experimental group with 3- and 5-minute treatments and observed a significant decrease in the number of bacteria as the treatment time increased. The treatment time required to completely eradicate bacteria depends on the power setting of the plasma, discharge gas, bacterial type, type of medium, and the distance between the cell and plasma tip [24,28]. In addition to this study, Miura et al. [28] also tested the bactericidal effect of APPJ on *P. gingivalis* with different plasma durations and distances on machined surfaces in an *in vitro* setting. However, they did not suggest a proper treatment time for *P. gingivalis*. Carreiro et al. [18] applied an Ar-APPJ to an SLA surface

for 1 and 3 minutes and reported a reduction in *P. gingivalis* colonization. However, the plasma treatment was ineffective compared to chlorhexidine irrigation. Shi et al. [29] performed plasma treatment on contaminated implants in an experimental peri-implantitis model with beagle dogs and observed an increase in bone formation and re-osseointegration in the group treated by plasma for 3 minutes compared to that in the control group, which was irrigated with chlorhexidine and saline for 3 minutes. The authors demonstrated a decrease in periodontal pathogenic bacterial load after plasma treatment that was significantly different for 3 months after treatment. In this study, the 3-minute He-APPJ treatment was effective for debridement of the SLA surface, but the 5-minute treatment completely killed *P. gingivalis*, including both directly plasma-exposed and unexposed sites. However, there is no consensus on the most suitable plasma treatment time for achieving bactericidal effects because of limited evidence on the plasma type, settings, and treatment parameters. *P. gingivalis* is a representative periodontal pathogen that contributes to destructive periodontitis and peri-implantitis. *P. gingivalis* has a unique cell membrane structure, with polysaccharide capsular materials that generally show higher resistance than those of other Gram-negative bacteria [30,31]. Furthermore, dental microorganisms usually exist in the form of biofilm in the periodontal pocket, and bacteria in biofilms are usually stable and resistant to outer attack because they produce a polymeric matrix [30,32]. These features make *P. gingivalis* very resistant to antibacterial treatment. Considering those conditions, the relatively

longer killing time of *P. gingivalis* in this study can be compared to that of other Gram-negative bacteria such as *E. Coli*, 120 seconds. [25].

In this study, the efficacy of plasma treatment was determined by evaluating the number of bacteria after treatment. The changes in bacterial viability after plasma treatment were evaluated using live/dead staining and CLSM. The morphological changes in the bacterial biofilm after plasma treatment were evaluated by SEM. In the cell assay, bacteria were not detected after direct plasma exposure for more than 3 minutes. The CLSM results were in accordance with those of the cell assay, and either the bacteria were undetected or most of the residual bacteria were dead in the 3- and 5-minute plasma groups. Other studies have shown a similar bactericidal effectiveness of plasma treatment by live/dead staining and CLSM and reported that plasma could diffuse on surfaces and even penetrate surface irregularities and porous structures, thereby efficiently disinfecting the implants [33,34]. In contrast, the unexposed area of the 3-minute plasma group showed yellow fluorescence, rather than green or red fluorescence, in the CLSM observations (red fluorescence indicates dying or dead bacteria). The expression of yellow fluorescence might have meant that propidium iodide was in the middle of the penetrated and ruptured cells; therefore, the mixture of green and red colors could have appeared yellow. One of the limitations of CLSM is that staining of bacterial cells does not always produce distinct red and green fluorescence, leading to difficulty in the interpretation of results [35].

The SEM images clearly show that bacteria were not found in the plasma-exposed area of the SLA titanium discs in this study. Intriguingly, most bacteria were removed, except a few ruptured bacteria in the adjacent, unexposed area. The effect of plasma seemed to gradually decrease as the distance from the plasma area increased. This phenomenon is in contrast with the localization of plasma in a previous low-resolution study, in which high-resolution observations showed that the influence of plasma on adjacent, unexposed areas was prominent [31]. A possible mechanism is that bacterial death in unexposed areas was caused by the displacement of the covering medium, resulting in the desiccation of bacteria. Those desiccated bacteria were observed by SEM and were called “dead frozen bacteria”; similar observations were made in our study [36]. Displacement of the covering medium is unavoidable if the plasma treatment time is over 3 minutes; therefore, desiccation may be one of the ways in which the bacteria were killed.

We found it interesting that the bacterial counts of the He gas group were considerably lower than those of the untreated control group in this study. Simple gas blowing without plasma generation seemed to attenuate bacterial biofilms to some degree. Plasma radicals can hardly penetrate a thick biofilm; thus, removal of the superficial layer of biofilm can be clinically attractive [34,37]. In addition, He gas blowing is thought to spread charged plasma particles over a more extensive area, thereby widening the plasma-affected region and resulting in a shift in the sterilization or disinfection area [38]. In the SEM investigation, the He group also showed disruption of the

superficial biofilm and a lower bacterial count. However, the remaining bacteria exhibited intact morphology, and most of them expressed green fluorescence in the CLSM images, indicating a viable state. These viable residual bacteria can recover rapidly and cause the recurrence of peri-implantitis. Thus, He gas blowing without plasma can temporarily reduce the bacterial count, but the effect might not be long-lasting due to rapid bacterial regrowth.

This study has several limitations that need to be addressed in further studies. *P. gingivalis* exists as a complex biofilm in the peri-implant pocket, not as a monoculture or as planktonic cells. *P. gingivalis* shows different phenotypes in biofilms of *in vitro* cultures [32]. Therefore, the response of *P. gingivalis* to the plasma treatment may not be the same in the mixed-species subgingival environment. Further research is needed to determine the response of *P. gingivalis* in an *in vivo* environment. Plasma treatment time was 3 or 5 minutes in this study. For clinical applications, it is believed that further investigation with more sophisticated time settings and evaluations are needed. Several mechanisms of action have been suggested previously, and further investigation is necessary to determine which mechanism contributes more to our plasma system. Additionally, the safety of plasma application in the mouth in contact with living cells should be evaluated in future studies.

V. CONCLUSIONS

In conclusion, a He-APPJ was effective for the decontamination of *P. gingivalis* biofilms on SLA surfaces without surface alterations. Additionally, the He-APPJ resulted in removal of bacteria from both direct plasma-exposed and unexposed sites. Based on these results, further research should investigate the clinical applicability of He-AAPJ as a decontamination technique for peri-implantitis treatment.

VI. REFERENCES

1. Mombelli A. Microbiology and antimicrobial therapy of peri-implantitis. *Periodontol 2000* 2002;28:177–89.
2. Schwarz F, Sculean A, Romanos G, Hertzen M, Horn N, Scherbaum W, et al. Influence of different treatment approaches on the removal of early plaque biofilms and the viability of SAOS2 osteoblasts grown on titanium implants. *Clin Oral Investig* 2005;9:111–7.
3. Persson LG, Mouhyi J, Berglundh T, Sennerby L, Lindhe J. Carbon dioxide laser and hydrogen peroxide conditioning in the treatment of periimplantitis: an experimental study in the dog. *Clin Implant Dent Relat Res* 2004;6:230–8.
4. Schwarz F, Nuesry E, Bieling K, Hertzen M, Becker J. Influence of an erbium, chromium-doped yttrium, scandium, gallium, and garnet (Er, Cr :YSGG) laser on the reestablishment of the biocompatibility of contaminated titanium implant surfaces. *J Periodontol* 2006;77:1820–7.
5. Schou S, Holmstrup P, Jørgensen T, Skovgaard LT, Stoltze K, Hjørting-Hansen E, et al. Implant surface preparation in the surgical treatment of experimental peri-implantitis with autogenous bone graft and ePTFE membrane in cynomolgus monkeys. *Clin Oral Implants Res* 2003;14:412–22.
6. Schwarz F, Ferrari D, Popovski K, Hartig B, Becker J. Influence of different air-abrasive powders on cell viability at biologically

contaminated titanium dental implants surfaces. J Biomed Mater Res B Appl Biomater 2009;88:83–91.

7. Kreisler M, Kohnen W, Christoffers AB, Götz H, Jansen B, Duschner H, et al. *In vitro* evaluation of the biocompatibility of contaminated implant surfaces treated with an Er : YAG laser and an air powder system. Clin Oral Implants Res 2005;16:36–43.

8. Mendis DA, Rosenberg M, Azam F. A note on the possible electrostatic disruption of bacteria. IEEE Trans Plasma Sci 2000;28:1304–6.

9. Laroussi M. Nonthermal decontamination of biological media by atmospheric-pressure plasmas: review, analysis, and prospects. IEEE Trans Plasma Sci 2002;30:1409–15.

10. Kim JH, Lee MA, Han GJ, Cho BH. Plasma in dentistry: a review of basic concepts and applications in dentistry. Acta Odontol Scand 2014;72:1–12.

11. Lackmann JW, Bandow JE. Inactivation of microbes and macromolecules by atmospheric-pressure plasma jets. Appl Microbiol Biotechnol 2014;98:6205–13.

12. Jablonowski L, Koban I, Berg MH, Kindel E, Duske K, Schröder K, et al. Elimination of *E. Faecalis* by a new non-thermal atmospheric pressure plasma handheld device for endodontic treatment. A preliminary investigation. Plasma Process Polym 2013;10:499–505.

13. Yang B, Chen J, Yu Q, Li H, Lin M, Mustapha A, et al. Oral bacterial deactivation using a low-temperature atmospheric argon plasma brush. *J Dent* 2011;39:48–56.
14. Cho BH, Han GJ, Oh KH, Chung SN, Chun BH. The effect of plasma polymer coating using atmospheric-pressure glow discharge on the shear bond strength of composite resin to ceramic. *J Mater Sci* 2010;46:2755–63.
15. Nam SH, Lee HJ, Hong JW, Kim GC. Efficacy of nonthermal atmospheric pressure plasma for tooth bleaching. *Sci World J* 2015;2015:581731.
16. Canullo L, Peñarrocha D, Clementini M, Iannello G, Micarelli C. Impact of plasma of argon cleaning treatment on implant abutments in patients with a history of periodontal disease and thin biotype: radiographic results at 24-month follow-up of a RCT. *Clin Oral Implants Res* 2015;26:8–14.
17. Koban I, Holtfreter B, Hübner NO, Matthes R, Sietmann R, Kindel E, et al. Antimicrobial efficacy of non-thermal plasma in comparison to chlorhexidine against dental biofilms on titanium discs *in vitro* – proof of principle experiment. *J Clin Periodontol* 2011;38:956–65.
18. Carreiro AF, Delben JA, Guedes S, Silveira EJ, Janal MN, Vergani CE, et al. Low-temperature plasma on peri-implant-related biofilm and gingival tissue. *J Periodontol* 2019;90:507–15.
19. Yoon SY, Kim KH, Seol YJ, Kim SJ, Bae B, Huh SR, et al. Effects of metastable species in helium and argon atmospheric

pressure plasma jets (APPJs) on inactivation of periodontopathogenic bacteria. *J Korean Phys Soc* 2016;68:1176–91.

20. Berglundh T, Gotfredsen K, Zitzmann NU, Lang NP, Lindhe J. Spontaneous progression of ligature induced peri-implantitis at implants with different surface roughness: an experimental study in dogs. *Clin Oral Implants Res* 2007;18:655–61.

21. Ehlbeck J, Schnabel U, Polak M, Winter J, von Woedtke T, Brandenburg R, et al. Low temperature atmospheric pressure plasma sources for microbial decontamination. *J Phys D Appl Phys* 2010;44:013002.

22. Laroussi M, Mendis DA, Rosenberg M. Plasma interaction with microbes. *New J Phys* 2003;5:41.1–41.10.

23. Venezia RA, Orrico M, Houston E, Yin SM, Naumova YY. Lethal activity of nonthermal plasma sterilization against microorganisms. *Infect Control Hosp Epidemiol* 2008;29:430–6.

24. Laroussi M, Sayler GS, Glascock BB, McCurdy B, Pearce ME, Bright NG, et al. Images of biological samples undergoing sterilization by a glow discharge at atmospheric pressure. *IEEE Trans Plasma Sci* 1999;27:34–5.

25. Lee K, Paek KH, Ju WT, Lee Y. Sterilization of bacteria, yeast, and bacterial endospores by atmospheric-pressure cold plasma using helium and oxygen. *J Microbiol* 2006;44:269–75.

26. Montie TC, Kelly-Wintenberg K, Roth JR. An overview of research using the one atmosphere uniform glow discharge plasma

(OAUGDP) for sterilization of surfaces and materials. IEEE Trans Plasma Sci 2000;28:41-50.

27. Kim SJ, Yoon SY, Kim GH. Bullet velocity distribution of a helium atmospheric-pressure plasma jet in various N₂/O₂ mixed ambient conditions. IEEE Trans Plasma Sci 2015;43:2054-63.

28. Miura T, Egawa M, Ito T, Eguro T, Tanabe K, Yoshinari M. Debridement effect on periodontal pathogen *Porphyromonas gingivalis* cultured on titanium by application of atmospheric-pressure plasma. J Biomed Sci Eng 2017;10:51-9.

29. Shi Q, Song K, Zhou X, Xiong Z, Du T, Lu X, Cao Y. Effects of non-equilibrium plasma in the treatment of ligature-induced peri-implantitis. J Clin Periodontol 2015; 42: 478 - 487. doi: 10.1111/jcpe.12403.

30. Lamont RJ, Jenkinson HF. Life below the gum line: pathogenic mechanisms of *Porphyromonas gingivalis*. Microbiol Mol Biol Rev 1998;62:1244-63.

31. Mahasneh A, Darby M, Tolle SL, Hynes W, Laroussi M, Karakas E. Inactivation of *Porphyromonas gingivalis* by low-temperature atmospheric pressure plasma. Plasma Med 2011;1:191-204.

32. Marsh PD. Dental plaque as a microbial biofilm. Caries Res 2004;38:204-11.

33. Preissner S, Wirtz HC, Tietz AK, Abu-Sirhan S, Herbst SR, Hartwig S, et al. Bactericidal efficacy of tissue tolerable plasma on

microrough titanium dental implants: An *in-vitro*-study. J Biophotonics 2016;9:637-44.

34. Sladek RE, Filoche SK, Sissons CH, Stoffels E. Treatment of *Streptococcus mutans* biofilms with a nonthermal atmospheric plasma. Lett Appl Microbiol 2007;45:318-23.

35. Berney M, Hammes F, Bosshard F, Weilenmann HU, Egli T. Assessment and interpretation of bacterial viability by using the LIVE/DEAD BacLight Kit in combination with flow cytometry. Appl Environ Microbiol 2007;73:3283-90.

36. Shashurin A, Keidar M, Bronnikov S, Jurjus RA, Stepp MA. Living tissue under treatment of cold plasma atmospheric jet. Appl Phys Lett 2008;93:181501.

37. Vleugels M, Shama G, Deng XT, Greenacre E, Brocklehurst T, Kong MG. Atmospheric plasma inactivation of biofilm-forming bacteria for food safety control. IEEE Trans Plasma Sci 2005;33:824-8.

38. Fridman G, Brooks AD, Balasubramanian M, Fridman A, Gutsol A, Vasilets VN, et al. Comparison of direct and indirect effects of non-thermal atmospheric-pressure plasma on bacteria. Plasma Process Polym 2007;4:370-5.

Table and Figures

Table 1. Number of *Porphyromonas gingivalis* colonies (CFU/mL) in the biofilm according to the He-APPJ treatment time

	SLA surface		<i>P</i> value	
	He gas	<i>P</i> value	He-APPJ	<i>P</i> value
Untreated control group	14,392 ± 5,215		13,520 ± 3,790	0.829 ^{b)}
3-min treatment group	1,476 ± 269 *	<0.001 ^{a)}	754 ± 1,299 *	<0.001 ^{a)} 0.426 ^{b)}
5-min treatment group	1,773 ± 946 *	<0.001 ^{a)}	0 *	<0.001 ^{a)} 0.033 ^{b)} *

Data are shown as mean ± standard deviation.

CFU: colony-forming unit, SLA: sandblasted and acid-etched, He: helium,

He-APPJ: helium atmospheric-pressure plasma jet.

^{a)}Statistical significance compared to the control group (1-way analysis of variance);

^{b)}Statistical significance between the He group and the He-APPJ group (Student *t*-test).

* *P* value < 0.05 was considered to be statistically significant.




Figure 1 A. Scanning electron microscopy images of machined titanium discs after He-APPJ treatment for 3 (b) or 5 (c) minutes to evaluate the surface change of titanium discs after He-APPJ treatment (scale bar = 100 μm). Untreated disc was control (a, No treatment). Surface changes were not found in comparison to the untreated control. (original magnification $\times 500$) **B.** Three-dimensional confocal laser scanning microscopy showing the roughness (Ra) of the SLA titanium discs treated with He-APPJ for 3 (b) or 5 (c) minutes. Untreated disc was control (a, No treatment). Marked changes of the roughness depending on the treatment time was difficult to detect.

He-APPJ : helium atmospheric-pressure plasma jet

SLA : sandblasted and acid-etched




Figure 2 A. Scanning electron microscopy images of the growth of *P. gingivalis* after pre-treatment of sterilized discs with He-APPJ for 3 (b) or 5 (c) minutes followed by cultivating *P. gingivalis* (scale bar = 5 μm). Untreated disc was control (a, No treatment). The results showed that He-APPJ pre-treatment did not affect the growth of *P. gingivalis*. (original magnification $\times 10,000$) **B.** Confocal laser scanning microscopy images of the osteoblast cultivated for 1 day after He-APPJ treatment for 3 (b) or 5 (c) minutes. Untreated discs was control (a, No treatment). The results showed that cells attached well regardless of exposure time and inferred that He-APPJ treatment did not impede the adhesion and growth of osteoblasts.

He-APPJ : helium atmospheric-pressure plasma jet




Figure 3. Photograph of the apparatus for He-APPJ generation and He-APPJ application to titanium discs in this study.

He-APPJ: helium atmospheric-pressure plasma jet.




Figure 4. Scanning electron microscopy images of the central and marginal

areas after He gas or He-APPJ treatment (scale bar = 5 μm). The control group (no treatment) showed heavy bacterial accumulation on the discs. In the He gas - only treatment group, remaining bacterial biofilm was observed in both the central and marginal areas in the 3-minute treatment group, while bacterial rods had been eliminated from the central area of the 5-minute treatment group, but not in the marginal area. However, the He-APPJ group showed a clear titanium surface without a bacterial biofilm in both the 3-minute and 5-minute treatment groups (original magnification $\times 10,000$).

He-APPJ: helium atmospheric-pressure plasma jet.




Figure 5. Confocal laser scanning microscopy images of the central and marginal areas after He gas or He-APPJ application (scale bar = 10 μm). Heavy live bacterial accumulation on the discs was observed in the control group. In the He gas -only treatment group, the amount of bacteria was higher in the central area than in the marginal area in the 3-minute treatment group. In the 5-minute treatment group, the amount of bacteria decreased in both the central and marginal areas, although more live bacteria were observed in the marginal area. In the He-APPJ treatment group, the amount of bacteria was lower in the central area than in the marginal area in both the 3-minute and 5-minute treatment groups. Only a few viable

bacteria were observed in the central area of the 3-minute treatment group. In the marginal area, a mixture of live and dead bacteria was found, whereas most bacteria were dead in the 5-minute treatment group (original magnification $\times 63$).

He: helium, He-APPJ: helium atmospheric-pressure plasma jet.

국문 초록

Sandblasting 후 산부식 처리한
티타늄 디스크 상의
Porphyromonas gingivalis
바이오필름에 대한
Atmospheric-Pressure Plasma
Jet의 살균 효과

이 지 윤

서울대학교 대학원

치의과학과 치의과학 전공

(지도교수 설 양 조)

Atmospheric-pressure plasma jet (APPJ)를 세균에 직접 처리하는 것이 세균의 멸균에 효과적인 방법이라는 것은 이미 많은 연구를 통해 알려진 사실이다. 이번 연구는 Sandblasting 후 산부식 처리한 (sandblasted and acid-etched; SLA) 티타늄 디스크 위에 형성된 *Porphyromonas gingivalis* (*P. gingivalis*) 바이오필름에 헬륨 가스를 이용한 APPJ (He-APPJ)를 직접 적용하여 그 살균 효과를 연구하고자 하였다.

P. gingivalis 바이오필름으로 덮인 SLA 티타늄 디스크 표면에 He-APPJ를 3, 5분간 적용한 군, 플라즈마 자체의 효과를 평가하기 위해 플라즈마 없이 헬륨 가스만 동일 시간 적용한 군을 실험군으로, 아무 처리도 하지 않은 군을 대조군으로 하였다. He-APPJ의 살균 효과는 집락 형성 단위 (colony forming units)의 수로 판단하였다. 세균 생존력 (viability)은 공초점 레이저 주사 현미경 (confocal laser scanning microscopy :CLSM)으로, 세균의 형태 변화는 주사 전자 현미경 (scanning electron microscopy :SEM)으로 관찰하였다.

실험 결과 플라즈마 적용 시간이 길어질수록 *P. gingivalis*의 수는 감소하였고, 그 차이는 통계적으로 유의하였다. SEM 영상을 통해서 관찰한 결과 He-APPJ로 3분 이상 치료했을 경우 SLA 티타늄 디스크의 세균 바이오필름 구조가 대조군과 비교했을 때 유의미하게 파괴된 것으로 관찰되었고, CLSM 영상도 집락 형성 단위의 수, SEM과 일치된 결과를 보여주었다. He-APPJ가 직접적으로 노출된 부위로부터 떨어진 영역에서도 SEM 영상과 CLSM 영상 모두에서 살균 효과가 관찰되었다.

결론적으로, 이번 연구에서 시험관 (in vitro) 실험 하에서 He-APPJ가 SLA 티타늄 디스크 위에 형성된 *P. gingivalis* 바이오필름을 제거하는데 효과적인 방법임을 입증하였다.

Keywords: 세균 부하; 공초점 현미경; 플라즈마 기체; *Porphyromonas gingivalis*; 주사 전자 현미경; 티타늄

Student Number: 2010-30638