

Current status of dental caries diagnosis using cone beam computed tomography

Young-Seok Park, Jin-Soo Ahn*, Ho-Beom Kwon**, Seung-Pyo Lee

Department of Oral Anatomy and Dental Research Institute, School of Dentistry, Seoul National University, Seoul, Korea

*Department of Biomaterials Science and Dental Research Institute, School of Dentistry, Seoul National University, Seoul, Korea

**Department of Prosthodontics and Dental Research Institute, School of Dentistry, Seoul National University, Seoul, Korea

ABSTRACT

Purpose : The purpose of this article is to review the current status of dental caries diagnosis using cone beam computed tomography (CBCT).

Materials and Methods : An online PubMed search was performed to identify studies on caries research using CBCT.

Results : Despite its usefulness, there were inherent limitations in the detection of caries lesions through conventional radiograph mainly due to the two-dimensional (2D) representation of caries lesions. Several efforts were made to investigate the three-dimensional (3D) image of lesion, only to gain little popularity. Recently, CBCT was introduced and has been used for diagnosis of caries in several reports. Some of them maintained the superiority of CBCT systems, however it is still under controversies.

Conclusion : The CBCT systems are promising, however they should not be considered as a primary choice of caries diagnosis in everyday practice yet. Further studies under more standardized condition should be performed in the near future. (*Imaging Sci Dent 2011; 41 : 43-51*)

KEY WORDS : Cone-Beam Computed Tomography; Dental Caries; Diagnosis

Introduction

As a chronic infectious disease, dental caries is one of the most common problems encountered in clinical dentistry,¹ and shows a very high incidence of over 80% of all adults in Korea according to the 2006 national survey.² Early and accurate diagnosis of caries is essential for clinicians, who require exact knowledge of the depth of caries in order to determine the appropriate type of restoration and treatment planning.^{3,4} Among the various types of methods in the diagnosis of caries, probing, visual examination, intraoral film, and digital sensors are commonly used in routine clinical practice.⁵ Such diagnostic methods in the management of dental caries are used to determine the pre-

sence of caries and its extent, to monitor the course of caries progression, and to evaluate the effectiveness of treatment.⁶ However, a general observation from a review was that current methods tended to be more specific than sensitive (that is, their use results in relatively more false negative findings than false positive findings).⁷ Actually, several studies have shown that between 25% and 42% of caries lesions remain undetected by clinical examination performed without radiographic examination.⁸⁻¹⁰

When it comes to visual examinations with the adoption of the International Caries Detection and Assessment System (ICDAS II) clinical criteria in 2005, it has been attempted to correlate the clinical image of teeth to their histological status.¹¹ However, the impact of these criteria on the diagnostic performance of visual examination is being examined on a limited number of studies, while the differences in caries location, lesion extent and experimental setup make extrapolation conduction difficult.¹²⁻¹⁵

In everyday clinical situation, if we exclude the recently introduced sophisticated methods for diagnosis of carious

*This research was supported by Basic Science Research Program through the National Research Foundation of Korea (NRF) funded by the Ministry of Education, Science, and Technology (2010-0023586).

Received March 16, 2011; Revised April 19, 2011; Accepted April 29, 2011

Correspondence to : Prof. Seung-Pyo Lee

Department of Oral Anatomy, School of Dentistry, Seoul National University, 28 Yeongeong-dong, Jongno-gu, Seoul 110-749, Korea

Tel) 82-2-740-8671, Fax) 82-2-740-8674, E-mail) orana9@snu.ac.kr

lesion such as fiberoptic transillumination (FOTI), electrical conductance (EC), laser fluorescence, and so on, the radiographic examination is the most frequently recommended method as a supplement to the clinical inspections.^{16,17} Regarding proximal non-cavitated caries, in spite of the variety of diagnostic modalities available, radiography is the most widely used method.¹¹ Bitewing radiography has been available for more than 80 years,^{12,18} while in the more recent years digital radiographic modalities attempted to substitute conventional film radiography. The advantages of a digital system are the abilities to manipulate the image contrast and brightness and to magnify the images. However, extensive researches regarding conventional and various digital radiographic modalities failed to detect significant differences in their diagnostic performances, only to show comparable results.¹⁹⁻²⁸

Despite its usefulness, the detection of caries lesions through conventional radiograph remains rather an elusive process.²⁹ The limitations inherent in conventional radiography are mainly due to the 2D representation of caries lesions, which are 3D structures in reality, and this might lead to loss of valuable information.^{30,31} Small lesions remain undetected when the relative amount of mineral loss is low, resulting in low subject and image contrast. Moreover, the radiographic appearance of a lesion can change dramatically as a function of the chosen projection geometry. The replacement of film by digital detectors does not address these fundamental limitations.^{21,28,32,33}

Dentistry has largely used the same method of 2D imaging since the first intraoral radiograph obtained in 1896. According to the review by Tyndall et al,³⁴ only one or two significant advances in dental imaging have been made since then in the sense of imaging geometry. These advances include panoramic imaging and tomography, with the former being far more useful for dental applications, and the latter historically being limited primarily to temporomandibular joint and implant site imaging.³⁴

Computed tomography (CT), which was invented by Hounsfield³⁵ in 1973, is considered as a technical breakthrough. It is a well-known medical technique for the nondestructive examination of internal structures and its introduction to the dentistry has been revolutionary in the sense that it can provide true 3D imaging.²⁹ Cone beam computed tomography (CBCT) is a new application of CT that generates 3D data at lower cost and absorbed doses than conventional “fan beam” CT found in the practice of medical field.³⁶ Data from the craniofacial region are often collected at higher resolution in the axial plane than those from conventional CT systems.³⁷ In addition,

these systems do require relatively small amount of space and can easily fit into most dental clinics today. Although most of the CBCT usage has been confined to the applications for dental implant placement, orthodontics, surgery, and temporomandibular joint disease³⁸⁻⁴⁴ so far, several studies focusing on the diagnosis of dental caries have been reported.³⁴

The aim of this article is to review the brief history of the usage of 3D concept in diagnosis of the dental caries and the current status of diagnosis using CBCT. This review builds on the findings of several recent articles related to caries diagnosis using CBCT, and seeks to outline the possible advanced applications that might be used in the future.

Materials and Methods

A PubMed search from 1965 up to February of 2011 was conducted to identify articles published in dental literature, and limited to human trials, using the search terms “caries”, “diagnosis”, “3 dimensional”, “tomography”, and “computed tomography”. Manual searches of the bibliographies of all full text articles and related reviews selected from the electronic search were also performed.

Results

Early trials to realize the 3D image

An ideal diagnostic tool would enable the clinician to accurately assess the presence or absence of a lesion, to quantify its size and depth, and to determine its activity. Whereas the physics underlying the radiographic image formation process is well suited for imaging the dental structures, the sampling level of traditional intraoral imaging is not sufficient to fulfill these requirements. In order to acquire 3D information, the level of sampling needs to be increased.²⁹

Earlier attempts have been made to improve the diagnosis of dentoalveolar conditions with 3D imaging using variations of tomosynthesis.³⁴ This has been the underlying premise of tuned aperture computed tomography (TACT), which uses a limited number of basis projections to generate 2D slices at various depths.⁴⁵ Although TACT provided some incremental benefit for periodontal and endodontic applications, improvements in caries detection and characterization were limited to simulated recurrent caries. In case of proximal lesions, a significant increase in detection rate could not be demonstrated according to several

reports.⁴⁶⁻⁵²

Unfortunately, TACT has provided limited application in the practice of dentistry thus far, partly due to the advent of CBCT, ironically. The development of CBCT has been innovative because complete (360°) sampling is now possible without increasing the patient dose to unacceptable levels.²⁹

A benchtop-based CT device using an intraoral detector as the image receptor, which was developed by van Daatselaar and coworkers,⁵³ was another attempt used in caries diagnosis. It is referred as a local computed tomography (LCT), and has similar basic working principles with the commercial CBCT systems on the market, although it is not automated. They used a high resolution charge-coupled device (CCD) detector and rotating turntable with a fixed anode intraoral radiograph source. Both TACT and LCT generate a series of images that can be reconstructed into a series of cross-sectional images. Whereas TACT uses a fixed object and CCD sensor⁴⁵ while moving the X-ray source, LCT uses a fixed CCD sensor and source and a rotating object.⁵⁴

In their studies,⁵³⁻⁵⁶ they have shown the feasibility of LCT and the improved accuracy in caries detection compared with conventional radiography. The term, LCT is somewhat confusing with the local cone-beam computed tomography (LCBCT), which is frequently used in contrast to full volume CBCT. Recently, Kalathingal et al²⁹ used LCT in their research and found no difference in the detection of carious lesions, however they did find that LCT was superior for assessment of caries depth.

Researches using clinical CBCT systems

Astounding speed of development and improvement of dental imaging technology makes the proper verification difficult. When it comes to application of clinical CBCT systems in dental caries research, the study of Akdeniz et al⁵⁷ was the first English literature available via Pubmed search. They found that Accuitomo 3DX (Morita Co. Ltd, Tokyo, Japan) CBCT system was superior in caries depth assessment compared with conventional film radiography or storage phosphor (SP) images. There were many studies^{5,6,25,57-61} dealing with the caries researches using clinical CBCT system including the study by Akdeniz et al.⁵⁷ Caries researches have been traditionally classified according to the location of caries lesion. Their studies dealt with only proximal lesions,^{5,6,57,59} occlusal lesions exclusively,⁶⁰ and both proximal and occlusal lesions.^{25,58,61} Various CBCT systems with different settings have been used, al-

though Accuitomo 3DX was the most frequently used one. They were compared with the conventional imaging modalities which included film radiography and digital intraoral radiography using CCD and photostimulable phosphor (PSP) plates. All the studies were in vitro studies, and they used premolar and molar teeth with various stages of caries lesions. Intraoral radiographs were taken using paralleling technique and rectangular collimation. For comparison of film radiography, the film was developed automatically. For the rating of the caries lesions, five-scale score was used when using histology examinations.

However, aforementioned studies have not been standardized in every aspect. For conventional imaging modalities, there were differences not only in systems but also in their X-ray taking environment such as voltage, current, exposure time, and focus to film (or receptor) distance. Regarding these differences, the authors might follow the manufacturer's instructions or clinical experiences. Storage method of teeth, the width of acrylic block mimicking the soft tissue, image software and comparison analysis were different among the studies. In addition, the status of caries lesion was diverse according to the studies. Several studies used the image of X-ray microcomputed tomography (MicroCT) as a gold standard instead of conventional histology. The number of observers was also different from study to study. Detailed information about imaging modalities and environments are described in Tables 1-3.

Table 1. Product information. The manufacturer's name follows the original articles' description

Kodak Insight (Eastman Kodak company, Rochester, NY, USA)
Kodak Ectaspeed (Eastman Kodak company, Rochester, NY, USA)
Diagora-fmx (Soredex, Tuusula, Finland)
Diagora Optime (Soredex/Orion Corp., Helsinki, Finland)
#2 CCD (E2V Technologies Inc., Elmsford, NY, USA)
Progeny Vision DX (Progeny Dental, Buffalo Grove, IL, USA)
MAX CCD (Benlioğlu Dental, Ankara, Turkey)
3DX Accuitomo (Morita Co. Ltd., Tokyo, Japan)
NewTom 3G (Quantitative Radiology, Verona, Italy)
NewTom 9000 (Quantitative Radiology, Verona, Italy)
Kodak 9000 3D (Carestream Health, Inc., Rochester, NY, USA)
Kodak 9500 3D (Carestream Health, Inc., Rochester, NY, USA)
Promax 3D (Planmeca Oy, Helsinki, Finland)
DCT PRO (VATECH, Co., Ltd., Yongin-Si, Korea)
ILUMA (Imtec Imaging, Ardmore, OK, USA)
Gendex 1000 X-ray unit (General Electric Co., Milwaukee, WI, USA)
Gendex Oralix DC X-ray unit (Gendex Dental Systems, Milan, Italy)
HD-70 X-ray unit (Asahi Roentgen Ind. Co., Kyoto, Japan)
Trophy Trex X-ray unit (Croissy, Beaubourg, France)
Trophy ETX X-ray unit (Trophy Radiologie, France)
AET-Orix 70 X-ray unit (Ardet, Buccinasco, Italy)

Table 2. Detailed information about imaging modalities used in studies regarding detection of lesions in approximal surfaces

Research	Gold standard	Image modalities	Additional information
Kaypimaz et al ^{61,*}	Histology	Film (Kodak Ektaspeed) PSP (Diagora Optime) Kodak 9500 FOV (9 × 15 cm)	Trophy ETX X-ray unit, 65 kVp, 8 mA, FFD [†] 30 cm Voxel size 0.2 mm, Scan time 10.8 s
Şenel et al ⁵	Histology	Kodak E speed CCD (Progeny Vision DX) PSP (Diagora Optime) ILUMA FOV (21.1 × 14.2 cm)	Trophy Trex X-ray unit, 65 kVp, 8 mA, FFD [†] 20 cm 120 kVp, 3.8 mA 0.3 mm voxel, Scan time 40 s
Qu et al ⁵⁹	Histology	NewTom9000 FOV (15 × 15 cm) 3DX Accuitomo FOV (3 × 4 cm) Kodak 9000 3D FOV (5 × 3.7 cm) Promax 3D FOV (8 × 8 cm) DCT Pro FOV (20 × 19 cm)	110 kVp, 2.1 mA, Scan time 36 s, Voxel size 0.3 mm, Slice thickness 0.4 mm 80 kVp, 5 mA, Scan time 17.5 s, Voxel size 0.125 mm, Slice thickness 1 mm 70 kVp, 10 mA, Scan time 24 s, Voxel size 0.076 mm, Slice thickness 0.076 mm 76 kVp, 6 mA, Scan time 18 s, Voxel size 18 mm, Slice thickness 0.32 mm 90 kVp, 3.5 mA, Scan time 24 s, Voxel size 0.322 mm, Slice thickness 1 mm
Young et al ^{58,*}	MicroCT	CCD (#2 CCD) 3DX accuitomo FOV (4 cm)	Gendex 1000 X-ray unit, 70 kVp, 10 mA 80 kVp, 5 mA
Haiter-Neto et al ^{25,*}	Histology	PSP (Diagora-fmx) Film (Kodak Insight) NewTom FOV (12 inch) NewTom FOV (9 inch) NewTom FOV (6 inch) Accuitomo FOV (4 cm)	Gendex DC X-ray unit, 65 kVp, 10 mA, FFD [†] 32 cm Voxel size 0.36 mm Voxel size 0.25 mm Voxel size 0.16 mm Automated adjusted milliamperes, Scan time 36 s Voxel size 0.125 mm, 60 kVp, Scan time 18 s
Tsuchida et al ⁶	MicroCT	Film (Kodak Insight) 3DX Accuitomo FOV 4 cm	HD-70 X-ray unit (60 kVp, 7 mA FFD [†] 40 cm) 80 kVp, 4 mA, Scan time 18 s
Akdeniz et al ⁵⁷	Histology	Film (Kodak F-speed) Diagora fmx (Soredex) 3DX Accuitomo FOV 4 cm	Gendex Oralix DC X-ray unit (60 kVp, 7 mA FFD [†] 25 cm) 80 kVp, 1.5 mA, Scan time 17.5 s

* denotes the study dealt with both approximal and occlusal caries lesions.

† denotes the focus to film distance or focus to receptor distance.

Table 3. Detailed information about imaging modalities used in studies regarding detection of lesions in occlusal surfaces

Research	Gold standard	Image modalities	Additional information
Kaypimaz et al ^{61,*}	Histology	Film (Kodak Ektaspeed) PSP (Diagora Optime) Kodak 9500 FOV (9 × 15 cm)	Trophy ETX X-ray unit, 65 kVp, 8 mA, FFD [†] 30 cm Voxel size 0.2 mm, Scan time 10.8 s
Kamburoğlu et al ⁶⁰	Histology	Film (Kodak E speed) ILUMA FOV (21.1 × 14.2 cm)	AET-Orix 70 X-ray unit, 70 kVp, 8 mA, FFD [†] 20 cm Voxel size 0.1 mm 120 kVp, 3.8 mA 0.3 mm voxel, Scan time 40 s Voxel size 0.3 mm
Young et al ^{58,*}	MicroCT	CCD (#2 CCD) 3DX accuitomo FOV (4 cm)	Gendex 1000 X-ray unit (70 kVp, 10 mA) 80 kVp, 5 mA
Haiter-Neto et al ^{25,*}	Histology	PSP (Diagora-fmx) Film (Kodak Insight) NewTom FOV (12 inch) NewTom FOV (9 inch) NewTom FOV (6 inch) Accuitomo FOV (4 cm)	Gendex DC X-ray unit, 65 kVp, 10 mA, FFD [†] 32 cm Voxel size 0.36 mm Voxel size 0.25 mm Voxel size 0.16 mm Automated adjusted milliamperes, Scan time 36 s Voxel size 0.125 mm, 60 kVp, Scan time 18 s

* denotes the study dealt with both approximal and occlusal caries lesions.

† denotes the focus to film distance or focus to receptor distance.

Methods of comparing analysis

Several studies^{5,6,59,61} used Az value for comparison which means area under the receiver operating character-

istic (ROC) analysis. The ROC curve is a fundamental tool for diagnostic test evaluation. It allows to a complete sensitivity/specificity report. In a ROC curve the true positive rate (sensitivity) is plotted in function of the false positive

rate (1-specificity) for different cut-off points of a parameter.⁶² Each point on the ROC curve represents a sensitivity/specificity pair corresponding to a particular decision threshold. The area under the ROC curve, Az, is a measure of how well a parameter can distinguish between two diagnostic groups (diseased/normal).⁶³ Some aforementioned studies^{25,58} used sensitivity and specificity for comparing image modalities, however they did not present the ROC curve and Az values.

The study of Kamburoğlu K et al⁶⁰ introduced the relative treatment effect (RTE) value as a statistical parameter according to the lesion depth. A 'treatment effect' is the average causal effect of a binary (0-1) variable on an outcome variable of scientific or policy interest. The term 'treatment effect' originates in a medical literature concerned with the causal effects of binary, yes-or-no 'treatments', such as an experimental drug or a new surgical procedure.⁶⁴ However, the term is now used much more generally.

The study of Akdeniz et al⁵⁷ is somewhat peculiar in

that it compared the accuracy of determining the depth of proximal caries lesion, not just the accuracy of detecting the existence of lesion. Actually, this kind of studies, which deal with the defining the 3 dimensional region of the caries lesion, are now being actively performed by various research groups using X-ray microcomputed tomography instead of CBCT.⁶⁵

Summary of comparisons

Tsuchida et al⁶ used the noncavitated proximal incipient lesions and found that no significant differences between CBCT and film images. This result might reflect difficulty of detecting incipient lesion.

Haiter-Neto et al²⁵ compared NewTom 3G system with 3 fields of view as a full-volume CBCT with Accuitomo 3DX as a local CBCT (LCBCT). The intraoral radiography was also compared. The results showed that the NewTom 12-inches and 9-inches images had significantly lower sensitivities than the Accuitomo systems, whereas the NewTom 9-inches and 6-inches images had significantly lower

Table 4. Brief summary of results from studies regarding detection of lesion in approximal surfaces

Research	N	Observer	Image modalities	Az	Sensitivity	Specificity	Lesion depth
Kaypimaz et al ^{61,*}	72	2	Film (Kodak Ektaspeed)	0.782±0.0561			
			PSP (Diagora Optime)	0.689±0.0637			
			Kodak 9500 FOV (9×15 cm)	0.705±0.0627			
Şenel et al ⁵	230	3	Kodak E speed	0.835			
			CCD (Progeny Vision DX)	0.861			
			PSP (Diagora Optime)	0.823			
			ILLUMA FOV (21.1×14.2 cm)	0.883			
Qu et al ⁵⁹	39	7	NewTom9000 FOV 15×15 cm	0.541±0.033			
			3DX Accuitomo FOV (3×4 cm)	0.555±0.044			
			Kodak 9000 3D FOV (5×3.7 cm)	0.577±0.038			
			Promax 3D FOV (8×8 cm)	0.545±0.024			
			DCT Pro FOV (20×19 cm)	0.549±0.028			
Young et al ^{58,*}	146	8	CCD (#2 CCD)		0.18±0.15 (E)	0.96±0.08 (E)	
					0.33±0.06 (D)	0.96±0.02 (D)	
			3DX accuitomo FOV (4 cm)		0.24±0.08 (E)	0.95±0.05 (E)	
					0.61±0.05 (D)	0.94±0.05 (D)	
Haiter-Neto et al ^{25,*}	100	6	PSP (Diagora-fmx)		0.17	0.91	
			Film (Kodak Insight)		0.18	0.92	
			NewTom FOV (12 inch)		0.13	0.88	
			NewTom FOV (9 inch)		0.14	0.85	
			NewTom FOV (6 inch)		0.18	0.84	
			Accuitomo FOV (4 cm)		0.21	0.89	
Tsuchida et al ⁶	50	7	Film (Kodak Insight)	0.633±0.029			
			3DX Accuitomo FOV 4 cm	0.625±0.018			
Akdeniz et al ⁵⁷	30	2	Film (Kodak F-speed)				0.44 mm (95%CI 1.0-1.8)
			Diagora fmx (Soredex)				0.7 mm (95%CI 1.2-2.5)
			3DX Accuitomo FOV 4 cm				0.5 mm (95%CI 1.4-2.5)

* denotes the study dealt with both approximal and occlusal caries lesions.

(E) means the lesion confined to the Enamel, (D) means the lesion penetrated into the dentin.

Table 5. Brief summary of results from studies regarding detection of lesion in occlusal surfaces

Research	N	Observer	Image modalities	Az	Sensitivity	Specificity	Relative treatment effect [†]				
							SE	DE	SD	DD	H
Kayipmaz et al ^{61,*}	72	2	Film (Kodak Ektaspeed)	0.666±0.0723							
			PSP (Diagora Optime)	0.649±0.0730							
			Kodak 9500 FOV (9×15 cm)	0.849±0.0554							
Kamburoğlu et al ⁶⁰	130	3	Max CCD (Benlioğlu dental)				0.26	0.34	0.61	0.67	0.82
			ILLUMA Voxel size 0.1 mm				0.28	0.44	0.71	0.76	0.82
			ILLUMA Voxel size 0.2 mm				0.29	0.52	0.76	0.79	0.85
			ILLUMA Voxel size 0.3 mm				0.31	0.62	0.84	0.85	0.84
Young et al ^{58,*}	146	8	CCD (#2 CCD)		0.18±0.20 (E)	0.88±0.18 (E)					
					0.64±0.10 (D)	0.90±0.07 (D)					
			3DX accuitomo FOV (4 cm)		0.64±0.15 (E)	0.61±0.18 (E)					
					0.93±0.06 (D)	0.63±0.16 (D)					
Haitei-Neto et al ^{25,*;§}	100	6	PSP (Diagora-fmx)		0.17	0.91					
			Film (Kodak Insight)		0.18	0.92					
			NewTom FOV (12 inch)		0.13	0.88					
			NewTom FOV (9 inch)		0.14	0.85					
			NewTom FOV (6 inch)		0.18	0.84					
			Accuitomo FOV (4 cm)		0.21	0.89					

* denotes the study dealt with both approximal and occlusal caries lesions.

§ The results of Haitei-Neto et al in this table is not from occlusal lesion exclusively, but from occlusal+approximal surface.

† In the subdivisions of Relative Treatment Effect, SE: superficial enamel, DE: deep enamel, SD: superficial dentin, DD: deep dentin, H: healthy

specificities than the conventional radiography in detecting proximal lesions. The Accuitomo images were reported to be comparable with the conventional radiography. For occlusal caries detection, this LCBCT system presented a higher sensitivity than the other systems. In addition, it was determined to be equal to the intraoral systems, however the overall true score (true positives and negatives) was not higher for detection of dentinal lesions.

Young et al⁵⁸ compared the efficacy of CBCT and conventional CCD image in detecting proximal and occlusal lesions. They used 3DX Accuitomo systems as a CBCT and found a significant difference in the average sensitivity score between CBCT and CCD regarding detection of proximal caries. They concluded that by using CBCT, it was able to improve the detection of proximal surface caries extending into the dentin, but not the occlusal caries.

Currently, there are two types of image detectors employed in CBCT equipments.⁵⁹ One is the image intensifier with CCD, and the other is the flat panel detector including amorphous silicon flat panel and complementary metal oxide semiconductor panel. Qu et al⁵⁹ evaluated the diagnostic accuracy of approximal lesions among the 5 clinical CBCT systems. The systems showed no statistical significant difference, and there was no significant difference between the types of the detectors of CBCT systems.

Kamburoğlu K et al⁶⁰ assessed the diagnostic ability of

intraoral digital CCD sensor images and CBCT images at different voxel resolutions in detection of occlusal caries. They found that the modalities of imaging performance were different in deep enamel, superficial dentin, and deep dentin. However, there was no difference in healthy and superficial enamel caries groups. They concluded that at all voxel sizes, CBCT images could be considered a tool for use in the diagnosis of occlusal caries.

The study by Şenel et al⁵ assessed the diagnostic ability of visual inspection, film, CCD sensor, PSP plate, and CBCT in the detection of proximal caries. The authors concluded that all the methods performed similarly in the detection of proximal caries.

Kayipmaz et al⁶¹ compared the effectiveness of conventional radiograph, PSP plate, and CBCT systems in determination of occlusal and approximal caries. In determining occlusal caries, CBCT was statistically superior to the other two conventional methods. However, no significant difference was verified in determining approximal caries.

Brief summary of results are described in Tables 4 and 5.

Discussion

Until now, only a small amount of research has been undertaken regarding caries diagnosis using CBCT. Although some studies maintained the superiority or the promise of using clinical CBCT system for diagnosis of den-

tal caries, we could not find the consensus of the research up to now. There was a tendency among the studies to insist that the accuracy of CBCT systems was higher than the conventional methods in detecting occlusal caries and deep lesions into the dentin, however evidences are still insufficient. Sensitivity may increase, but mostly with simultaneous decrease of specificity. As mentioned above, the conditions of the experiments were different from study to study and not standardized. Thus, it is appropriate to say that whether the CBCT is superior to the conventional modalities in diagnosis of dental caries is under controversy at this stage. Furthermore, it goes without saying that the routine use of CBCT system instead of conventional radiography should not be accepted.

When considering a comparison of different modalities, it should be reminded that an increase in efficacy or lack of thereof does not always mean superiority or inferiority. We should consider cost, time, and effort. Above all, we cannot but consider the radiation dose despite the improvement of image.

The imaging of CBCT has fewer problems of geometric distortion than those of conventional methods in theory, thus the real 3D representation of caries lesion was available, which was impossible before. However, there are surely several limitations in clinical situations. It should also be reminded that all the studies mentioned were in vitro studies. The images were obtained under ideal geometry with no closed contact, cone cut, soft tissues, and projection distortions. In addition, the presence of any metal restoration in clinical situations might affect the quality of CBCT image. Further studies are compulsory not only to elucidate the accuracy of current systems, but also to verify and keep face with the future systems including in vivo studies.

References

- Pereira AC, Verdonschot EH, Huysmans MC. Caries detection methods: can they aid decision making for invasive sealant treatment? *Caries Res* 2001; 35 : 83-9.
- Ministry of Health and Welfare. National survey of oral health in 2006. Seoul: Ministry of Health and Welfare; 2007. p. 59-70.
- Attrill DC, Ashley PF. Occlusal caries detection in primary teeth: a comparison of DIAGNOdent with conventional methods. *Br Dent J* 2001; 190 : 440-3.
- Ohki M, Okano T, Nakamura T. Factors determining the diagnostic accuracy of digitized conventional intra-oral radiographs. *Dentomaxillofac Radiol* 1994; 23 : 77-82.
- Senel B, Kamburoglu K, Uçok O, Yüksel SP, Ozen T, Avsever H. Diagnostic accuracy of different imaging modalities in detection of proximal caries. *Dentomaxillofac Radiol* 2010; 39 : 501-11.
- Tsuchida R, Araki K, Okano T. Evaluation of a limited cone-beam volumetric imaging system: comparison with film radiography in detecting incipient proximal caries. *Oral Surg Oral Med Oral Pathol Oral Radiol Endod* 2007; 104 : 412-6.
- Bader JD, Shugars DA. A systematic review of the performance of a laser fluorescence device for detecting caries. *J Am Dent Assoc* 2004; 135 : 1413-26.
- Haak R, Wicht MJ, Noack MJ. Conventional, digital and contrast-enhanced bite-wing radiographs in the decision to restore proximal carious lesions. *Caries Res* 2001; 35 : 193-9.
- Møystad A, Svanaes DB, Risnes S, Larheim TA, Grøndahl HG. Detection of approximal caries with a storage phosphor system. A comparison of enhanced digital images with dental X-ray film. *Dentomaxillofac Radiol* 1996; 25 : 202-6.
- Tam LE, McComb D. Diagnosis of occlusal caries: Part II. Recent diagnostic technologies. *J Can Dent Assoc* 2001; 67 : 459-63.
- Mitropoulos P, Rahiotis C, Stamatakis H, Kakaboura A. Diagnostic performance of the visual caries classification system ICDAS II versus radiography and micro-computed tomography for proximal caries detection: an in vitro study. *J Dent* 2010; 38 : 859-67.
- Jablonski-Momeni A, Stachniss V, Ricketts DN, Heinzl-Gutenbrunner M, Pieper K. Reproducibility and accuracy of the ICDAS-II for detection of occlusal caries in vitro. *Caries Res* 2008; 42 : 79-87.
- Pitts NB, Rimmer PA. An in vivo comparison of radiographic and directly assessed clinical caries status of posterior approximal surfaces in primary and permanent teeth. *Caries Res* 1992; 26 : 146-52.
- Peers A, Hiff FJ, Mitropoulos CM, Holloway PJ. Validity and reproducibility of clinical examination, fiber-optic transillumination, and bitewing radiology for the diagnosis of small approximal carious lesion: an in vitro study. *Caries Res* 1993; 27 : 307-11.
- Ismail AI, Sohn W, Tellez M, Amaya A, Sen A, Hasson H, et al. The International Caries Detection and Assessment System (ICDAS): an integrated system for measuring dental caries. *Community Dent Oral Epidemiol* 2007; 35 : 170-8.
- Yang J, Dutra V. Utility of radiology, laser fluorescence, and transillumination. *Dent Clin North Am* 2005; 49 : 739-52.
- Souza-Zaroni WC, Ciccone JC, Souza-Gabriel AE, Ramos RP, Corona SA, Palma-Dibb RG. Validity and reproducibility of different combinations of methods for occlusal caries detection: an in vitro comparison. *Caries Res* 2006; 40 : 194-201.
- Raper HR. A new kind of X-ray examination for preventive dentistry. *Int J Orthod Oral Surg* 1925; 11 : 76-86.
- Tyndall DA, Ludlow JB, Platin E, Nair M. A comparison of Kodak Ektaspeed Plus film and the Siemens Sidexis digital imaging system for caries detection using receiver operating characteristic analysis. *Oral Surg Oral Med Oral Pathol Oral Radiol Endod* 1998; 85 : 113-8.
- Syriopoulos K, Sanderink GC, Velders XL, van der Stelt PF. Radiographic detection of approximal caries: a comparison of dental films and digital imaging systems. *Dentomaxillofac Radiol* 2000; 29 : 312-8.

21. Abreu M Jr, Mol A, Ludlow JB. Performance of RVGui sensor and Kodak Ektaspeed Plus film for proximal caries detection. *Oral Surg Oral Med Oral Pathol Oral Radiol Endod* 2001; 91 : 381-5.
22. Nair MK, Nair UP. An in-vitro evaluation of Kodak Insight and Ektaspeed Plus film with a CMOS detector for natural proximal caries: ROC analysis. *Caries Res* 2001; 35 : 354-9.
23. Khan EA, Tyndall DA, Caplan D. Extraoral imaging for proximal caries detection: Bitewings vs scanogram. *Oral Surg Oral Med Oral Pathol Oral Radiol Endod* 2004; 98 : 730-7.
24. Wenzel A, Hintze H. The choice of gold standard for evaluating tests for caries diagnosis. *Dentomaxillofac Radiol* 1999; 28 : 132-6.
25. Haiter-Neto F, Wenzel A, Gotfredsen E. Diagnostic accuracy of cone beam computed tomography scans compared with intra-oral image modalities for detection of caries lesions. *Dentomaxillofac Radiol* 2008; 37 : 18-22.
26. Møystad A, Svanaes DB, Larheim TA, Grondahl HG. Effect of image magnification of digitized bitewing radiographs on approximal caries detection: an in vitro study. *Dentomaxillofac Radiol* 1995; 24 : 255-9.
27. Wenzel A, Pitts N, Verdonschot EH, Kalsbeek H. Developments in radiographic caries diagnosis. *J Dent* 1993; 21 : 131-40.
28. White SC, Yoon DC. Comparative performance of digital and conventional images for detecting proximal surface caries. *Dentomaxillofac Radiol* 1997; 26 : 32-8.
29. Kalathingal SM, Mol A, Tyndall DA, Caplan DJ. In vitro assessment of cone beam local computed tomography for proximal caries detection. *Oral Surg Oral Med Oral Pathol Oral Radiol Endod* 2007; 104 : 699-704.
30. Hintze H, Wenzel A, Danielsen B, Nyvad B. Reliability of visual examination, fibre-optic transillumination, and bitewing radiography, and reproducibility of direct visual examination following tooth separation for the identification of cavitated carious lesions in contacting approximal surfaces. *Caries Res* 1998; 32 : 204-9.
31. Wenzel A. Bitewing and digital bitewing radiography for detection of caries lesions. *J Dent Res* 2004; 83 Spec No C : C72-5.
32. Hintze H, Wenzel A, Jones C. In vitro comparison of D- and E-speed film radiography, RVG, and visualix digital radiography for the detection of enamel approximal and dentinal occlusal caries lesions. *Caries Res* 1994; 28 : 363-7.
33. Wenzel A, Borg E, Hintze H, Grondahl HG. Accuracy of caries diagnosis in digital images from charge-coupled device and storage phosphor systems: an in vitro study. *Dentomaxillofac Radiol* 1995; 24 : 250-4.
34. Tyndall DA, Rathore S. Cone-beam CT diagnostic applications: caries, periodontal bone assessment, and endodontic applications. *Dent Clin North Am* 2008; 52 : 825-41.
35. Hounsfield GN. Computerized transverse axial scanning (tomography). 1. Description of system. *Br J Radiol* 1973; 46 : 1016-22.
36. Lo EC, Zhi QH, Itthagarun A. Comparing two quantitative methods for studying remineralization of artificial caries. *J Dent* 2010; 38 : 352-9.
37. Ito K, Gomi Y, Sato S, Arai Y, Shinoda K. Clinical application of a new compact CT system to assess 3-D images for the preoperative treatment planning of implants in the posterior mandible. A case report. *Clin Oral Implants Res* 2001; 12 : 539-42.
38. Lascala CA, Panella J, Marques MM. Analysis of the accuracy of linear measurements obtained by cone beam computed tomography (CBCT-NewTom). *Dentomaxillofac Radiol* 2004; 33 : 291-4.
39. Honda K, Arai Y, Kashima M, Takano Y, Sawada K, Ejima K, et al. Evaluation of the usefulness of the limited cone-beam CT (3DX) in the assessment of the thickness of the roof of the glenoid fossa of the temporomandibular joint. *Dentomaxillofac Radiol* 2004; 33 : 391-5.
40. Ziegler CM, Woertche R, Brief J, Hassfeld S. Clinical indications for digital volume tomography in oral and maxillofacial surgery. *Dentomaxillofac Radiol* 2002; 31 : 126-30.
41. Mozzo P, Procacci C, Tacconi A, Martini PT, Andreis IA. A new volumetric CT machine for dental imaging based on the cone-beam technique: preliminary results. *Eur Radiol* 1998; 8 : 1558-64.
42. Danforth RA. Cone beam volume tomography: a new digital imaging option for dentistry. *J Calif Dent Assoc* 2003; 31 : 814-5.
43. Sukovic P. Cone beam computed tomography in craniofacial imaging. *Orthod Craniofac Res* 2003; 6(Suppl 1) : 31-6.
44. Scarfe WC. Imaging of maxillofacial trauma: evolutions and emerging revolutions. *Oral Surg Oral Med Oral Pathol Oral Radiol Endod* 2005; 100(2 Suppl) : S75-96.
45. Webber RL, Horton RA, Tyndall DA, Ludlow JB. Tuned-aperture computed tomography (TACT). Theory and application for three-dimensional dento-alveolar imaging. *Dentomaxillofac Radiol* 1997; 26 : 53-62.
46. Tyndall DA, Clifton TL, Webber RL, Ludlow JP, Horton RA. TACT imaging of primary caries. *Oral Surg Oral Med Oral Pathol Oral Radiol Endod* 1997; 84 : 214-25.
47. Nair MK, Tyndall DA, Ludlow JB, May K, Ye F. The effects of restorative material and location on the detection of simulated recurrent caries. A comparison of dental film, direct digital radiography and tuned aperture computed tomography. *Dentomaxillofac Radiol* 1998; 27 : 80-4.
48. Abreu Júnior M, Tyndall DA, Platin E, Ludlow JB, Phillips C. Two- and three-dimensional imaging modalities for the detection of caries. A comparison between film, digital radiography and tuned aperture computed tomography. *Dentomaxillofac Radiol* 1999; 28 : 152-7.
49. Nance RS, Tyndall DA, Levin LG, Trope M. Diagnosis of external root resorption using TACT (tuned-aperture computed tomography). *Endod Dent Traumatol* 2000; 16 : 24-8.
50. Nance R, Tyndall D, Levin LG, Trope M. Identification of root canals in molars by tuned-aperture computed tomography. *Int Endod J* 2000; 33 : 392-6.
51. Ramesh A, Ludlow JB, Webber RL, Tyndall DA, Paquette D. Evaluation of tuned aperture computed tomography (TACT) in the localization of simulated periodontal defects. *Dentomaxillofac Radiol* 2001; 30 : 319-24.
52. Terakado M, Hashimoto K, Arai Y, Honda M, Sekiwa T, Sato H. Diagnostic imaging with newly developed ortho cubic super-high resolution computed tomography (Ortho-CT). *Oral Surg Oral Med Oral Pathol Oral Radiol Endod* 2000; 89 : 509-18.

53. van Daatselaar AN, Dunn SM, Spoelder HJ, Germans DM, Renambot L, Bal HE, et al. Feasibility of local CT of dental tissues. *Dentomaxillofac Radiol* 2003; 32 : 173-80.
54. van Daatselaar AN, Tyndall DA, van der Stelt PF. Detection of caries with local CT. *Dentomaxillofac Radiol* 2003; 32 : 235-41.
55. van Daatselaar AN, Tyndall DA, Verheij H, van der Stelt PF. Minimum number of basis projections for caries detection with local CT. *Dentomaxillofac Radiol* 2004; 33 : 355-60.
56. van Daatselaar AN, van der Stelt PF, Weenen J. Effect of number of projections on image quality of local CT. *Dentomaxillofac Radiol* 2004; 33 : 361-9.
57. Akdeniz BG, Grondahl HG, Magnusson B. Accuracy of proximal caries depth measurements: comparison between limited cone beam computed tomography, storage phosphor and film radiography. *Caries Res* 2006; 40 : 202-7.
58. Young SM, Lee JT, Hodges RJ, Chang TL, Elashoff DA, White SC. A comparative study of high-resolution cone beam computed tomography and charge-coupled device sensors for detecting caries. *Dentomaxillofac Radiol* 2009; 38 : 445-51.
59. Qu X, Li G, Zhang Z, Ma X. Detection accuracy of in vitro approximal caries by cone beam computed tomography images. *Eur J Radiol* (in press).
60. Kamburoğlu K, Murat S, Yüksel SP, Cebeci AR, Paksoy CS. Occlusal caries detection by using a cone-beam CT with different voxel resolutions and a digital intraoral sensor. *Oral Surg Oral Med Oral Pathol Oral Radiol Endod* 2010; 109 : e63-9.
61. Kayipmaz S, Sezgin OS, Saricaoğlu ST, Can G. An in vitro comparison of diagnostic abilities of conventional radiography, storage phosphor, and cone beam computed tomography to determine occlusal and approximal caries. *Eur J Radiol* (in press).
62. Hanley JA, McNeil BJ. A method of comparing the areas under receiver operating characteristic curves derived from the same cases. *Radiology* 1983; 148 : 839-43.
63. McNeil BJ, Hanley JA. Statistical approaches to the analysis of receiver operating characteristic (ROC) curves. *Med Decis Making* 1984; 4 : 137-50.
64. Rothwell PM. Can overall results of clinical trials be applied to all patients? *Lancet* 1995; 345 : 1616-9.
65. Park YS, Bae KH, Chang J, Shon WJ. Theory of X-ray micro-computed tomography in dental research: application for the caries research. *J Korean Acad Conserv Dent* 2011; 36 : 98-107.